

Clear table

Instrument ID: IC-2										Analyst : IZ		Method: 300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst		
STD1	0.0000	0.0000	0.0000	0.0000	0.0090	0.0000	0.0000	0.0000_2023-07-18_	7/18/23 11:54		IZ/MA		
STD2	0.5190	0.6440	0.6590	2.1930	0.5650	1.2310	3.4630	2023-07-18_	7/18/23 12:25		IZ/MA		
STD3	0.7210	1.2220	1.2040	3.9830	1.0030	1.9630	6.0460	2023-07-18_	7/18/23 12:55		IZ/MA		
STD4	1.0180	1.5110	1.5050	5.1180	1.2620	2.5080	7.5420	2023-07-18_	7/18/23 13:26		IZ/MA		
STD5	1.9840	2.9150	2.9260	9.6980	2.4210	4.8360	14.4150	2023-07-18_	7/18/23 13:57		IZ/MA		
STD6	3.7660	5.9300	5.9290	19.6870	4.9030	9.6670	29.4280	2023-07-18_	7/18/23 14:27		IZ/MA		
STD7	5.1680	7.5780	7.5770	25.3210	6.3470	12.7960	38.1050	2023-07-18_	7/18/23 14:58		IZ/MA		
ICV	2.0250	2.9710	2.9980	9.8960	2.5030	4.9290	14.7030	2023-07-18_	7/18/23 15:29		IZ/MA		
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-07-18_	7/18/23 16:00		IZ/MA		
CCV	2.0310	3.1920	3.1320	10.5020	2.6060	5.2680	15.5020	2023-08-07_	8/7/23 9:31		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-07_	8/7/23 10:05		IZ/MA		
LB126792BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-07_	8/7/23 10:36		IZ/MA		
LB126792BSW	2.0430	3.1980	3.1790	10.5300	2.6250	5.4930	15.6440	2023-08-07_	8/7/23 11:07		IZ/MA		
O3923-01	0.0000	0.0000	0.0000	0.0000	0.1870	0.0000	0.9660	2023-08-07_	8/7/23 11:44		IZ/MA		
O3923-01MS	1.7470	2.9210	2.9080	9.5340	2.4330	5.0780	14.2300	2023-08-07_	8/7/23 12:14		IZ/MA		
O3923-01MSD	1.8320	2.8480	2.8650	9.3870	2.4080	5.0830	13.9540	2023-08-07_	8/7/23 12:45		IZ/MA		
CCV	2.0470	3.2140	3.1940	10.4770	2.6500	5.6480	15.9530	2023-08-07_	8/7/23 13:16		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-08-07_	8/7/23 13:46		IZ/MA		

Clear table

Instrument ID: IC-2 Analyst : IZ Method: 300.0 / 9056A


ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2023-07-18_	7/18/23 11:54		IZ/MA
STD2	0.5190	0.6440	0.6590	2.1930	0.5650	1.2310	3.4630	_2023-07-18_	7/18/23 12:25		IZ/MA
STD3	0.7210	1.2220	1.2040	3.9830	1.0030	1.9630	6.0460	_2023-07-18_	7/18/23 12:55		IZ/MA
STD4	1.0180	1.5110	1.5050	5.1180	1.2620	2.5080	7.5420	_2023-07-18_	7/18/23 13:26		IZ/MA
STD5	1.9840	2.9150	2.9260	9.6980	2.4210	4.8360	14.4150	_2023-07-18_	7/18/23 13:57		IZ/MA
STD6	3.7660	5.9300	5.9290	19.6870	4.9030	9.6670	29.4280	_2023-07-18_	7/18/23 14:27		IZ/MA
STD7	5.1680	7.5780	7.5770	25.3210	6.3470	12.7960	38.1050	_2023-07-18_	7/18/23 14:58		IZ/MA

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	29.7500	7.3333	9.8333	9.6500	13.0000	23.1000	15.4333
STD3	-9.8750	1.8333	0.3333	-0.4250	0.3000	-1.8500	0.7667
STD4	1.8000	0.7333	0.3333	2.3600	0.9600	0.3200	0.5600
STD5	-0.8000	-2.8333	-2.4667	-3.0200	-3.1600	-3.2800	-3.9000
STD6	-5.8500	-1.1667	-1.1833	-1.5650	-1.9400	-3.3300	-1.9067
STD7	3.3600	1.0400	1.0267	1.2840	1.5520	2.3680	2.9865

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0893931 \cdot A + 3.11575$
 RSD: 7.560 %
 Correlation coefficient: 0.996880




K3 = 0 K2 = 0 K1 = 0.0893931 K0 = 3.11575
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	1.069	85.37	0.4	20	6.387	
3	1.887	124.8	0.8	20	6.387	
4	2.787	195.8	1	20	6.387	
5	6.176	409	2	20	6.387	
6	12.27	807.7	4	20	6.387	
7	17.4	1121	5	20	6.387	

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.129379 \cdot A + 2.57559$
 RSD: 2.187 %
 Correlation coefficient: 0.999739




K3 = 0 K2 = 0 K1 = 0.129379 K0 = 2.57559
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.821	79.67	0.6	20	8.957		
3	11.94	169	1.2	20	8.957		
4	15.21	213.7	1.5	20	8.957		
5	31.35	430.7	3	20	8.957		
6	65.62	896.7	6	20	8.957		
7	84.19	1152	7.5	20	8.957		

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0571447 \cdot A + 2.7868$
 RSD: 2.142 %
 Correlation coefficient: 0.999750




K3 = 0 K2 = 0 K1 = 0.0571447 K0 = 2.7868
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.91	181.7	0.6	20	10.82		
3	22.02	372.5	1.2	20	10.82		
4	28.54	478.1	1.5	20	10.82		
5	58.22	975.2	3	20	10.82		
6	117	2026	6	20	10.82		
7	150.1	2603	7.5	20	10.82		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.288518 \cdot A + 10.6539$
 RSD: 2.666 %
 Correlation coefficient: 0.999612




K3 = 0 K2 = 0 K1 = 0.288518 K0 = 10.6539
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	
2	6.375	115.1	2	20	13.99	
3	13.22	239.2	4	20	13.99	
4	17.19	317.8	5	20	13.99	
5	35.66	635.4	10	20	13.99	
6	74.24	1328	20	20	13.99	
7	95.06	1718	25	20	13.99	

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0499072 \cdot A + 2.72307$
 RSD: 3.120 %
 Correlation coefficient: 0.999469




K3 = 0 K2 = 0 K1 = 0.0499072 K0 = 2.72307
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.794	171.7	0.5	20	15.52		
3	15.93	347.4	1	20	15.52		
4	20.7	451.3	1.25	20	15.52		
5	42.9	915.4	2.5	20	15.52		
6	91.09	1910	5	20	15.52		
7	118.8	2489	6.25	20	15.52		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.0597633 \cdot A + 12.5173$
 RSD: 4.814 %
 Correlation coefficient: 0.998736




K3 = 0 K2 = 0 K1 = 0.0597633 K0 = 12.5173
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.234	202.4	1	20	22.61		
3	9.564	447.6	2	20	22.61		
4	13.76	629.8	2.5	20	22.61		
5	32.81	1409	5	20	22.61		
6	75.03	3026	10	20	22.61		
7	103.9	4073	12.5	20	22.61		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-071823.mtw
 Equation: $Q = 0.162956 \cdot A + 15.5782$
 RSD: 3.392 %
 Correlation coefficient: 0.999373



K3 = 0 K2 = 0 K1 = 0.162956 K0 = 15.5782
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0.5229		20.4	0	20		26.01
2	11.32		329.4	3	20		26.01
3	22.74		646.5	6	20		26.01
4	29.39		830	7.5	20		26.01
5	60.94		1674	15	20		26.01
6	130.3		3516	30	20		26.01
7	169.7		4581	37.5	20		26.01

Report date: 7/19/2023 11:11:53 AM
Printed by: wet


Ident: STD1
Analysis from: 7/18/2023 11:54:38 AM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34856

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:01 AM
Printed by: wet


Ident: STD2
Analysis from: 7/18/2023 12:25:16 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34857

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.32	1.167	1.03	2.17	81.326	7.00
2	9.12	0.201	5.82	12.26	79.672	6.86
3	10.98	0.233	10.91	22.98	181.715	15.65
4	14.19	0.275	6.37	13.42	115.119	9.91
5	16.21	0.326	7.79	16.41	171.712	14.78
6	21.82	0.586	4.23	8.91	202.440	17.43
7	25.12	0.434	11.32	23.83	329.435	28.36
7	28.00	0.460	47.48	99.99	1161.418	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:10 AM
Printed by: wet


Ident: STD3
Analysis from: 7/18/2023 12:55:54 PM
File: _2023-07-18_

Last save: 7/18/2023 3:32:11 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34858

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.27	1.020	1.90	1.95	126.563	5.39
2	9.13	0.202	11.94	12.27	168.983	7.19
3	10.98	0.237	22.02	22.62	372.529	15.86
4	14.17	0.274	13.22	13.59	239.159	10.18
5	16.17	0.323	15.93	16.37	347.382	14.79
6	21.86	0.580	9.56	9.83	447.587	19.06
7	25.16	0.426	22.74	23.37	646.495	27.53
7	28.00	0.437	97.31	100.00	2348.698	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:18 AM
Printed by: wet


Ident: STD4
Analysis from: 7/18/2023 1:26:31 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34859

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.35	1.020	2.76	2.16	192.844	6.19
2	9.15	0.202	15.21	11.92	213.663	6.86
3	11.01	0.235	28.54	22.38	478.126	15.36
4	14.25	0.280	17.19	13.47	317.821	10.21
5	16.26	0.322	20.70	16.23	451.305	14.49
6	22.01	0.566	13.76	10.79	629.820	20.23
7	25.33	0.424	29.39	23.04	830.037	26.66
7	28.00	0.436	127.54	99.99	3113.615	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:26 AM
Printed by: wet


Ident: STD5
Analysis from: 7/18/2023 1:57:09 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34860

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.27	0.962	6.17	2.30	409.042	6.34
2	9.07	0.198	31.35	11.70	430.714	6.68
3	10.89	0.235	58.22	21.72	975.219	15.12
4	14.05	0.269	35.66	13.30	635.355	9.85
5	15.98	0.314	42.90	16.00	915.443	14.20
6	21.75	0.533	32.81	12.24	1408.769	21.85
7	25.05	0.413	60.94	22.73	1673.615	25.95
7	28.00	0.418	268.05	100.00	6448.158	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:33 AM
Printed by: wet


Ident: STD6
Analysis from: 7/18/2023 2:27:46 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34861

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.26	0.960	12.27	2.17	807.679	5.98
2	9.07	0.196	65.62	11.60	896.723	6.64
3	10.88	0.244	117.03	20.69	2026.370	15.00
4	14.04	0.269	74.24	13.13	1327.749	9.83
5	15.93	0.308	91.08	16.11	1910.213	14.14
6	21.75	0.503	75.02	13.27	3025.506	22.39
7	25.08	0.407	130.25	23.03	3516.225	26.03
7	28.00	0.412	565.52	100.00	13510.466	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:12:39 AM
Printed by: wet


Ident: STD7
Analysis from: 7/18/2023 2:58:23 PM
File: _2023-07-18_

Last save: 7/18/2023 3:27:26 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34862

Last save: 7/18/2023 2:55:

SAMPLE:
: IZ/MA
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.34	0.949	17.40	2.35	1121.440	6.32
2	9.09	0.196	84.19	11.39	1151.573	6.49
3	10.89	0.245	150.10	20.31	2603.211	14.68
4	14.06	0.273	95.06	12.86	1718.350	9.69
5	15.93	0.306	118.78	16.07	2488.841	14.03
6	21.80	0.491	103.93	14.06	4072.751	22.96
7	25.15	0.407	169.72	22.96	4581.095	25.83
7	28.00	0.410	739.17	100.00	17737.261	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:14:06 AM
Printed by: wet


Ident: ICV
Analysis from: 7/18/2023 3:29:09 PM
File: _2023-07-18_

Last save: 7/18/2023 3:57:09 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34863

Last save: 7/18/2023 3:26:

SAMPLE:
: IZ/MA
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.26	0.954	6.38	2.31	418.174	6.33
2	9.03	0.197	32.27	11.67	439.302	6.65
3	10.83	0.233	60.17	21.76	1000.409	15.15
4	13.96	0.265	36.92	13.35	649.049	9.83
5	15.86	0.310	44.57	16.12	948.693	14.36
6	21.71	0.528	33.75	12.20	1440.075	21.80
7	25.01	0.412	62.48	22.59	1708.970	25.88
7	28.00	0.414	276.55	100.00	6604.672	100.00

This report has been created by IC Net
METROHM LTD

Report date: 7/19/2023 11:14:16 AM
Printed by: wet


Ident: ICB
Analysis from: 7/18/2023 4:00:04 PM
File: _2023-07-18_

Last save: 7/19/2023 8:57:46 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 34864

Last save: 7/18/2023 3:26:

SAMPLE:
: IZ/MA
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:24:19 AM
Printed by: wet


Ident: CCV
Analysis from: 8/7/2023 9:31:48 AM
File: _2023-08-07_

Last save: 8/7/2023 9:59:48 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35052

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.60	1.198	5.30	2.05	419.515	6.01
2	9.53	0.221	31.47	12.17	473.555	6.78
3	11.52	0.267	57.65	22.29	1047.306	15.00
4	15.00	0.312	33.20	12.84	691.044	9.90
5	17.14	0.379	38.69	14.96	989.619	14.17
6	22.15	0.599	33.44	12.93	1553.560	22.25
7	25.53	0.463	58.85	22.76	1807.024	25.88
7	28.00	0.491	258.58	100.00	6981.623	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:24:40 AM
Printed by: wet


Ident: CCB
Analysis from: 8/7/2023 10:05:51 AM
File: _2023-08-07_

Last save: 8/7/2023 10:33:51 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35053

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:25:11 AM
Printed by: wet


Ident: LB126792BLW
Analysis from: 8/7/2023 10:36:29 AM
File: _2023-08-07_

Last save: 8/9/2023 9:23:47 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35054

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:25:30 AM
Printed by: wet


Ident: LB126792BSW
Analysis from: 8/7/2023 11:07:06 AM
File: _2023-08-07_

Last save: 8/9/2023 9:23:47 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35055

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.50	1.118	5.64	2.10	422.150	5.94
2	9.36	0.216	32.22	11.98	474.409	6.68
3	11.27	0.260	59.63	22.17	1063.774	14.97
4	14.62	0.300	34.76	12.93	693.029	9.76
5	16.67	0.362	40.83	15.18	997.254	14.04
6	21.92	0.588	35.63	13.25	1628.747	22.93
7	25.27	0.458	60.22	22.39	1824.454	25.68
7	28.00	0.472	268.93	100.00	7103.818	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:25:45 AM
Printed by: wet


Ident: 03923-01
Analysis from: 8/7/2023 11:44:20 AM
File: _2023-08-07_

Last save: 8/7/2023 12:12:20 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35056

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 23
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.61	0.377	0.81	60.22	20.359	47.01
6	0.00	0.000	0.00	0.00	0.000	0.00
7	25.20	0.644	0.53	39.48	22.950	52.99
7	28.00	0.146	1.34	99.70	43.309	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:25:56 AM
Printed by: wet


Ident: 03923-01MS
Analysis from: 8/7/2023 12:14:57 PM
File: _2023-08-07_

Last save: 8/7/2023 2:05:12 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35057

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 24
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.37	1.109	4.82	1.96	355.999	5.53
2	9.24	0.214	29.54	12.01	431.640	6.70
3	11.13	0.261	54.03	21.96	968.897	15.04
4	14.43	0.293	32.18	13.08	623.997	9.69
5	16.42	0.354	38.73	15.74	920.470	14.29
6	21.75	0.601	31.86	12.95	1490.012	23.13
7	25.06	0.455	54.89	22.31	1650.900	25.63
7	28.00	0.470	246.06	100.00	6441.915	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:26:11 AM
Printed by: wet


Ident: O3923-01MSD
Analysis from: 8/7/2023 12:45:35 PM
File: _2023-08-07_

Last save: 8/7/2023 1:13:35 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35058

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 25
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.45	1.103	5.10	2.09	374.984	5.88
2	9.27	0.215	28.90	11.86	420.301	6.59
3	11.16	0.257	54.07	22.19	953.811	14.95
4	14.46	0.293	31.65	12.99	613.803	9.62
5	16.45	0.353	38.17	15.67	910.294	14.26
6	21.83	0.598	32.03	13.15	1491.695	23.37
7	25.15	0.455	53.73	22.05	1617.061	25.34
7	28.00	0.468	243.64	100.00	6381.949	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:26:29 AM
Printed by: wet


Ident: CCV
Analysis from: 8/7/2023 1:16:18 PM
File: _2023-08-07_

Last save: 8/7/2023 1:44:19 PM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35059

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.35	1.039	5.99	2.15	423.164	5.87
2	9.18	0.211	33.31	11.98	476.853	6.61
3	11.04	0.254	61.34	22.05	1069.196	14.83
4	14.28	0.289	36.12	12.98	689.330	9.56
5	16.25	0.348	42.73	15.36	1007.575	13.98
6	21.62	0.578	37.31	13.41	1680.813	23.31
7	24.94	0.459	61.34	22.05	1862.394	25.83
7	28.00	0.454	278.15	100.00	7209.324	100.00

This report has been created by IC Net
METROHM LTD

Report date: 8/9/2023 9:27:30 AM
Printed by: wet


Ident: CCB
Analysis from: 8/7/2023 1:46:56 PM
File: _2023-08-07_

Last save: 8/9/2023 9:27:24 AM

Method: AnionsIC2-071823.mtw
Run operator: wet
Analysis number: 35060

Last save: 8/4/2023 11:24:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD