

66129514

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/13/2024 14:12

Reviewed by : RM Instrument ID : Konelab

Test: Total CN

Sample Id	<i>M/L</i> Result	Dil. 1 +	Response	Errors
ICV1	95.266	0.0	0.080	
ICB1	0.730	0.0	0.002	
CCV1	245.958	0.0	0.204	
CCB1	0.650	0.0	0.001	
RL CHECK	4.942	0.0	0.005	
PB159001BL	0.421	0.0	0.001	
P1404-02	0.459	0.0	0.001	
P1404-02DUP	0.399	0.0	0.001	
CCV2	245.994	0.0	0.204	
CCB2	0.410	0.0	0.001	

N	10
Mean	59.523
SD	102.5672
CV%	172.32

Aquakem v. 7.2AQ1

Results from time period:

Tue Feb 13 13:48:16 2024

Tue Feb 13 14:11:08 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.5401	µg/l	2/13/2024 10:44:00	
5.OPPBCN	A	Total CN	P	4.9211	µg/l	2/13/2024 10:44:01	
10PPBCN	A	Total CN	P	9.9666	µg/l	2/13/2024 10:44:02	
50PPBCN	A	Total CN	P	49.0879	µg/l	2/13/2024 10:44:03	
100PPBCN	A	Total CN	P	99.2612	µg/l	2/13/2024 10:44:04	
250PPBCN	A	Total CN	P	251.9653	µg/l	2/13/2024 10:44:05	
500PPBCN	A	Total CN	P	499.2578	µg/l	2/13/2024 10:44:06	
ICV1	S	Total CN	P	95.2658	µg/l	2/13/2024 13:48:17	
ICB1	S	Total CN	P	0.7304	µg/l	2/13/2024 13:48:18	
CCV1	S	Total CN	P	245.9579	µg/l	2/13/2024 13:48:21	
CCB1	S	Total CN	P	0.6502	µg/l	2/13/2024 13:48:25	
RL CHECK	S	Total CN	P	4.9423	µg/l	2/13/2024 13:55:56	
PB159001BL	S	Total CN	P	0.4213	µg/l	2/13/2024 14:03:26	
P1404-02	S	Total CN	P	0.4587	µg/l	2/13/2024 14:03:32	
P1404-02DUP	S	Total CN	P	0.3986	µg/l	2/13/2024 14:03:34	
CCV2	S	Total CN	P	245.9941	µg/l	2/13/2024 14:11:03	
CCB2	S	Total CN	P	0.4097	µg/l	2/13/2024 14:11:08	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/13/2024 11:08

Reviewed by : RM

Instrument ID : Konelab


Test Total CN

Accepted 2/13/2024 10:53

Factor 1213
Bias 0.001

Coeff. of det. 0.999970

Errors



	Calibrator	Response	Calc. con.	Conc.	^{Re} Errors
1	0.0PPBCN	0.001	0.5401	0.0000	-1.6
2	5.0PPBCN	0.005	4.9211	5.0000	-0.3
3	10PPBCN	0.009	9.9666	10.0000	-1.8
4	50PPBCN	0.041	49.0879	50.0000	-0.7
5	100PPBCN	0.083	99.2612	100.0000	0.8
6	250PPBCN	0.209	251.9653	250.0000	-0.1
7	500PPBCN	0.413	499.2578	500.0000	

02/13/2024
RM