

(15129 519)

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12/11 Instrument ID : Konelab

2/13/2024 16:52

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.942	0.0	0.210	
ICB1	0.023	0.0	0.018	
CCV1	0.937	0.0	0.209	
CCB1	0.017	0.0	0.016	
RL CHEK	0.113	0.0	0.036	
PB158890BL	0.019	0.0	0.017	
PB158890BS	0.948	0.0	0.211	
P1404-01	0.033	0.0	0.020	
P1404-01DUP	0.033	0.0	0.020	
P1404-01MS	0.905	0.0	0.202	
P1404-01MSD	0.924	0.0	0.206	
P1410-01	5.975	0.0	1.261	Test limit high
P1410-03	12.676	0.0	2.660	Init abs., Test limit high
P1410-05	5.817	0.0	1.228	Test limit high
CCV2	0.942	0.0	0.210	
CCB2	0.023	0.0	0.018	
P1429-01	4.969	0.0	1.050	Test limit high
P1429-02	5.062	0.0	1.070	Test limit high
P1434-01	0.901	0.0	0.201	
P1452-01	0.451	0.0	0.107	
P1452-03	1.351	0.0	0.295	
P1458-01	0.019	0.0	0.017	
CCV3	0.975	0.0	0.216	
CCB3	0.027	0.0	0.018	
P1410-01DLX5	1.215	0.0	0.266	
P1410-03DLX10	1.204	0.0	0.264	
P1410-05DLX5	1.153	0.0	0.254	
P1429-01DLX5	0.990	0.0	0.220	
P1429-02DLX5	0.993	0.0	0.220	
CCV4	0.960	0.0	0.213	
CCB4	0.022	0.0	0.017	

113% (50-150)
02/13/2024 RM

N 31
Mean 1.633
SD 2.6712
CV% 163.59

Aquakem v. 7.2AQ1

Results from time period:

Tue Feb 13 14:55:24 2024

Tue Feb 13 16:52:34 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-I P		0.0169	mg/l	2/13/2024 14:55:24	
0.1PPM	A	Ammonia-I P		0.1123	mg/l	2/13/2024 14:55:25	
0.2PPM	A	Ammonia-I P		0.2005	mg/l	2/13/2024 14:55:26	
0.4PPM	A	Ammonia-I P		0.3917	mg/l	2/13/2024 14:55:27	
1.0PPM	A	Ammonia-I P		0.9778	mg/l	2/13/2024 14:55:28	
1.3PPM	A	Ammonia-I P		1.2994	mg/l	2/13/2024 14:55:29	
2.0PPM	A	Ammonia-I P		2.0347	mg/l	2/13/2024 14:55:30	
ICV1	S	Ammonia-I P		0.9422	mg/l	2/13/2024 15:32:49	
ICB1	S	Ammonia-I P		0.0229	mg/l	2/13/2024 15:32:51	
CCV1	S	Ammonia-I P		0.9373	mg/l	2/13/2024 15:32:53	
CCB1	S	Ammonia-I P		0.017	mg/l	2/13/2024 15:32:55	
RL CHEK	S	Ammonia-I P		0.1132	mg/l	2/13/2024 15:32:57	
PB158890BL	S	Ammonia-I P		0.0186	mg/l	2/13/2024 15:32:59	
PB158890BS	S	Ammonia-I P		0.948	mg/l	2/13/2024 15:43:33	
P1404-01	S	Ammonia-I P		0.0325	mg/l	2/13/2024 15:43:34	
P1404-01DUP	S	Ammonia-I P		0.0327	mg/l	2/13/2024 15:43:35	
P1404-01MS	S	Ammonia-I P		0.9052	mg/l	2/13/2024 15:43:36	
P1404-01MSD	S	Ammonia-I P		0.9236	mg/l	2/13/2024 15:43:37	
P1410-01	S	Ammonia-I P		5.9753	mg/l	2/13/2024 15:43:38	
P1410-03	S	Ammonia-I P		12.6761	mg/l	2/13/2024 15:43:42	
P1410-05	S	Ammonia-I P		5.8168	mg/l	2/13/2024 15:43:43	
CCV2	S	Ammonia-I P		0.9421	mg/l	2/13/2024 15:54:13	
CCB2	S	Ammonia-I P		0.0233	mg/l	2/13/2024 15:54:16	
P1429-01	S	Ammonia-I P		4.9687	mg/l	2/13/2024 15:54:19	
P1429-02	S	Ammonia-I P		5.0625	mg/l	2/13/2024 15:54:20	
P1434-01	S	Ammonia-I P		0.901	mg/l	2/13/2024 15:54:21	
P1452-01	S	Ammonia-I P		0.4505	mg/l	2/13/2024 15:54:22	
P1452-03	S	Ammonia-I P		1.3508	mg/l	2/13/2024 15:54:23	
P1458-01	S	Ammonia-I P		0.0189	mg/l	2/13/2024 15:54:24	
CCV3	S	Ammonia-I P		0.9752	mg/l	2/13/2024 16:01:29	
CCB3	S	Ammonia-I P		0.027	mg/l	2/13/2024 16:01:31	
P1410-01DLX5	S	Ammonia-I P		1.2145	mg/l	2/13/2024 16:52:27	
P1410-03DLX10	S	Ammonia-I P		1.2036	mg/l	2/13/2024 16:52:28	
P1410-05DLX5	S	Ammonia-I P		1.1534	mg/l	2/13/2024 16:52:30	
P1429-01DLX5	S	Ammonia-I P		0.99	mg/l	2/13/2024 16:52:31	
P1429-02DLX5	S	Ammonia-I P		0.9935	mg/l	2/13/2024 16:52:32	
CCV4	S	Ammonia-I P		0.9596	mg/l	2/13/2024 16:52:33	
CCB4	S	Ammonia-I P		0.0217	mg/l	2/13/2024 16:52:34	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

2/13/2024 14:56


Test Ammonia-N

Accepted 2/13/2024 14:56

Factor 4.788
Bias 0.013

Coeff. of det. 0.999004

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.016	0.0169	0.0000	
2	NH3-2PPM	0.036	0.1123	0.1000	12.3
3	NH3-2PPM	0.055	0.2005	0.2000	0.3
4	NH3-2PPM	0.095	0.3917	0.4000	-2.1
5	NH3-2PPM	0.217	0.9778	1.0000	-2.2
6	NH3-2PPM	0.284	1.2994	1.3333	0.0
7	NH3-2PPM	0.438	2.0347	2.0000	1.7

02/13/2024
RM