


7471B

LB129877

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.7675e-004

C= -1.7498e-002

Rho= 0.9999830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.013	-0.013	24	0.000	24					
0.2	0.200	0.198	-0.002	1219	0.0 %	1219					
2.5	2.500	2.490	-0.010	14188	0.0 %	14188					
5.0	5.000	5.045	0.045	28644	0.0 %	28644					
7.5	7.500	7.501	0.001	42538	0.0 %	42538					
10.0	10.000	9.979	-0.021	56557	0.0 %	56557					

0/0 D

1

0

1

0

0

LB129877 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Mar 2024 10:31:32

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	24	-	0.00007471B		PPB	14 Mar 2024 10:47:02	S	Std
0.2 - 1	SD-2	1219	-	0.20007471B		PPB	14 Mar 2024 10:49:19	S	Std
2.5 - 1	SD-5	14188	-	2.50007471B		PPB	14 Mar 2024 10:51:36	S	Std
5.0 - 1	SD-5	28644	-	5.00007471B		PPB	14 Mar 2024 10:53:53	S	Std
7.5 - 1	SD-5	42538	-	7.50007471B		PPB	14 Mar 2024 10:56:13	S	Std
10.0 - 1	SD-5	56557	-	10.00007471B		PPB	14 Mar 2024 11:01:24	S	Std
ICV48 - 1	ICV48	21059	3.7046	-7471B		PPB	14 Mar 2024 11:04:17	U	SMPL
ICB48 - 1	ICB48	-380	-0.0847	-7471B		PPB	14 Mar 2024 11:06:34	U	SMPL
CCV66 - 1	CCV66	28478	5.0159	-7471B		PPB	14 Mar 2024 11:08:54	U	SMPL
CCB66 - 1	CCB66	-174	-0.0483	-7471B		PPB	14 Mar 2024 11:11:09	U	SMPL
CRA - 1	CRA	1181	0.1912	-7471B		PPB	14 Mar 2024 11:13:32	U	SMPL
HighStd - 1	HighStd	54556	9.6252	-7471B		PPB	14 Mar 2024 11:15:48	U	SMPL
ChkStd - 1	ChkStd	38194	6.7332	-7471B		PPB	14 Mar 2024 11:18:03	U	SMPL
PB159590BL - 1	PBS	-239	-0.0597	-7471B		PPB	14 Mar 2024 11:20:34	U	SMPL
PB159590BS - 1	LCSS	20882	3.6734	-7471B		PPB	14 Mar 2024 11:23:02	U	SMPL
P1187-01 - 1	LOD-MDL-SOIL-01-QT1-2024	339	0.0424	-7471B		PPB	14 Mar 2024 11:28:52	U	SMPL
P1187-02 - 1	LOQ-SOIL-02-QT1-2024	1376	0.2257	-7471B		PPB	14 Mar 2024 11:31:08	U	SMPL
P1733-01 - 1	WC-1	4516	0.7807	-7471B		PPB	14 Mar 2024 11:33:25	U	SMPL
P1733-05 - 1	WC-2	4519	0.7812	-7471B		PPB	14 Mar 2024 11:35:42	U	SMPL
P1733-09 - 1	WC-3	10223	1.7894	-7471B		PPB	14 Mar 2024 11:38:32	U	SMPL
CCV67 - 1	CCV67	27964	4.9251	-7471B		PPB	14 Mar 2024 11:43:25	U	SMPL
CCB67 - 1	CCB67	-178	-0.0490	-7471B		PPB	14 Mar 2024 11:45:42	U	SMPL
P1744-01 - 1	RT-3488	4113	0.7095	-7471B		PPB	14 Mar 2024 11:48:04	U	SMPL
P1744-01DUP - 1	RT-3488DUP	3817	0.6571	-7471B		PPB	14 Mar 2024 11:53:01	U	SMPL
P1744-01MS - 1	RT-3488MS	24641	4.3377	-7471B		PPB	14 Mar 2024 11:55:16	U	SMPL
P1744-01MSD - 1	RT-3488MSD	25710	4.5267	-7471B		PPB	14 Mar 2024 11:58:29	U	SMPL
P1744-01LX5 - 1		702	0.1066	-7471B		PPB	14 Mar 2024 12:00:45	U	SMPL
P1744-01A - 1		25637	4.5138	-7471B		PPB	14 Mar 2024 12:03:46	U	SMPL
CCV68 - 1	CCV68	29720	5.2354	-7471B		PPB	14 Mar 2024 12:06:03	U	SMPL
CCB68 - 1	CCB68	-136	-0.0415	-7471B		PPB	14 Mar 2024 12:08:26	U	SMPL