

Lb129903

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

3/15/2024 12:59

Test: Phenolics-

Sample Id	<i>Null</i> Result	Dil. 1 +	Response	Errors
ICV1	0.970	0.0	0.103	
ICB1	0.004	0.0	0.007	
CCV1	0.982	0.0	0.104	
CCB1	-0.007	0.0	0.006	
PB159615BL	-0.008	0.0	0.006	
PB159615BS	0.967	0.0	0.103	
P1764-10	-0.025	0.0	0.004	
P1764-11	-0.012	0.0	0.005	
P1764-11DUP	-0.003	0.0	0.006	
P1764-11MS	1.028	0.0	0.109	
P1764-11MSD	1.027	0.0	0.108	
IDOC-1	0.962	0.0	0.102	
IDOC-2	0.962	0.0	0.102	
IDOC-3	0.968	0.0	0.103	
IDOC-4	0.973	0.0	0.103	
CCV2	0.983	0.0	0.104	
CCB2	-0.002	0.0	0.006	

N 17
Mean 0.575
SD 0.5024
CV% 87.43

Aquakem v. 7.2AQ1

Results from time period:

Fri Mar 15 11:47:40 2024

Fri Mar 15 12:47:16 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0055	mg/l	3/15/2024 11:47:40	
0.05PPM	A	Phenolics-	P	0.061	mg/l	3/15/2024 11:47:41	
0.1PPM	A	Phenolics-	P	0.1089	mg/l	3/15/2024 11:47:42	
0.25PPM	A	Phenolics-	P	0.2579	mg/l	3/15/2024 11:47:43	
0.50PPM	A	Phenolics-	P	0.486	mg/l	3/15/2024 11:47:44	
1.0PPM	A	Phenolics-	P	0.9797	mg/l	3/15/2024 11:47:45	
2.0PPM	A	Phenolics-	P	2.0119	mg/l	3/15/2024 11:47:46	
ICV1	S	Phenolics-	P	0.9702	mg/l	3/15/2024 12:31:20	
ICB1	S	Phenolics-	P	0.0041	mg/l	3/15/2024 12:31:22	
CCV1	S	Phenolics-	P	0.9816	mg/l	3/15/2024 12:31:23	
CCB1	S	Phenolics-	P	-0.0071	mg/l	3/15/2024 12:31:25	
PB159615BL	S	Phenolics-	P	-0.0076	mg/l	3/15/2024 12:31:27	
PB159615BS	S	Phenolics-	P	0.9669	mg/l	3/15/2024 12:31:30	
P1764-10	S	Phenolics-	P	-0.0249	mg/l	3/15/2024 12:41:21	
P1764-11	S	Phenolics-	P	-0.0122	mg/l	3/15/2024 12:41:24	
P1764-11DUP	S	Phenolics-	P	-0.0027	mg/l	3/15/2024 12:41:26	
P1764-11MS	S	Phenolics-	P	1.0284	mg/l	3/15/2024 12:41:27	
P1764-11MSD	S	Phenolics-	P	1.0268	mg/l	3/15/2024 12:41:29	
IDOC-1	S	Phenolics-	P	0.9623	mg/l	3/15/2024 12:41:31	
IDOC-2	S	Phenolics-	P	0.9617	mg/l	3/15/2024 12:41:32	
IDOC-3	S	Phenolics-	P	0.9677	mg/l	3/15/2024 12:47:11	
IDOC-4	S	Phenolics-	P	0.973	mg/l	3/15/2024 12:47:12	
CCV2	S	Phenolics-	P	0.9831	mg/l	3/15/2024 12:47:13	
CCB2	S	Phenolics-	P	-0.0022	mg/l	3/15/2024 12:47:16	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

3/15/2024 11:51


Test Phenolics-

Accepted 3/15/2024 11:51

Factor 10.07
Bias 0.006

Coeff. of det. 0.999669

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	-0.0055	0.0000	-
2	Phenol-2	0.013	0.0610	0.0500	22.0
3	Phenol-2	0.017	0.1089	0.1000	8.9
4	Phenol-2	0.032	0.2579	0.2500	3.2
5	Phenol-2	0.055	0.4860	0.5000	-2.8
6	Phenol-2	0.104	0.9797	1.0000	-1.0
7	Phenol-2	0.206	2.0119	2.0000	0.6

03/15/2024
RM