

66130132

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

4/2/2024 15:15

Reviewed by : RM Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.991	0.0	0.116	
ICB1	0.003	0.0	0.008	
CCV1	1.000	0.0	0.117	
CCB1	0.001	0.0	0.008	
PB159952BL	-0.009	0.0	0.007	
PB159952BS	0.999	0.0	0.117	
P1911-25	0.424	0.0	0.054	
CCV2	1.004	0.0	0.117	
CCB2	0.002	0.0	0.008	

N	9
Mean	0.490
SD	0.5004
CV%	102.02

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 02 14:08:43 2024

Tue Apr 02 15:09:34 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0002	mg/l	4/2/2024 14:28:12	
0.05PPM	A	Phenolics-	P	0.0533	mg/l	4/2/2024 14:28:13	
0.1PPM	A	Phenolics-	P	0.1055	mg/l	4/2/2024 14:28:14	
0.25PPM	A	Phenolics-	P	0.2549	mg/l	4/2/2024 14:28:15	
0.50PPM	A	Phenolics-	P	0.5037	mg/l	4/2/2024 14:28:16	
1.0PPM	A	Phenolics-	P	0.9684	mg/l	4/2/2024 14:28:17	
2.0PPM	A	Phenolics-	P	2.0139	mg/l	4/2/2024 14:28:18	
ICV1	S	Phenolics-	P	0.991	mg/l	4/2/2024 15:01:44	
ICB1	S	Phenolics-	P	0.0027	mg/l	4/2/2024 15:01:46	
CCV1	S	Phenolics-	P	1.0003	mg/l	4/2/2024 15:01:47	
CCB1	S	Phenolics-	P	0.0007	mg/l	4/2/2024 15:01:50	
PB159952BL	S	Phenolics-	P	-0.0094	mg/l	4/2/2024 15:01:51	
PB159952BS	S	Phenolics-	P	0.9985	mg/l	4/2/2024 15:01:53	
P1911-25	S	Phenolics-	P	0.4241	mg/l	4/2/2024 15:09:26	
CCV2	S	Phenolics-	P	1.0041	mg/l	4/2/2024 15:09:29	
CCB2	S	Phenolics-	P	0.0023	mg/l	4/2/2024 15:09:32	

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

4/2/2024 14:28

Reviewed by : RC

Instrument ID : Konelab

Test Phenolics-


Accepted 4/2/2024 14:28

Factor 9.141

Bias 0.008

Coeff. of det. 0.999598

Errors



	Calibrator	Response	Calc. con.	Conc.	<u>RC</u> Errors
1	0.00PPM	0.008	0.0002	0.0000	-
2	Phenol-2	0.013	0.0533	0.0500	6.6
3	Phenol-2	0.019	0.1055	0.1000	5.5
4	Phenol-2	0.036	0.2549	0.2500	2.0
5	Phenol-2	0.063	0.5037	0.5000	0.7
6	Phenol-2	0.114	0.9684	1.0000	-3.2
7	Phenol-2	0.228	2.0139	2.0000	0.7

04/02/2024
RM (w)