

6131938

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

8/10/2024 12:39

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.968	0.0	0.172	
ICB1	0.002	0.0	0.014	
CCV1	0.980	0.0	0.174	
CCB1	0.006	0.0	0.015	
RL CHECK	0.091	0.0	0.028	
PB162567BL	0.004	0.0	0.014	
PB162567BS	0.975	0.0	0.173	
P3490-01	0.005	0.0	0.014	
P3490-01DUP	0.005	0.0	0.014	
P3490-01MS	1.030	0.0	0.182	
P3490-01MSD	1.022	0.0	0.181	
P3510-01	1.778	0.0	0.304	
P3510-03	0.296	0.0	0.062	
P3510-05	1.384	0.0	0.240	
CCV2	0.940	0.0	0.167	
CCB2	0.023	0.0	0.018	
P3536-04	1.496	0.0	0.258	
P3546-01	3.592	0.0	0.600	
P3546-02	3.668	0.0	0.612	Test limit high
PB162523BL	0.005	0.0	0.015	Test limit high
PB162523BS	0.981	0.0	0.174	
P3390-03	0.080	0.0	0.027	
CCV3	0.965	0.0	0.171	
CCB3	0.006	0.0	0.015	
P3546-01DLX2	1.843	0.0	0.314	
P3546-02DLX2	1.981	0.0	0.337	
CCV4	0.953	0.0	0.169	
CCB4	0.003	0.0	0.014	
N	28			
Mean	0.896			
SD	1.0049			
CV%	112.18			

91% (50-150)
8/10/2024
RM

Aquakem v. 7.2AQ1

Results from time period:

Sat Aug 10 10:05:59 2024

Sat Aug 10 12:31:00 2024

Sample Id	Sam/Ctr/c#	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-† P		0.008	mg/l	8/10/2024 10:58:43	
0.1PPM	A	Ammonia-† P		0.0967	mg/l	8/10/2024 10:58:44	
0.2PPM	A	Ammonia-† P		0.1936	mg/l	8/10/2024 10:58:45	
0.4PPM	A	Ammonia-† P		0.3987	mg/l	8/10/2024 10:58:46	
1.0PPM	A	Ammonia-† P		1.0053	mg/l	8/10/2024 10:58:47	
1.3PPM	A	Ammonia-† P		1.3313	mg/l	8/10/2024 10:58:48	
2.0PPM	A	Ammonia-† P		1.9998	mg/l	8/10/2024 10:58:49	
ICV1	S	Ammonia-† P		0.968	mg/l	8/10/2024 11:39:01	
ICB1	S	Ammonia-† P		0.0022	mg/l	8/10/2024 11:39:03	
CCV1	S	Ammonia-† P		0.98	mg/l	8/10/2024 11:39:06	
CCB1	S	Ammonia-† P		0.0059	mg/l	8/10/2024 11:39:07	
RL CHECK	S	Ammonia-† P		0.0908	mg/l	8/10/2024 11:39:10	
PB162567BL	S	Ammonia-† P		0.0038	mg/l	8/10/2024 11:39:12	
PB162567BS	S	Ammonia-† P		0.9749	mg/l	8/10/2024 11:49:45	
P3490-01	S	Ammonia-† P		0.0046	mg/l	8/10/2024 11:49:47	
P3490-01DUP	S	Ammonia-† P		0.0048	mg/l	8/10/2024 11:49:50	
P3490-01MS	S	Ammonia-† P		1.0303	mg/l	8/10/2024 11:49:51	
P3490-01MSD	S	Ammonia-† P		1.0225	mg/l	8/10/2024 11:49:52	
P3510-01	S	Ammonia-† P		1.7776	mg/l	8/10/2024 11:49:53	
P3510-03	S	Ammonia-† P		0.296	mg/l	8/10/2024 11:49:54	
P3510-05	S	Ammonia-† P		1.3839	mg/l	8/10/2024 11:49:55	
CCV2	S	Ammonia-† P		0.9397	mg/l	8/10/2024 11:49:56	
CCB2	S	Ammonia-† P		0.0234	mg/l	8/10/2024 12:00:27	
P3536-04	S	Ammonia-† P		1.4958	mg/l	8/10/2024 12:00:29	
P3546-01	S	Ammonia-† P		3.5924	mg/l	8/10/2024 12:00:30	
P3546-02	S	Ammonia-† P		3.6682	mg/l	8/10/2024 12:00:31	
PB162523BL	S	Ammonia-† P		0.0051	mg/l	8/10/2024 12:00:33	
PB162523BS	S	Ammonia-† P		0.9814	mg/l	8/10/2024 12:00:34	
P3390-03	S	Ammonia-† P		0.0797	mg/l	8/10/2024 12:00:36	
CCV3	S	Ammonia-† P		0.9651	mg/l	8/10/2024 12:08:19	
CCB3	S	Ammonia-† P		0.0064	mg/l	8/10/2024 12:08:20	
P3546-01DLX2	S	Ammonia-† P		1.8426	mg/l	8/10/2024 12:30:54	
P3546-02DLX2	S	Ammonia-† P		1.9809	mg/l	8/10/2024 12:30:56	
CCV4	S	Ammonia-† P		0.9532	mg/l	8/10/2024 12:30:57	
CCB4	S	Ammonia-† P		0.0032	mg/l	8/10/2024 12:31:00	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

8/10/2024 10:59


Test Ammonia-N

Accepted 8/10/2024 10:59

Factor 6.127
Bias 0.014

Coeff. of det. 0.999956

Errors



	Calibrator	Response	Calc. con.	Conc.	^{RM} Errors
1	0.00PPM	0.015	0.0080	0.0000	—
2	NH3-2PPM	0.029	0.0967	0.1000	-3.3
3	NH3-2PPM	0.045	0.1936	0.2000	-3.2
4	NH3-2PPM	0.079	0.3987	0.4000	-6.3
5	NH3-2PPM	0.178	1.0053	1.0000	0.5
6	NH3-2PPM	0.231	1.3313	1.3333	2.4
7	NH3-2PPM	0.340	1.9998	2.0000	0.0

08/10/2024
RM