

66132111

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

8/21/2024 13:11

Reviewed by : RM

Instrument ID : Konelab

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.964	0.0	0.175	
ICB1	0.010	0.0	0.016	
CCV1	0.926	0.0	0.169	
CCB1	0.008	0.0	0.015	
RL CHECK	0.091	0.0	0.029	
PB162845BL	0.007	0.0	0.015	
PB162845BS	0.998	0.0	0.181	
P3656-01	0.019	0.0	0.017	
P3656-01DUP	0.021	0.0	0.017	
P3656-01MS	1.018	0.0	0.185	
P3656-01MSD	0.984	0.0	0.179	
P3656-03	2.094	0.0	0.365	Test limit high
P3656-05	2.054	0.0	0.358	Test limit high
P3656-07	11.952	0.0	2.017	Test limit high
CCV2	0.957	0.0	0.174	
CCB2	0.011	0.0	0.016	
P3656-03DLX2	1.005	0.0	0.182	
P3656-05DLX2	1.033	0.0	0.187	
P3656-07DLX10	1.175	0.0	0.211	
CCV3	0.946	0.0	0.172	
CCB3	0.011	0.0	0.016	
N	21			
Mean	1.252			
SD	2.5359			
CV%	202.62			

91% (50-150)

08/21/2024

RM

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 21 11:53:47 2024

Wed Aug 21 13:04:03 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-† P		0.0091	mg/l	8/21/2024 11:53:47	
0.1PPM	A	Ammonia-† P		0.1092	mg/l	8/21/2024 11:53:48	
0.2PPM	A	Ammonia-† P		0.1948	mg/l	8/21/2024 11:53:49	
0.4PPM	A	Ammonia-† P		0.398	mg/l	8/21/2024 11:53:50	
1.0PPM	A	Ammonia-† P		0.9832	mg/l	8/21/2024 11:53:51	
1.3PPM	A	Ammonia-† P		1.3239	mg/l	8/21/2024 11:53:52	
2.0PPM	A	Ammonia-† P		2.0151	mg/l	8/21/2024 11:53:53	
ICV1	S	Ammonia-† P		0.9636	mg/l	8/21/2024 12:25:35	
ICB1	S	Ammonia-† P		0.0102	mg/l	8/21/2024 12:25:37	
CCV1	S	Ammonia-† P		0.9265	mg/l	8/21/2024 12:25:38	
CCB1	S	Ammonia-† P		0.008	mg/l	8/21/2024 12:25:40	
RL CHECK	S	Ammonia-† P		0.0906	mg/l	8/21/2024 12:25:43	
PB162845BL	S	Ammonia-† P		0.0075	mg/l	8/21/2024 12:25:44	
PB162845BS	S	Ammonia-† P		0.9981	mg/l	8/21/2024 12:36:17	
P3656-01	S	Ammonia-† P		0.0187	mg/l	8/21/2024 12:36:18	
P3656-01DUP	S	Ammonia-† P		0.0207	mg/l	8/21/2024 12:36:20	
P3656-01MS	S	Ammonia-† P		1.0181	mg/l	8/21/2024 12:36:22	
P3656-01MSD	S	Ammonia-† P		0.9837	mg/l	8/21/2024 12:36:23	
P3656-03	S	Ammonia-† P		2.0939	mg/l	8/21/2024 12:36:24	
P3656-05	S	Ammonia-† P		2.0544	mg/l	8/21/2024 12:36:25	
P3656-07	S	Ammonia-† P		11.9517	mg/l	8/21/2024 12:36:26	
CCV2	S	Ammonia-† P		0.9566	mg/l	8/21/2024 12:36:27	
CCB2	S	Ammonia-† P		0.0105	mg/l	8/21/2024 12:41:46	
P3656-03DLX2	S	Ammonia-† P		1.0045	mg/l	8/21/2024 13:03:55	
P3656-05DLX2	S	Ammonia-† P		1.0329	mg/l	8/21/2024 13:03:57	
P3656-07DLX10	S	Ammonia-† P		1.1746	mg/l	8/21/2024 13:03:58	
CCV3	S	Ammonia-† P		0.9459	mg/l	8/21/2024 13:04:01	
CCB3	S	Ammonia-† P		0.0115	mg/l	8/21/2024 13:04:03	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

8/21/2024 11:54

Reviewed by : RM

Instrument ID : Konelab

Test Ammonia-N


Accepted 8/21/2024 11:54

Factor 5.967

Bias 0.014

Coeff. of det. 0.999763

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.015	0.0091	0.0000	-
2	NH3-2PPM	0.032	0.1092	0.1000	9.2
3	NH3-2PPM	0.047	0.1948	0.2000	-2.6
4	NH3-2PPM	0.081	0.3980	0.4000	-0.5
5	NH3-2PPM	0.179	0.9832	1.0000	-1.7
6	NH3-2PPM	0.236	1.3239	1.3333	1.0
7	NH3-2PPM	0.352	2.0151	2.0000	0.0

08/21/2024
RM