

66132256

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

9/3/2024 15:18

Reviewed by : RM Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.028	0.0	0.116	
ICB1	0.003	0.0	0.008	
CCV1	0.981	0.0	0.111	
CCB1	0.006	0.0	0.009	
PB162894BL	-0.001	0.0	0.008	
PB162894BS	1.023	0.0	0.116	
P3390-12	0.023	0.0	0.010	
CCV2	0.964	0.0	0.109	
CCB2	-0.006	0.0	0.007	

N	9
Mean	0.447
SD	0.5244
CV%	117.38

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 03 14:27:55 2024

Tue Sep 03 15:07:21 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0025	mg/l	9/3/2024 14:27:55	
0.05PPM	A	Phenolics-	P	0.0509	mg/l	9/3/2024 14:27:56	
0.1PPM	A	Phenolics-	P	0.1038	mg/l	9/3/2024 14:27:57	
0.25PPM	A	Phenolics-	P	0.2504	mg/l	9/3/2024 14:27:58	
0.50PPM	A	Phenolics-	P	0.5002	mg/l	9/3/2024 14:27:59	
1.0PPM	A	Phenolics-	P	0.9948	mg/l	9/3/2024 14:28:00	
2.0PPM	A	Phenolics-	P	2.0023	mg/l	9/3/2024 14:28:01	
ICV1	S	Phenolics-	P	1.0281	mg/l	9/3/2024 15:00:58	
ICB1	S	Phenolics-	P	0.0028	mg/l	9/3/2024 15:00:59	
CCV1	S	Phenolics-	P	0.9812	mg/l	9/3/2024 15:01:02	
CCB1	S	Phenolics-	P	0.006	mg/l	9/3/2024 15:01:03	
PB162894BL	S	Phenolics-	P	-0.0009	mg/l	9/3/2024 15:01:05	
PB162894BS	S	Phenolics-	P	1.0229	mg/l	9/3/2024 15:01:07	
P3390-12	S	Phenolics-	P	0.0228	mg/l	9/3/2024 15:07:15	
CCV2	S	Phenolics-	P	0.9644	mg/l	9/3/2024 15:07:18	
CCB2	S	Phenolics-	P	-0.0061	mg/l	9/3/2024 15:07:21	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

9/3/2024 14:33


Test Phenolics-

Accepted 9/3/2024 14:33

Factor 9.511
Bias 0.008

Coeff. of det. 0.999983

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	-0.0025	0.0000	-
2	Phenol-2	0.013	0.0509	0.0500	1.8
3	Phenol-2	0.019	0.1038	0.1000	3.8
4	Phenol-2	0.034	0.2504	0.2500	0.2
5	Phenol-2	0.061	0.5002	0.5000	0.0
6	Phenol-2	0.113	0.9948	1.0000	-0.5
7	Phenol-2	0.219	2.0023	2.0000	0.1

09/03/2024
RM