

L6132855

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/10/2024 9:43

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.366	0.0	0.075	
ICB1	-0.375	0.0	0.002	
CCV1	238.703	0.0	0.184	
CCB1	-0.128	0.0	0.002	
PB163993BL	-0.484	0.0	0.002	
P4353-06	-0.652	0.0	0.001	
P4353-06DUP	-0.757	0.0	0.001	
P4359-02	-1.078	0.0	0.001	
PB164012BL	-1.083	0.0	0.001	
P4372-01	-0.603	0.0	0.001	
P4372-01DUP	-0.432	0.0	0.002	
P4372-02	-0.505	0.0	0.002	
CCV2	256.525	0.0	0.197	
CCB2	-0.196	0.0	0.002	

N	14
Mean	41.807
SD	90.9800
CV%	217.62

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/10/2024 9:04


Test Total CN

Accepted 10/10/2024 9:04

Factor 1312
Bias 0.002

Coeff. of det. 0.999898

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-1.1125	0.0000	-
2	5.0PPBCN	0.005	4.3215	5.0000	-13.6
3	10PPBCN	0.009	9.1050	10.0000	-9.0
4	50PPBCN	0.040	49.8212	50.0000	-0.4
5	100PPBCN	0.079	101.2924	100.0000	1.3
6	250PPBCN	0.195	253.5768	250.0000	1.4
7	500PPBCN	0.381	497.9957	500.0000	-0.4

10/10/2024

RM

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 10 09:03:39 2024

Thu Oct 10 09:42:28 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.1125	µg/l	10/10/2024 9:03:39	
5.0PPBCN	A	Total CN	P	4.3215	µg/l	10/10/2024 9:03:40	
10PPBCN	A	Total CN	P	9.105	µg/l	10/10/2024 9:03:41	
50PPBCN	A	Total CN	P	49.8212	µg/l	10/10/2024 9:03:42	
100PPBCN	A	Total CN	P	101.2924	µg/l	10/10/2024 9:03:43	
250PPBCN	A	Total CN	P	253.5768	µg/l	10/10/2024 9:03:44	
500PPBCN	A	Total CN	P	497.9957	µg/l	10/10/2024 9:03:45	
ICV1	S	Total CN	P	96.3665	µg/l	10/10/2024 9:30:06	
ICB1	S	Total CN	P	-0.3752	µg/l	10/10/2024 9:30:08	
CCV1	S	Total CN	P	238.7032	µg/l	10/10/2024 9:30:09	
CCB1	S	Total CN	P	-0.1285	µg/l	10/10/2024 9:30:11	
PB163993BL	S	Total CN	P	-0.484	µg/l	10/10/2024 9:30:14	
P4353-06	S	Total CN	P	-0.6522	µg/l	10/10/2024 9:30:15	
P4353-06DUP	S	Total CN	P	-0.757	µg/l	10/10/2024 9:37:39	
P4359-02	S	Total CN	P	-1.0785	µg/l	10/10/2024 9:37:40	
PB164012BL	S	Total CN	P	-1.0826	µg/l	10/10/2024 9:37:43	
P4372-01	S	Total CN	P	-0.6034	µg/l	10/10/2024 9:37:44	
P4372-01DUP	S	Total CN	P	-0.4317	µg/l	10/10/2024 9:37:45	
P4372-02	S	Total CN	P	-0.5048	µg/l	10/10/2024 9:37:46	
CCV2	S	Total CN	P	256.5251	µg/l	10/10/2024 9:42:25	
CCB2	S	Total CN	P	-0.1958	µg/l	10/10/2024 9:42:27	