

15136003

=====

Test results	Aquakem 7.2AQ1	Page: 1
--------------	----------------	---------

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

6/4/2025 14:20 Reviewed by : RM Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.005	0.0	0.069	
ICB1	0.002	0.0	0.008	
CCV1	1.004	0.0	0.069	
CCB1	-0.002	0.0	0.008	
PB168281BL	0.001	0.0	0.008	
PB168281BS	1.004	0.0	0.069	
Q2126-01	0.033	0.0	0.010	
Q2126-02	0.059	0.0	0.011	
PB168283BL	0.005	0.0	0.008	
PB168283BS	1.011	0.0	0.069	
Q2126-07	0.030	0.0	0.010	
Q2126-08	0.055	0.0	0.011	
CCV2	1.011	0.0	0.069	
CCB2	0.002	0.0	0.008	

N	14
Mean	0.373
SD	0.4908
CV%	131.64

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 04 12:40:31 2025

Wed Jun 04 14:11:07 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0015	mg/l	6/4/2025 12:40:31	
0.05PPM	A	Phenolics-	P	0.0546	mg/l	6/4/2025 12:40:32	
0.1PPM	A	Phenolics-	P	0.1033	mg/l	6/4/2025 12:40:33	
0.25PPM	A	Phenolics-	P	0.2557	mg/l	6/4/2025 12:40:34	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	6/4/2025 12:40:35	
1.0PPM	A	Phenolics-	P	0.9746	mg/l	6/4/2025 12:40:36	
2.0PPM	A	Phenolics-	P	2.0122	mg/l	6/4/2025 12:40:37	
ICV1	S	Phenolics-	P	1.0045	mg/l	6/4/2025 13:56:09	
ICB1	S	Phenolics-	P	0.0017	mg/l	6/4/2025 13:56:11	
CCV1	S	Phenolics-	P	1.0038	mg/l	6/4/2025 13:56:13	
CCB1	S	Phenolics-	P	-0.0019	mg/l	6/4/2025 13:56:14	
PB168281BL	S	Phenolics-	P	0.0013	mg/l	6/4/2025 13:56:16	
PB168281BS	S	Phenolics-	P	1.0035	mg/l	6/4/2025 13:56:19	
Q2126-01	S	Phenolics-	P	0.0334	mg/l	6/4/2025 14:06:11	
Q2126-02	S	Phenolics-	P	0.0587	mg/l	6/4/2025 14:06:13	
PB168283BL	S	Phenolics-	P	0.0053	mg/l	6/4/2025 14:06:15	
PB168283BS	S	Phenolics-	P	1.0109	mg/l	6/4/2025 14:06:17	
Q2126-07	S	Phenolics-	P	0.0298	mg/l	6/4/2025 14:06:18	
Q2126-08	S	Phenolics-	P	0.0551	mg/l	6/4/2025 14:06:20	
CCV2	S	Phenolics-	P	1.0112	mg/l	6/4/2025 14:11:05	
CCB2	S	Phenolics-	P	0.0018	mg/l	6/4/2025 14:11:06	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

6/4/2025 12:43


Test Phenolics-

Accepted 6/4/2025 12:43

Factor 16.51
Bias 0.008

Coeff. of det. 0.999727

Errors



	Calibrator	Response	Calc. con.	Conc.	^{Re} Errors
1	0.00PPM	0.008	0.0015	0.0000	-
2	Phenol-2	0.011	0.0546	0.0500	9.2
3	Phenol-2	0.014	0.1033	0.1000	3.3
4	Phenol-2	0.023	0.2557	0.2500	2.3
5	Phenol-2	0.038	0.4981	0.5000	-0.4
6	Phenol-2	0.067	0.9746	1.0000	-2.5
7	Phenol-2	0.130	2.0122	2.0000	0.6

06/04/2025
RM