

Cb136004

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

6/4/2025 15:13

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.003	0.0	0.069	
ICB1	0.005	0.0	0.008	
CCV1	1.010	0.0	0.069	
CCB1	0.003	0.0	0.008	
PB168282BL	0.003	0.0	0.008	
PB168282BS	1.010	0.0	0.069	
Q2126-10	0.029	0.0	0.010	
Q2126-11	0.056	0.0	0.011	
CCV2	1.007	0.0	0.069	
CCB2	0.002	0.0	0.008	

N	10
Mean	0.413
SD	0.5122
CV%	124.04

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 04 14:41:45 2025

Wed Jun 04 14:48:38 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0015	mg/l	6/4/2025 12:40:31	
0.05PPM	A	Phenolics-	P	0.0546	mg/l	6/4/2025 12:40:32	
0.1PPM	A	Phenolics-	P	0.1033	mg/l	6/4/2025 12:40:33	
0.25PPM	A	Phenolics-	P	0.2557	mg/l	6/4/2025 12:40:34	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	6/4/2025 12:40:35	
1.0PPM	A	Phenolics-	P	0.9746	mg/l	6/4/2025 12:40:36	
2.0PPM	A	Phenolics-	P	2.0122	mg/l	6/4/2025 12:40:37	
ICV1	S	Phenolics-	P	1.003	mg/l	6/4/2025 14:41:46	
ICB1	S	Phenolics-	P	0.0046	mg/l	6/4/2025 14:41:48	
CCV1	S	Phenolics-	P	1.0104	mg/l	6/4/2025 14:41:49	
CCB1	S	Phenolics-	P	0.0029	mg/l	6/4/2025 14:41:52	
PB168282BL	S	Phenolics-	P	0.0028	mg/l	6/4/2025 14:41:54	
PB168282BS	S	Phenolics-	P	1.0103	mg/l	6/4/2025 14:41:55	
Q2126-10	S	Phenolics-	P	0.0294	mg/l	6/4/2025 14:48:32	
Q2126-11	S	Phenolics-	P	0.0561	mg/l	6/4/2025 14:48:34	
CCV2	S	Phenolics-	P	1.0072	mg/l	6/4/2025 14:48:36	
CCB2	S	Phenolics-	P	0.0025	mg/l	6/4/2025 14:48:38	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

6/4/2025 12:43

Test Phenolics-


Accepted 6/4/2025 12:43

Factor 16.51

Bias 0.008

Coeff. of det. 0.999727

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	0.0015	0.0000	-
2	Phenol-2	0.011	0.0546	0.0500	9.2
3	Phenol-2	0.014	0.1033	0.1000	3.3
4	Phenol-2	0.023	0.2557	0.2500	2.3
5	Phenol-2	0.038	0.4981	0.5000	-0.4
6	Phenol-2	0.067	0.9746	1.0000	-2.5
7	Phenol-2	0.130	2.0122	2.0000	0.6

06/04/2025
RM