

LB136246


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43953



A= 0.0000e+000

B= 2.3148e-004

C= 4.3351e-002

Rho= 0.9988639

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/C/D
0.0	0.000	0.055	0.055	49	0.000	49	0				
0.2	0.200	0.219	0.019	759	0.0 %	759					
2.5	2.500	2.502	0.002	10622	0.0 %	10622					
5.0	5.000	5.068	0.068	21706	0.0 %	21706					
7.5	7.500	7.139	-0.361	30654	0.0 %	30654					
10.0	10.000	10.217	0.217	43953	0.0 %	43953					

LB136246 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jun 2025 14:21:55

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	49	-	0.0000	7470A	PPB	24 Jun 2025 14:29:03	S	Std
0.2 - 1	Fi 2	759	-	0.2000	7470A	PPB	24 Jun 2025 14:31:21	S	Std
2.5 - 1	S2.5	10622	-	2.5000	7470A	PPB	24 Jun 2025 14:33:38	S	Std
5.0 - 1	S5	21706	-	5.0000	7470A	PPB	24 Jun 2025 14:38:32	S	Std
7.5 - 1	S7.5	30654	-	7.5000	7470A	PPB	24 Jun 2025 14:40:48	S	Std
10.0 - 1	S10	43953	-	10.0000	7470A	PPB	24 Jun 2025 14:47:28	S	Std
ICV03 - 1	ICV03	17071	3.9949	-	7470A	PPB	24 Jun 2025 14:53:57	U	SMPL
ICB03 - 1	ICB03	-186	0.0003	-	7470A	PPB	24 Jun 2025 14:56:12	U	SMPL
CCV08 - 1	CCV08	20929	4.8879	-	7470A	PPB	24 Jun 2025 14:58:29	U	SMPL
CCB08 - 1	CCB08	-186	0.0003	-	7470A	PPB	24 Jun 2025 15:00:44	U	SMPL
CRA - 1	CRA	646	0.1929	-	7470A	PPB	24 Jun 2025 15:03:02	U	SMPL
HighStd - 1	HighStd	41811	9.7216	-	7470A	PPB	24 Jun 2025 15:05:17	U	SMPL
ChkStd - 1	ChkStd	26890	6.2677	-	7470A	PPB	24 Jun 2025 15:11:52	U	SMPL
PB168602BL - 1	PBW	-247	-0.0138	-	7470A	PPB	24 Jun 2025 15:14:07	U	SMPL
PB168602BS - 1	LCSW	14442	3.3863	-	7470A	PPB	24 Jun 2025 15:16:26	U	SMPL
Q2126-09 - 1	MDL-WATER-03-QT2-2025	28	0.0498	-	7470A	PPB	24 Jun 2025 15:18:42	U	SMPL
Q2382-01 - 1	FMI109	738	0.2142	-	7470A	PPB	24 Jun 2025 15:21:00	U	SMPL
Q2382-01DUP - 1	FMI109DUP	714	0.2086	-	7470A	PPB	24 Jun 2025 15:23:17	U	SMPL
Q2382-01MS - 1	FMI109MS	8975	2.1208	-	7470A	PPB	24 Jun 2025 15:25:34	U	SMPL
Q2382-01MSD - 1	FMI109MSD	10446	2.4613	-	7470A	PPB	24 Jun 2025 15:27:49	U	SMPL
CCV09 - 1	CCV09	19475	4.5513	-	7470A	PPB	24 Jun 2025 15:35:38	U	SMPL
CCB09 - 1	CCB09	-316	-0.0298	-	7470A	PPB	24 Jun 2025 15:37:52	U	SMPL
Q2126-07 - 1		75	0.0607	-	7470A	PPB	24 Jun 2025 15:47:52	U	SMPL
Q2126-07RE - 1		520	0.1637	-	7470A	PPB	24 Jun 2025 15:50:07	U	SMPL
Q2382-01LX5 - 1		138	0.0753	-	7470A	PPB	24 Jun 2025 15:52:23	U	SMPL
Q2382-01A - 1		11070	2.6058	-	7470A	PPB	24 Jun 2025 15:54:39	U	SMPL
CCV10 - 1	CCV10	21223	4.9560	-	7470A	PPB	24 Jun 2025 16:05:09	U	SMPL
CCB10 - 1	CCB10	-273	-0.0198	-	7470A	PPB	24 Jun 2025 16:07:25	U	SMPL