

LB136298


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41311



A= 0.0000e+000

B= 2.3349e-004

C= 3.4379e-002

Rho= 0.9962468

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	34	0.000	34				
0.2	0.200	0.216	0.016	779	0.0 %	779				
2.5	2.500	2.575	0.075	10880	0.0 %	10880				
5.0	5.000	4.601	-0.399	19558	0.0 %	19558				
7.5	7.500	8.086	0.586	34482	0.0 %	34482				
10.0	10.000	9.680	-0.320	41311	0.0 %	41311				

lb136298 instrument id : cv1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jun 2025 14:10:54

Sample ID	Extended ID	μ Abs.	Conc.	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	26 Jun 2025 14:11:09	S	Std
0.2 - 1	50.2	779	-	0.2000	7471B	PPB	26 Jun 2025 14:13:26	S	Std
2.5 - 1	52.5	10880	-	2.5000	7471B	PPB	26 Jun 2025 14:15:43	S	Std
5.0 - 1	55	19558	-	5.0000	7471B	PPB	26 Jun 2025 14:18:00	S	Std
7.5 - 1	57.5	34482	-	7.5000	7471B	PPB	26 Jun 2025 14:20:19	S	Std
10.0 - 1	59	41311	-	10.0000	7471B	PPB	26 Jun 2025 14:31:44	S	Std
ICV08 - 1	ICV08	16631	3.9176	-	7471B	PPB	26 Jun 2025 14:34:57	U	SMPL
ICB08 - 1	ICB08	-216	-0.0161	-	7471B	PPB	26 Jun 2025 14:37:11	U	SMPL
CCV23 - 1	CCV23	21009	4.9398	-	7471B	PPB	26 Jun 2025 14:39:29	U	SMPL
CCB23 - 1	CCB23	-313	-0.0387	-	7471B	PPB	26 Jun 2025 14:41:44	U	SMPL
CRA - 1	CRA	755	0.2107	-	7471B	PPB	26 Jun 2025 14:44:03	U	SMPL
HighStd - 1	HighStd	40954	9.5967	-	7471B	PPB	26 Jun 2025 14:46:18	U	SMPL
ChkStd - 1	ChkStd	32077	7.5240	-	7471B	PPB	26 Jun 2025 14:48:33	U	SMPL
PB168633BL - 1	PBS	-213	-0.0154	-	7471B	PPB	26 Jun 2025 14:51:01	U	SMPL
PB168633BS - 1	LCS	16163	3.8083	-	7471B	PPB	26 Jun 2025 14:58:26	U	SMPL
Q2412-01 - 1	TR-05-062425	2652	0.6536	-	7471B	PPB	26 Jun 2025 15:00:42	U	SMPL
Q2413-01 - 1	TP-35	1289	0.3353	-	7471B	PPB	26 Jun 2025 15:03:01	U	SMPL
Q2413-02 - 1	TP-34	1290	0.3356	-	7471B	PPB	26 Jun 2025 15:05:19	U	SMPL
Q2413-03 - 1	TP-77	871	0.2377	-	7471B	PPB	26 Jun 2025 15:07:36	U	SMPL
Q2413-04 - 1	TP-74	529	0.1579	-	7471B	PPB	26 Jun 2025 15:09:52	U	SMPL
CCV24 - 1	CCV24	20356	4.7873	-	7471B	PPB	26 Jun 2025 15:12:07	U	SMPL
CCB24 - 1	CCB24	-174	-0.0062	-	7471B	PPB	26 Jun 2025 15:14:22	U	SMPL
Q2413-05 - 1	TP-73	1270	0.3309	-	7471B	PPB	26 Jun 2025 15:16:41	U	SMPL
Q2413-06 - 1	TP-72	1009	0.2700	-	7471B	PPB	26 Jun 2025 15:18:57	U	SMPL
Q2414-01 - 1	WC-1	1609	0.4101	-	7471B	PPB	26 Jun 2025 15:21:12	U	SMPL
Q2415-01 - 1	WC-1	2438	0.6036	-	7471B	PPB	26 Jun 2025 15:23:27	U	SMPL
Q2416-01 - 1	MH-G/H	112	0.0605	-	7471B	PPB	26 Jun 2025 15:25:43	U	SMPL
Q2419-01 - 1	EO-03-06252025	148636	34.7394	-	7471B	PPB	26 Jun 2025 15:28:00	U	SMPL
Q2420-01 - 1	72-11933	585	0.1710	-	7471B	PPB	26 Jun 2025 15:30:17	U	SMPL
Q2423-01 - 1	RT915	4231	1.0223	-	7471B	PPB	26 Jun 2025 15:33:07	U	SMPL
Q2425-01 - 1	COMP-1	124631	29.1345	-	7471B	PPB	26 Jun 2025 15:35:25	U	SMPL
Q2425-02 - 1	COMP-2	126663	29.6090	-	7471B	PPB	26 Jun 2025 15:37:43	U	SMPL
CCV25 - 1	CCV25	19923	4.6862	-	7471B	PPB	26 Jun 2025 15:40:27	U	SMPL
CCB25 - 1	CCB25	-238	-0.0212	-	7471B	PPB	26 Jun 2025 15:43:23	U	SMPL
Q2425-03 - 1	COMP-3	5394	1.2938	-	7471B	PPB	26 Jun 2025 15:45:42	U	SMPL
Q2425-03DUP - 1	COMP-3DUP	5441	1.3048	-	7471B	PPB	26 Jun 2025 15:47:58	U	SMPL
Q2425-03MS - 1	COMP-3MS	23716	5.5718	-	7471B	PPB	26 Jun 2025 15:54:27	U	SMPL
Q2425-03MSD - 1	COMP-3MSD	23844	5.6017	-	7471B	PPB	26 Jun 2025 15:56:43	U	SMPL
Q2425-03LX5 - 1		857	0.2345	-	7471B	PPB	26 Jun 2025 15:59:04	U	SMPL
Q2425-03A - 1		23346	5.4854	-	7471B	PPB	26 Jun 2025 16:01:25	U	SMPL
Q2419-01DLX10 - 1		13041	3.0793	-	7471B	PPB	26 Jun 2025 16:03:42	U	SMPL
Q2425-01DLX10 - 1		11145	2.6366	-	7471B	PPB	26 Jun 2025 16:06:03	U	SMPL
Q2425-02DLX10 - 1		11519	2.7240	-	7471B	PPB	26 Jun 2025 16:11:22	U	SMPL
CCV26 - 1	CCV26	20243	4.7609	-	7471B	PPB	26 Jun 2025 16:20:08	U	SMPL
CCB26 - 1	CCB26	-230	-0.0193	-	7471B	PPB	26 Jun 2025 16:22:26	U	SMPL