

16137261

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

9/22/2025 13:24

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.991	0.0	0.206	
ICB1	0.017	0.0	0.017	
CCV1	0.999	0.0	0.208	
CCB1	0.014	0.0	0.016	
RL CHECK	0.090	0.0	0.031	
PB169778BL	0.015	0.0	0.017	
PB169778BS	0.972	0.0	0.202	
Q3155-01	0.024	0.0	0.018	
Q3155-01DUP	0.026	0.0	0.019	
Q3155-01MS	1.048	0.0	0.217	
Q3155-01MSD	1.021	0.0	0.212	
Q3155-02	0.009	0.0	0.015	
Q3155-03	0.023	0.0	0.018	
RL CHECK-	0.094	0.0	0.032	
CCV2	1.005	0.0	0.209	
CCB2	0.019	0.0	0.017	
PB169780BL	0.017	0.0	0.017	
PB169780BS	1.003	0.0	0.208	
Q3135-01	0.321	0.0	0.076	
Q3151-02	0.038	0.0	0.021	
Q3151-02DUP	0.039	0.0	0.021	
Q3151-02MS	1.044	0.0	0.217	
Q3151-02MSD	1.049	0.0	0.217	
CCV3	1.000	0.0	0.208	
CCB3	0.019	0.0	0.017	

90% (50-150) 09/22/2025 RM

94% (50-150) 09/22/2025 RM

N 25
Mean 0.436
SD 0.4851
CV% 111.28

Aquakem v. 7.2AQ1

Results from time period:

Mon Sep 22 10:37:26 2025

Mon Sep 22 13:21:27 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-† P		0.018	mg/l	9/22/2025 10:37:26	
0.1PPM	A	Ammonia-† P		0.1097	mg/l	9/22/2025 10:37:27	
0.2PPM	A	Ammonia-† P		0.2066	mg/l	9/22/2025 10:37:28	
0.4PPM	A	Ammonia-† P		0.3912	mg/l	9/22/2025 10:37:29	
1.0PPM	A	Ammonia-† P		0.9524	mg/l	9/22/2025 10:37:30	
1.3PPM	A	Ammonia-† P		1.3265	mg/l	9/22/2025 10:37:31	
2.0PPM	A	Ammonia-† P		2.029	mg/l	9/22/2025 10:37:32	
ICV1	S	Ammonia-† P		0.991	mg/l	9/22/2025 12:21:12	
ICB1	S	Ammonia-† P		0.0169	mg/l	9/22/2025 12:21:15	
CCV1	S	Ammonia-† P		0.9989	mg/l	9/22/2025 12:21:16	
CCB1	S	Ammonia-† P		0.014	mg/l	9/22/2025 12:21:19	
RL CHECK	S	Ammonia-† P		0.09	mg/l	9/22/2025 12:21:22	
PB169778BL	S	Ammonia-† P		0.0155	mg/l	9/22/2025 12:31:56	
PB169778BS	S	Ammonia-† P		0.9716	mg/l	9/22/2025 12:31:58	
Q3155-01	S	Ammonia-† P		0.0244	mg/l	9/22/2025 12:32:00	
Q3155-01DUP	S	Ammonia-† P		0.0264	mg/l	9/22/2025 12:32:02	
Q3155-01MS	S	Ammonia-† P		1.0477	mg/l	9/22/2025 12:32:04	
Q3155-01MSD	S	Ammonia-† P		1.0213	mg/l	9/22/2025 12:32:06	
Q3155-02	S	Ammonia-† P		0.009	mg/l	9/22/2025 12:42:40	
Q3155-03	S	Ammonia-† P		0.023	mg/l	9/22/2025 12:42:41	
RL CHECK	S	Ammonia-† P		0.0943	mg/l	9/22/2025 12:42:46	
CCV2	S	Ammonia-† P		1.0048	mg/l	9/22/2025 12:42:48	
CCB2	S	Ammonia-† P		0.019	mg/l	9/22/2025 12:42:51	
PB169780BL	S	Ammonia-† P		0.0171	mg/l	9/22/2025 12:53:22	
PB169780BS	S	Ammonia-† P		1.0025	mg/l	9/22/2025 12:53:25	
Q3135-01	S	Ammonia-† P		0.3208	mg/l	9/22/2025 12:53:26	
Q3151-02	S	Ammonia-† P		0.0382	mg/l	9/22/2025 12:53:28	
Q3151-02DUP	S	Ammonia-† P		0.0386	mg/l	9/22/2025 12:53:31	
Q3151-02MS	S	Ammonia-† P		1.0444	mg/l	9/22/2025 13:21:22	
Q3151-02MSD	S	Ammonia-† P		1.0486	mg/l	9/22/2025 13:21:23	
CCV3	S	Ammonia-† P		1.0001	mg/l	9/22/2025 13:21:24	
CCB3	S	Ammonia-† P		0.0186	mg/l	9/22/2025 13:21:27	

Alliance Technical Group
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

9/22/2025 11:01


Test Ammonia-N

Accepted 9/22/2025 11:01

Factor 5.147
Bias 0.014

Coeff. of det. 0.998904

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.017	0.0180	0.0000	-
2	NH3-2PPM	0.035	0.1097	0.1000	9.7
3	NH3-2PPM	0.054	0.2066	0.2000	3.3
4	NH3-2PPM	0.090	0.3912	0.4000	-2.2
5	NH3-2PPM	0.199	0.9524	1.0000	-4.8
6	NH3-2PPM	0.271	1.3265	1.3333	2.6
7	NH3-2PPM	0.408	2.0290	2.0000	1.5

09/22/2025
RM