

7470A


LB137599

INSTRUMENT ID: C41

Linear

 μ Abs.:

42776



A= 0.0000e+000

B= 2.3129e-004

C= -2.2572e-002

Rho= 0.9988979

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-3	0.000	-3	0			
0.2	0.200	0.165	-0.035	811	0.0 %	811	0			
2.5	2.500	2.370	-0.130	10343	0.0 %	10343	0			
5.0	5.000	5.376	0.376	23340	0.0 %	23340				
7.5	7.500	7.442	-0.058	32271	0.0 %	32271				
10.0	10.000	9.871	-0.129	42776	0.0 %	42776				

lb137599 instrument id : cv1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Oct 2025 12:39:01

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std CondMethod	Date
S	0.0 - 1	50	-	-3	PPB	Std	0.00007470A	21 Oct 2025 12:55:28
S	0.2 - 1	50.2	-	811	PPB	Std	0.20007470A	21 Oct 2025 12:57:52
S	2.5 - 1	50.5	-	10343	PPB	Std	2.50007470A	21 Oct 2025 13:00:09
S	5.0 - 1	51	-	23340	PPB	Std	5.00007470A	21 Oct 2025 13:02:26
S	7.5 - 1	51.5	-	32271	PPB	Std	7.50007470A	21 Oct 2025 13:07:33
S	10.0 - 1	51.8	-	42776	PPB	Std	10.00007470A	21 Oct 2025 13:12:47
U	ICV81 - 1	ICV81	3.8377	16690	PPB	SMPL	-7470A	21 Oct 2025 13:18:47
U	ICB81 - 1	ICB81	-0.1401	-508	PPB	SMPL	-7470A	21 Oct 2025 13:21:03
U	CCV65 - 1	CCV65	5.3062	23039	PPB	SMPL	-7470A	21 Oct 2025 13:26:02
U	CCB65 - 1	CCB65	-0.0989	-330	PPB	SMPL	-7470A	21 Oct 2025 13:28:19
U	CRA - 1	CRA	0.2196	1047	PPB	SMPL	-7470A	21 Oct 2025 13:35:03
U	HighStd - 1	HighStd	9.7658	42320	PPB	SMPL	-7470A	21 Oct 2025 13:40:02
U	ChkStd - 1	ChkStd	7.4022	32101	PPB	SMPL	-7470A	21 Oct 2025 13:45:26
U	PB170175BL - 1	PBW	-0.1019	-343	PPB	SMPL	-7470A	21 Oct 2025 13:47:43
U	PB170175BS - 1	LCSW	3.7436	16283	PPB	SMPL	-7470A	21 Oct 2025 13:50:07
U	Q3369-04 - 1	MH-22	-0.0933	-306	PPB	SMPL	-7470A	21 Oct 2025 13:53:00
U	Q3369-08 - 1	MH-21	-0.0265	-17	PPB	SMPL	-7470A	21 Oct 2025 13:55:20
U	Q3374-04 - 1	TP-4-WC	-0.0179	20	PPB	SMPL	-7470A	21 Oct 2025 13:57:37
U	Q3374-08 - 1	TP-5-WC	-0.0191	15	PPB	SMPL	-7470A	21 Oct 2025 13:59:53
U	Q3381-02 - 1	VNJ-222	-0.0133	40	PPB	SMPL	-7470A	21 Oct 2025 14:02:09
U	CCV66 - 1	CCV66	5.3411	23190	PPB	SMPL	-7470A	21 Oct 2025 14:07:18
U	CCB66 - 1	CCB66	-0.0927	-303	PPB	SMPL	-7470A	21 Oct 2025 14:09:33
U	Q3381-04 - 1	VNJ-245	-0.0390	-71	PPB	SMPL	-7470A	21 Oct 2025 14:11:51
U	Q3381-06 - 1	RBR200059	-0.0279	-23	PPB	SMPL	-7470A	21 Oct 2025 14:14:07
U	Q3381-08 - 1	VNJ-255	-0.0226	0	PPB	SMPL	-7470A	21 Oct 2025 14:16:22
U	Q3388-02 - 1	461	-0.0203	10	PPB	SMPL	-7470A	21 Oct 2025 14:18:38
U	Q3390-02 - 1	4175	-0.0182	19	PPB	SMPL	-7470A	21 Oct 2025 14:20:54
U	Q3392-04 - 1	MH-20	-0.0092	58	PPB	SMPL	-7470A	21 Oct 2025 14:23:10
U	Q3392-08 - 1	MH-19	-0.0055	74	PPB	SMPL	-7470A	21 Oct 2025 14:25:26
U	Q3392-08DUP - 1	MH-19DUP	-0.0152	32	PPB	SMPL	-7470A	21 Oct 2025 14:27:43
U	Q3392-08MS - 1	MH-19MS	2.1981	9601	PPB	SMPL	-7470A	21 Oct 2025 14:30:00
U	Q3392-08MSD - 1	MH-19MSD	2.1571	9424	PPB	SMPL	-7470A	21 Oct 2025 14:32:16
U	CCV67 - 1	CCV67	4.5307	19686	PPB	SMPL	-7470A	21 Oct 2025 14:34:34
U	CCB67 - 1	CCB67	-0.0996	-333	PPB	SMPL	-7470A	21 Oct 2025 14:36:53
U	PB170147TB - 1	PB170147TB	-0.0288	-27	PPB	SMPL	-7470A	21 Oct 2025 14:39:12
U	Q3392-08LX5 - 1		-0.0244	-8	PPB	SMPL	-7470A	21 Oct 2025 14:41:28
U	Q3392-08A - 1		4.0151	17457	PPB	SMPL	-7470A	21 Oct 2025 14:43:43
U	CCV68 - 1	CCV68	4.8501	21067	PPB	SMPL	-7470A	21 Oct 2025 14:45:59
U	CCB68 - 1	CCB68	-0.0880	-283	PPB	SMPL	-7470A	21 Oct 2025 14:48:17