

h


Instrument IC-1 Analyst : RM Method: 300.0 / 9056A

Ident	Con F-	Con CL-	Con NO ₂	Con BR-	Con NO ₃	Con HPO	Con SO ₄	Method name	date time	Initial Analyst
STD1	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-100925	10/9/2025 10:46	10 RM/IZ
STD2	0.440	0.646	0.652	2.141	0.541	1.118	3.299	IC1-100925	10/9/2025 11:08	10 RM/IZ
STD3	0.813	1.225	1.222	4.073	1.019	1.947	6.091	IC1-100925	10/9/2025 11:29	10 RM/IZ
STD4	0.993	1.497	1.496	4.994	1.249	2.669	7.507	IC1-100925	10/9/2025 11:51	10 RM/IZ
STD5	1.954	2.923	2.922	9.773	2.435	4.828	14.557	IC1-100925	10/9/2025 12:12	10 RM/IZ
STD6	3.928	5.923	5.920	19.752	4.934	9.517	29.510	IC1-100925	10/9/2025 12:34	10 RM/IZ
STD7	5.072	7.585	7.589	25.268	6.323	12.920	37.537	IC1-100925	10/9/2025 12:55	10 RM/IZ
ICV	1.970	2.918	2.942	9.842	2.398	5.003	14.702	IC1-100925	10/9/2025 13:16	10 RM/IZ
ICB	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-100925	10/9/2025 13:38	10 RM/IZ
CCV	2.031	3.077	3.060	10.306	2.564	4.739	15.199	IC1-100925	10/27/2025 11:38	10 RM/IZ
CCB	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-100925	10/27/2025 12:00	10 RM/IZ
LB137654BLW	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-100925	10/27/2025 12:21	10 RM/IZ
LB137654BSW	2.036	3.089	3.072	10.348	2.577	5.020	15.460	IC1-100925	10/27/2025 12:43	10 RM/IZ
Q3464-01	0.283	9.925	0.229	0.000	4.584	0.313	1081.143	IC1-100925	10/27/2025 13:04	10 RM/IZ
Q3464-01MS	2.294	12.642	3.147	10.107	6.948	5.334	1046.471	IC1-100925	10/27/2025 13:26	10 RM/IZ
Q3464-01MSD	2.203	12.517	3.013	9.659	6.838	4.861	1048.223	IC1-100925	10/27/2025 13:47	10 RM/IZ
Q3464-02	0.234	13.969	0.303	0.000	4.075	0.520	372.867	IC1-100925	10/27/2025 14:09	10 RM/IZ
Q3464-01DLX50	0.000	0.243	0.000	0.000	0.130	0.000	15.276	IC1-100925	10/27/2025 14:30	10 RM/IZ
Q3464-02DLX20	0.042	0.697	0.089	0.000	0.235	0.000	14.529	IC1-100925	10/27/2025 14:52	10 RM/IZ
CCV	2.007	3.067	3.066	10.307	2.562	4.790	15.184	IC1-100925	10/27/2025 15:13	10 RM/IZ
CCB	0.000	0.000	0.000	0.000	0.000	0.000	0.000	IC1-100925	10/27/2025 15:35	10 RM/IZ

Clear table

Instrument ID: IC-2 Analyst : IZ Method: 300.0 / 9056A											
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000 IC1-100925	10/9/2025 10:46	10	RM/IZ
STD2	0.4400	0.6460	0.6520	2.1410	0.5410	1.1180	3.2990	IC1-100925	10/9/2025 11:08	10	RM/IZ
STD3	0.8130	1.2250	1.2220	4.0730	1.0190	1.9470	6.0910	IC1-100925	10/9/2025 11:29	10	RM/IZ
STD4	0.9930	1.4970	1.4960	4.9940	1.2490	2.6690	7.5070	IC1-100925	10/9/2025 11:51	10	RM/IZ
STD5	1.9540	2.9230	2.9220	9.7730	2.4350	4.8280	14.5570	IC1-100925	10/9/2025 12:12	10	RM/IZ
STD6	3.9280	5.9230	5.9200	19.7520	4.9340	9.5170	29.5100	IC1-100925	10/9/2025 12:34	10	RM/IZ
STD7	5.0720	7.5850	7.5890	25.2680	6.3230	12.9200	37.5370	IC1-100925	10/9/2025 12:55	10	RM/IZ
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4				
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000				
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000				
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000				
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000				
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000				
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000				
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000				
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4				
STD1											
STD2	10.0000	7.6667	8.6667	7.0500	8.2000	11.8000	9.9667				
STD3	1.6250	2.0833	1.8333	1.8250	1.9000	-2.6500	1.5167				
STD4	-0.7000	-0.2000	-0.2667	-0.1200	-0.0800	6.7600	0.0933				
STD5	-2.3000	-2.5667	-2.6000	-2.2700	-2.6000	-3.4400	-2.9533				
STD6	-1.8000	-1.2833	-1.3333	-1.2400	-1.3200	-4.8300	-1.6333				
STD7	1.4400	1.1333	1.1867	1.0720	1.1680	3.3600	1.4514				

Fluoride (Anions)




Function: $A = -3.86859E-3 + 0.0146255 \times Q$

Relative standard deviation 2.744994 %

Correlation coefficient 0.999599

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.400	10.0	1.0	1.0	0.061	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	0.800	10.0	1.0	1.0	0.115	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.000	10.0	1.0	1.0	0.141	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	2.000	10.0	1.0	1.0	0.282	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	4.000	10.0	1.0	1.0	0.571	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	5.000	10.0	1.0	1.0	0.738	STD7	2025-10-09 12:55:27 UTC-4	used

Chloride (Anions)




Function: $A = -5.87555E-3 + 0.0101528 \times Q$

Relative standard deviation 2.276543 %

Correlation coefficient 0.999727

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.060	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.119	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.146	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.291	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	0.596	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	0.764	STD7	2025-10-09 12:55:27 UTC-4	used

Nitrite (Anions) $(\mu\text{S/cm}) \times \text{min}$ Function: $A = -0.0165604 + 0.0226258 \times Q$


Relative standard deviation 2.379226 %

Correlation coefficient 0.999705

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.131	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.260	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.322	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.645	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	1.323	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	1.700	STD7	2025-10-09 12:55:27 UTC-4	used

Bromide (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = -6.89623E-3 + 4.37562E-3 \times Q$


Relative standard deviation 2.114692 %

Correlation coefficient 0.999763

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	2.000	10.0	1.0	1.0	0.087	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	4.000	10.0	1.0	1.0	0.171	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	5.000	10.0	1.0	1.0	0.212	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	10.000	10.0	1.0	1.0	0.421	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	20.000	10.0	1.0	1.0	0.857	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	25.000	10.0	1.0	1.0	1.099	STD7	2025-10-09 12:55:27 UTC-4	used

Nitrate (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = -0.0141807 + 0.0252306 \times Q$


Relative standard deviation 2.346972 %

Correlation coefficient 0.999712

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.500	10.0	1.0	1.0	0.122	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.000	10.0	1.0	1.0	0.243	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.250	10.0	1.0	1.0	0.301	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	2.500	10.0	1.0	1.0	0.600	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	5.000	10.0	1.0	1.0	1.231	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	6.250	10.0	1.0	1.0	1.581	STD7	2025-10-09 12:55:27 UTC-4	used

Phosphate (Anions)

($\mu\text{S/cm}$) $\times$ min




Function: $A = -0.0117521 + 8.50315E-3 \times Q$

Relative standard deviation 6.489714 %

Correlation coefficient 0.997825

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	1.000	10.0	1.0	1.0	0.083	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	2.000	10.0	1.0	1.0	0.154	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	2.500	10.0	1.0	1.0	0.215	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	5.000	10.0	1.0	1.0	0.399	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	10.000	10.0	1.0	1.0	0.798	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	12.500	10.0	1.0	1.0	1.087	STD7	2025-10-09 12:55:27 UTC-4	used

Sulfate (Anions) $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -0.0307148 + 7.61403E-3 \times Q$

Relative standard deviation 2.830893 %

Correlation coefficient 0.999580


Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	3.000	10.0	1.0	1.0	0.220	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	6.000	10.0	1.0	1.0	0.433	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	7.500	10.0	1.0	1.0	0.541	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	15.000	10.0	1.0	1.0	1.078	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	30.000	10.0	1.0	1.0	2.216	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	37.000	10.0	1.0	1.0	2.827	STD7	2025-10-09 12:55:27 UTC-4	used

Sample data

Ident STD1
Sample type Standard 1
Determination start 2025-10-09 10:46:54 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.57 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident STD2
Sample type Standard 2
Determination start 2025-10-09 11:08:18 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.34 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.0605	0.366	0.440	Fluoride
2	6.680	0.0597	0.264	0.646	Chloride
3	8.865	0.1310	0.444	0.652	Nitrite
4	11.008	0.0868	0.245	2.141	Bromide
5	12.692	0.1222	0.301	0.541	Nitrate
6	14.212	0.0833	0.158	1.118	Phosphate
7	15.883	0.2204	0.426	3.299	Sulfate


Sample data

Ident STD3
Sample type Standard 3
Determination start 2025-10-09 11:29:42 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.63 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.388	0.1150	0.718	0.813	Fluoride
2	6.662	0.1185	0.531	1.225	Chloride
3	8.833	0.2598	0.891	1.222	Nitrite
4	10.962	0.1713	0.487	4.073	Bromide
5	12.637	0.2429	0.605	1.019	Nitrate
6	14.187	0.1538	0.307	1.947	Phosphate
7	15.888	0.4330	0.848	6.091	Sulfate


Sample data

Ident STD4
Sample type Standard 4
Determination start 2025-10-09 11:51:08 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.34 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.1414	0.897	0.993	Fluoride
2	6.655	0.1461	0.654	1.497	Chloride
3	8.818	0.3220	1.103	1.496	Nitrite
4	10.933	0.2116	0.604	4.994	Bromide
5	12.607	0.3009	0.750	1.249	Nitrate
6	14.185	0.2152	0.426	2.669	Phosphate
7	15.900	0.5409	1.049	7.507	Sulfate


Sample data

Ident STD5
Sample type Standard 5
Determination start 2025-10-09 12:12:34 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time	Area	Height	Concentration	Component name
	min	(µS/cm) x min	µS/cm	ppm	
1	3.390	0.2819	1.816	1.954	Fluoride
2	6.645	0.2909	1.308	2.923	Chloride
3	8.803	0.6446	2.221	2.922	Nitrite
4	10.910	0.4207	1.208	9.773	Bromide
5	12.582	0.6001	1.504	2.435	Nitrate
6	14.178	0.3988	0.783	4.828	Phosphate
7	15.905	1.0776	2.103	14.557	Sulfate


Sample data

Ident STD6
Sample type Standard 6
Determination start 2025-10-09 12:34:00 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.388	0.5707	3.761	3.928	Fluoride
2	6.638	0.5955	2.722	5.923	Chloride
3	8.797	1.3228	4.622	5.920	Nitrite
4	10.897	0.8574	2.493	19.752	Bromide
5	12.565	1.2307	3.128	4.934	Nitrate
6	14.170	0.7975	1.608	9.517	Phosphate
7	15.900	2.2162	4.397	29.510	Sulfate


Sample data

Ident STD7
Sample type Standard 7
Determination start 2025-10-09 12:55:27 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.01 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.395	0.7379	4.813	5.072	Fluoride
2	6.643	0.7642	3.464	7.585	Chloride
3	8.805	1.7004	5.904	7.589	Nitrite
4	10.905	1.0987	3.182	25.268	Bromide
5	12.572	1.5811	4.005	6.323	Nitrate
6	14.172	1.0869	2.232	12.920	Phosphate
7	15.905	2.8273	5.621	37.537	Sulfate


Sample data

Ident ICV
Sample type Check standard 1
Determination start 2025-10-09 13:16:55 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.23 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.383	0.2843	1.850	1.970	Fluoride
2	6.632	0.2904	1.314	2.918	Chloride
3	8.785	0.6491	2.249	2.942	Nitrite
4	10.887	0.4238	1.222	9.842	Bromide
5	12.553	0.5909	1.487	2.398	Nitrate
6	14.152	0.4137	0.815	5.003	Phosphate
7	15.880	1.0887	2.130	14.702	Sulfate

Sample data

Ident ICB
Sample type Sample
Determination start 2025-10-09 13:38:24 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.12 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2025-10-27 11:38:33 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.39 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.393	0.2932	2.098	2.031	Fluoride
2	6.555	0.3066	1.637	3.077	Chloride
3	8.638	0.6757	2.789	3.060	Nitrite
4	10.645	0.4441	1.558	10.306	Bromide
5	12.287	0.6328	1.943	2.564	Nitrate
6	13.800	0.3912	0.880	4.739	Phosphate
7	15.965	1.1265	2.594	15.199	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-10-27 12:00:03 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.22 MPa
Maximum pressure monitored yes
Temperature ---- °C


Anions

Sample data

Ident LB137654BLW
Sample type Sample
Determination start 2025-10-27 12:21:33 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.27 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident LB137654BSW
 Sample type Check standard 1
 Determination start 2025-10-27 12:43:04 UTC-4
 Method IC1-100925
 Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
 Channel Conductivity
 Recording time 19.0 min
 Integration Automatically
 Column type Metrosep A Supp 19 - 150/4.0
 Eluent composition not defined
 Flow 0.700 mL/min
 Maximum flow monitored yes
 Pressure 12.16 MPa
 Maximum pressure monitored yes
 Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.2939	2.106	2.036	Fluoride
2	6.535	0.3078	1.647	3.089	Chloride
3	8.602	0.6784	2.812	3.072	Nitrite
4	10.587	0.4459	1.571	10.348	Bromide
5	12.222	0.6361	1.962	2.577	Nitrate
6	13.795	0.4151	0.922	5.020	Phosphate
7	15.993	1.1464	2.623	15.460	Sulfate


Sample data

Ident Q3464-01
Sample type Sample
Determination start 2025-10-27 13:04:36 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.05 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	1.835	0.0080	0.044	invalid	
2	3.385	0.0375	0.224	0.283	Fluoride
3	6.535	1.0018	5.512	9.925	Chloride
4	7.152	0.0353	0.066	invalid	
5	8.590	0.0352	0.151	0.229	Nitrite
6	12.197	1.1425	3.713	4.584	Nitrate
7	13.798	0.0149	0.031	0.313	Phosphate
8	15.770	82.2878	189.149	1081.143	Sulfate


Sample data

Ident Q3464-01MS
Sample type Sample
Determination start 2025-10-27 13:26:10 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.11 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	1.860	0.0142	0.061	invalid	
2	2.800	0.0077	0.017	invalid	
3	3.392	0.3316	2.332	2.294	Fluoride
4	6.537	1.2777	7.027	12.642	Chloride
5	7.175	0.0337	0.063	invalid	
6	8.593	0.6955	2.982	3.147	Nitrite
7	10.565	0.4354	1.591	10.107	Bromide
8	12.190	1.7390	5.676	6.948	Nitrate
9	13.792	0.4418	1.015	5.334	Phosphate
10	15.790	79.6479	182.747	1046.471	Sulfate


Sample data

Ident Q3464-01MSD
Sample type Sample
Determination start 2025-10-27 13:47:43 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.11 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	1.855	0.0139	0.061	invalid	
2	2.732	0.0086	0.016	invalid	
3	3.388	0.3183	2.243	2.203	Fluoride
4	6.528	1.2650	6.985	12.517	Chloride
5	7.162	0.0339	0.064	invalid	
6	8.583	0.6651	2.864	3.013	Nitrite
7	10.552	0.4157	1.524	9.659	Bromide
8	12.173	1.7111	5.612	6.838	Nitrate
9	13.783	0.4016	0.913	4.861	Phosphate
10	15.787	79.7813	183.731	1048.223	Sulfate


Sample data

Ident Q3464-02
Sample type Sample
Determination start 2025-10-27 14:09:18 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.99 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	2.793	0.0057	0.013	invalid	
2	3.387	0.0304	0.223	0.234	Fluoride
3	6.540	1.4124	7.476	13.969	Chloride
4	7.113	0.0511	0.098	invalid	
5	8.575	0.0519	0.220	0.303	Nitrite
6	12.167	1.0139	3.226	4.075	Nitrate
7	13.787	0.0324	0.073	0.520	Phosphate
8	15.922	28.3595	69.851	372.867	Sulfate


Sample data

Ident Q3464-01DLX50
 Sample type Sample
 Determination start 2025-10-27 14:30:53 UTC-4
 Method IC1-100925
 Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
 Channel Conductivity
 Recording time 19.0 min
 Integration Automatically
 Column type Metrosep A Supp 19 - 150/4.0
 Eluent composition not defined
 Flow 0.700 mL/min
 Maximum flow monitored yes
 Pressure 12.16 MPa
 Maximum pressure monitored yes
 Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	6.515	0.0188	0.102	0.243	Chloride
2	12.165	0.0185	0.056	0.130	Nitrate
3	16.030	1.1324	2.598	15.276	Sulfate


Sample data

Ident Q3464-02DLX20
Sample type Sample
Determination start 2025-10-27 14:52:26 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.99 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.390	0.0023	0.012	0.042	Fluoride
2	6.508	0.0649	0.350	0.697	Chloride
3	8.547	0.0036	0.013	0.089	Nitrite
4	12.150	0.0451	0.137	0.235	Nitrate
5	16.033	1.0755	2.473	14.529	Sulfate


Sample data

Ident CCV
 Sample type Check standard 1
 Determination start 2025-10-27 15:13:59 UTC-4
 Method IC1-100925
 Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
 Channel Conductivity
 Recording time 19.0 min
 Integration Automatically
 Column type Metrosep A Supp 19 - 150/4.0
 Eluent composition not defined
 Flow 0.700 mL/min
 Maximum flow monitored yes
 Pressure 11.88 MPa
 Maximum pressure monitored yes
 Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.395	0.2897	2.091	2.007	Fluoride
2	6.507	0.3055	1.675	3.067	Chloride
3	8.548	0.6772	2.870	3.066	Nitrite
4	10.502	0.4441	1.600	10.307	Bromide
5	12.127	0.6323	1.992	2.562	Nitrate
6	13.788	0.3955	0.909	4.790	Phosphate
7	16.037	1.1254	2.624	15.184	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-10-27 15:35:29 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.99 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions

