



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB80560

Reagent/Standard	Lot/Log #
hydrazine sulfate solution 1	WP49646

Intercept: -0.0041

Slope: 1.0279

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.060	10/14/2016	08:55
2	CAL2	1	1	1.020	10/14/2016	08:58
3	CAL3	5	1	5.020	10/14/2016	08:58
4	CAL4	10	1	10.100	10/14/2016	08:58
5	CAL5	20	1	20.900	10/14/2016	08:58
6	CAL6	40	1	41.000	10/14/2016	08:58



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB80560

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	9.830	10/14/2016	08:58
2	ICB		1	0.056	0.058	10/14/2016	08:59
3	CCV1	10	1	10.000	9.733	10/14/2016	08:59
4	CCB1		1	0.061	0.063	10/14/2016	08:59
5	LB80560BL		1	0.068	0.070	10/14/2016	08:59
6	H1742-36		1	8.480	8.254	10/14/2016	08:59
7	H1786-01		1	0.118	0.119	10/14/2016	08:59
8	H1786-02		1	0.052	0.055	10/14/2016	08:59
9	H1786-02DUP		1	0.058	0.060	10/14/2016	08:59
10	CCV2	10	1	10.200	9.927	10/14/2016	09:00
11	CCB2		1	0.069	0.071	10/14/2016	09:00