


Chemtech Consulting Group

Analytical Review Report

Date Printed : 12/30/16
 Analyst : AK
 Data File : lb85134.csv

Approved By : 
 Approved Date :
 Worksheet # : 

m9012A-B Total Amenable el Reactive cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Reactive Cyanide											
0.0PPBCN	0.0PPBCN				W	12/30/16					
Reactive Cyanide		PASS	3.004		0.003						mg/L
5.0PPBCN	5.0PPBCN				W	12/30/16					
Reactive Cyanide		PASS	4.037		0.004						mg/L
10PPBCN	10PPBCN				W	12/30/16					
Reactive Cyanide		PASS	9.953		0.010						mg/L
50PPBCN	50PPBCN				W	12/30/16					
Reactive Cyanide		PASS	49.640		0.050						mg/L
100PPBCN	100PPBCN				W	12/30/16					
Reactive Cyanide		PASS	99.010		0.099						mg/L
250PPBCN	250PPBCN				W	12/30/16					
Reactive Cyanide		PASS	248.223		0.248						mg/L
500PPBCN	500PPBCN				W	12/30/16					
Reactive Cyanide		PASS	501.133		0.501						mg/L
ICV1	ICV1				W	12/30/16					
Reactive Cyanide		PASS	97.791		0.10	100.0	85	115			mg/L
ICB1	ICB1				W	12/30/16					
Reactive Cyanide		PASS	-0.009		0.000			+/-0.0050			mg/L
CCV1	CCV1				W	12/30/16					
Reactive Cyanide		PASS	244.389		0.24	96.0	90	110			mg/L
CCB1	CCB1				W	12/30/16					
Reactive Cyanide		PASS	0.205		0.000			+/-0.0050			mg/L
PB95679BLS	PB95679BLS				S	12/30/16					
Reactive Cyanide		PASS	-0.950		-0.010			+/-0.0500			mg/Kg
H6304-04	TP11-COMP8			1	S	12/30/16					
Reactive Cyanide		PASS	-1.105		-0.011						mg/Kg
H6304-04D	TP11-COMP8D			1	S	12/30/16					
Reactive Cyanide		PASS	-0.893		-0.009				0	20	mg/Kg
CCV2	CCV2				W	12/30/16					
Reactive Cyanide		PASS	251.436		0.25	100.0	90	110			mg/L
CCB2	CCB2				W	12/30/16					
Reactive Cyanide		PASS	-1.055		-0.001			+/-0.0050			mg/L