

Chemtech Consulting Group

Analytical Review Report

Date Printed : 12/30/16
Analyst : AK
Data File : lb85136.csv

Approved By : 
Approved Date : 1/3/17
Worksheet # :

M 9012 A-B Total Amenable of Reactive Cyanide-16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Cyanide											
0.0PPBCN	0.0PPBCN				W	12/30/16					
Cyanide		PASS	3.004		0.003						mg/L
5.0PPBCN	5.0PPBCN				W	12/30/16					
Cyanide		PASS	4.037		0.004						mg/L
10PPBCN	10PPBCN				W	12/30/16					
Cyanide		PASS	9.953		0.010						mg/L
50PPBCN	50PPBCN				W	12/30/16					
Cyanide		PASS	49.640		0.050						mg/L
100PPBCN	100PPBCN				W	12/30/16					
Cyanide		PASS	99.010		0.099						mg/L
250PPBCN	250PPBCN				W	12/30/16					
Cyanide		PASS	248.223		0.248						mg/L
500PPBCN	500PPBCN				W	12/30/16					
Cyanide		PASS	501.133		0.501						mg/L
ICV1	ICV1				W	12/30/16					
Cyanide		PASS	95.906		0.10	100.0	90	110			mg/L
ICB1	ICB1				W	12/30/16					
Cyanide		PASS	0.394		0.000			+/-0.0025			mg/L
CCV1	CCV1				W	12/30/16					
Cyanide		PASS	239.148		0.24	96.0	90	110			mg/L
CCB1	CCB1				W	12/30/16					
Cyanide		PASS	0.583		0.001			+/-0.0025			mg/L
LOWPB95704	LOWPB95704				W	12/30/16					
Cyanide		PASS	8.931		0.009						mg/L
HIGHPB95704	HIGHPB95704				W	12/30/16					
Cyanide		PASS	484.368		0.484						mg/L
PB95704BLS	PB95704BLS				S	12/30/16					
Cyanide		PASS	1.373		0.067			+/-0.0330			mg/Kg
PB95704BSS	PB95704BSS				S	12/30/16					
Cyanide		PASS	96.422		4.64	93.0	85.00	115.00			mg/Kg
H6304-03	TP11-COMP8			1	S	12/30/16					
Cyanide		PASS	-0.257		-0.016						mg/Kg
H6304-03D	TP11-COMP8D			1	S	12/30/16					
Cyanide		PASS	0.121		0.008				0	20	mg/Kg
H6304-03S	TP11-COMP8S			1	S	12/30/16					
Cyanide		PASS	30.884		1.990	77.0	75	125			mg/Kg
H6304-03SD	TP11-COMP8SD			1	S	12/30/16					
Cyanide		PASS	29.763		1.890	75.0	75	125	5.2	20	mg/Kg
CCV2	CCV2				W	12/30/16					
Cyanide		PASS	255.339		0.26	104.0	90	110			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 12/30/16
Analyst : AK
Data File : lb85136.csv

Approved By :
Approved Date :
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method				Line 2
Cyanide											
CCB2	CCB2			W	12/30/16						
Cyanide		PASS	1.409	0.001			+/-0.0025			mg/L	