


Chemtech Consulting Group

Analytical Review Report

Date Printed : 12/30/16
Analyst : AK
Data File : LB85144.csv

Approved By : 
Approved Date : 
Worksheet # :

m9065-Phenolics-11

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Phenol												
0.05PPM	0.05PPM				W	12/30/16						
Phenolics		PASS	0.055		0.055						mg/L	
0.1PPM	0.1PPM				W	12/30/16						
Phenolics		PASS	0.110		0.110						mg/L	
0.25PPM	0.25PPM				W	12/30/16						
Phenolics		PASS	0.250		0.250						mg/L	
0.50PPM	0.50PPM				W	12/30/16						
Phenolics		PASS	0.494		0.494						mg/L	
1.0PPM	1.0PPM				W	12/30/16						
Phenolics		PASS	0.980		0.980						mg/L	
2.0PPM	2.0PPM				W	12/30/16						
Phenolics		PASS	2.011		2.01						mg/L	
ICV1	ICV1				W	12/30/16						
Phenolics		PASS	0.920		0.92	92.0	90	110			mg/L	
ICB1	ICB1				W	12/30/16						
Phenolics		PASS	-0.010		-0.010			+/-0.0100			mg/L	
CCV1	CCV1				W	12/30/16						
Phenolics		PASS	0.914		0.91	91.0	90	110			mg/L	
CCB1	CCB1				W	12/30/16						
Phenolics		PASS	-0.004		-0.004			+/-0.0100			mg/L	
PB95683BLS	PB95683BLS				S	12/30/16						
Phenolics		PASS	-0.005		-0.260			+/-0.0780			mg/Kg	
H6304-03	TP11-COMP8		1		S	12/30/16						
Phenolics		PASS	0.001		0.079						mg/Kg	
H6304-03D	TP11-COMP8D		1		S	12/30/16						
Phenolics		PASS	0.000		0.020				0	20	mg/Kg	
H6304-03S	TP11-COMP8S		1		S	12/30/16						
Phenolics		PASS	0.865		55.800	86.0	75	125			mg/Kg	
H6304-03SD	TP11-COMP8SD		1		S	12/30/16						
Phenolics		PASS	0.868		56.600	87.0	75	125	1	20	mg/Kg	
PB95683BSS	PB95683BSS				S	12/30/16						
Phenolics		PASS	0.920		46.00	92.0	80.00	120.00			mg/Kg	
CCV2	CCV2				W	12/30/16						
Phenolics		PASS	0.908		0.91	91.0	90	110			mg/L	
CCB2	CCB2				W	12/30/16						
Phenolics		PASS	-0.015		-0.015			+/-0.0100			mg/L	