

LB85157

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/3/2017 13:37

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.690	0.0	0.094	
ICB1	1.090	0.0	0.006	
CCV1	236.965	0.0	0.219	
CCB1	-0.287	0.0	0.005	
PB95733BLS	-0.354	0.0	0.005	
H6322-04	-1.247	0.0	0.004	
H6322-04D	0.993	0.0	0.006	
H6325-02	-0.630	0.0	0.005	
CCV2	249.926	0.0	0.231	
CCB2	-0.336	0.0	0.005	

N	10
Mean	58.481
SD	102.2847
CV%	174.90

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 03 12:20:37 2017

Tue Jan 03 13:21:56 2017

ALL

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.1962	µg/l	1/3/2017 12:43:29	
5.OPPBCN	A	Reactive CN	P	4.9261	µg/l	1/3/2017 12:43:30	
10PPBCN	A	Reactive CN	P	11.5016	µg/l	1/3/2017 12:43:31	
50PPBCN	A	Reactive CN	P	49.453	µg/l	1/3/2017 12:43:32	
100PPBCN	A	Reactive CN	P	100.3035	µg/l	1/3/2017 12:43:33	
250PPBCN	A	Reactive CN	P	248.0946	µg/l	1/3/2017 12:43:34	
500PPBCN	A	Reactive CN	P	500.9174	µg/l	1/3/2017 12:43:35	
ICV1	S	Reactive CN	P	98.69	µg/l	1/3/2017 13:21:46	
ICB1	S	Reactive CN	P	1.0899	µg/l	1/3/2017 13:21:47	
CCV1	S	Reactive CN	P	236.9651	µg/l	1/3/2017 13:21:48	
CCB1	S	Reactive CN	P	-0.2874	µg/l	1/3/2017 13:21:49	
PB95733BLS	S	Reactive CN	P	-0.3539	µg/l	1/3/2017 13:21:50	
H6322-04	S	Reactive CN	P	-1.2471	µg/l	1/3/2017 13:21:51	
H6322-04D	S	Reactive CN	P	0.993	µg/l	1/3/2017 13:21:52	
H6325-02	S	Reactive CN	P	-0.6302	µg/l	1/3/2017 13:21:53	
CCV2	S	Reactive CN	P	249.9259	µg/l	1/3/2017 13:21:54	
CCB2	S	Reactive CN	P	-0.3364	µg/l	1/3/2017 13:21:55	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by :

1/3/2017 12:58


Test Total CN

Accepted 1/3/2017 12:58

Factor 1106
Bias 0.005

Coeff. of det. 0.999965

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	-0.1962	0.0000	
2	5.0PPBCN	0.010	4.9261	5.0000	
3	10PPBCN	0.016	11.5016	10.0000	
4	50PPBCN	0.050	49.4530	50.0000	
5	100PPBCN	0.096	100.3035	100.0000	
6	250PPBCN	0.229	248.0946	250.0000	
7	500PPBCN	0.458	500.9174	500.0000	