

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/4/17

Analyst : AK

Data File : lb85157.csv

Approved By : 

Approved Date : 1/4/17

Worksheet # :

079012 A-13, Total Amenable of Reactive cyanide - 16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	1/3/17						
Reactive Cyanide		PASS	-0.196		-0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	1/3/17						
Reactive Cyanide		PASS	4.926		0.005						mg/L	
10PPBCN	10PPBCN				W	1/3/17						
Reactive Cyanide		PASS	11.502		0.012						mg/L	
50PPBCN	50PPBCN				W	1/3/17						
Reactive Cyanide		PASS	49.453		0.049						mg/L	
100PPBCN	100PPBCN				W	1/3/17						
Reactive Cyanide		PASS	100.303		0.100						mg/L	
250PPBCN	250PPBCN				W	1/3/17						
Reactive Cyanide		PASS	248.095		0.248						mg/L	
500PPBCN	500PPBCN				W	1/3/17						
Reactive Cyanide		PASS	500.917		0.501						mg/L	
ICV1	ICV1				W	1/3/17						
Reactive Cyanide		PASS	98.690		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/3/17						
Reactive Cyanide		PASS	1.090		0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	1/3/17						
Reactive Cyanide		PASS	236.965		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/3/17						
Reactive Cyanide		PASS	-0.287		0.000			+/-0.0050			mg/L	
PB95733BLS	PB95733BLS				S	1/3/17						
Reactive Cyanide		PASS	-0.354		-0.004			+/-0.0500			mg/Kg	
H6322-04	TP9-COMP9			1	S	1/3/17						
Reactive Cyanide		PASS	-1.247		-0.012						mg/Kg	
H6322-04D	TP9-COMP9D			1	S	1/3/17						
Reactive Cyanide		PASS	0.993		0.010				0	20	mg/Kg	
H6325-02	LAR1020			1	S	1/3/17						
Reactive Cyanide		PASS	-0.630		-0.006						mg/Kg	
CCV2	CCV2				W	1/3/17						
Reactive Cyanide		PASS	249.926		0.25	100.0	90	110			mg/L	
CCB2	CCB2				W	1/3/17						
Reactive Cyanide		PASS	-0.336		0.000			+/-0.0050			mg/L	