


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/4/17

Analyst : AK

Data File : lb85158.csv

Approved By : 

Approved Date : 1/5/17

Worksheet # :

m9012 A-B, Total Amenable of Reactive cyanide - 6

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/3/17						
Cyanide		PASS	-0.196		-0.000						mg/L	
5.0PPBCN	5.0PPBCN				W	1/3/17						
Cyanide		PASS	4.926		0.005						mg/L	
10PPBCN	10PPBCN				W	1/3/17						
Cyanide		PASS	11.502		0.012						mg/L	
50PPBCN	50PPBCN				W	1/3/17						
Cyanide		PASS	49.453		0.049						mg/L	
100PPBCN	100PPBCN				W	1/3/17						
Cyanide		PASS	100.303		0.100						mg/L	
250PPBCN	250PPBCN				W	1/3/17						
Cyanide		PASS	248.095		0.248						mg/L	
500PPBCN	500PPBCN				W	1/3/17						
Cyanide		PASS	500.917		0.501						mg/L	
ICV1	ICV1				W	1/3/17						
Cyanide		PASS	99.802		0.10	100.0	90	110			mg/L	
ICB1	ICB1				W	1/3/17						
Cyanide		PASS	-0.176		0.000			+/-0.0025			mg/L	
CCV1	CCV1				W	1/3/17						
Cyanide		PASS	247.625		0.25	100.0	90	110			mg/L	
CCB1	CCB1				W	1/3/17						
Cyanide		PASS	1.786		0.002			+/-0.0025			mg/L	
LOWPB95735	LOWPB95735				W	1/3/17						
Cyanide		PASS	9.521		0.010						mg/L	
HIGHPB95735	HIGHPB95735				W	1/3/17						
Cyanide		PASS	498.959		0.499						mg/L	
PB95735BLS	PB95735BLS				S	1/3/17						
Cyanide		PASS	0.509		0.025			+/-0.0330			mg/Kg	
PB95735BSS	PB95735BSS				S	1/3/17						
Cyanide		PASS	102.056		5.10	102.0	85.00	115.00			mg/Kg	
H6322-03	TP9-COMP9			1	S	1/3/17						
Cyanide		PASS	1.580		0.119						mg/Kg	
H6322-03D	TP9-COMP9D			1	S	1/3/17						
Cyanide		FAIL	1.072		0.080				39.2	20	mg/Kg	
H6322-03S	TP9-COMP9S			1	S	1/3/17						
Cyanide		FAIL	3.213		0.244	4.0	75	125			mg/Kg	
H6322-03SD	TP9-COMP9SD			1	S	1/3/17						
Cyanide		FAIL	4.098		0.314	6.0	75	125	25.1	20	mg/Kg	
CCV2	CCV2				W	1/3/17						
Cyanide		PASS	268.700		0.27	108.0	90	110			mg/L	

Chemtech Consulting Group


Analytical Review Report


Date Printed : 1/4/17

Analyst : AK

Data File : lb85158.csv

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide											
CCB2	CCB2			W	1/3/17						
Cyanide		PASS		-0.283	0.000		+/-0.0025			mg/L	