

# Chemtech Consulting Group

## Analytical Review Report

Date Printed : 1/4/17

Analyst : AK

Data File : lb85166.csv

Approved By : 

Approved Date : 1/4/17

Worksheet # :

msm 4500 - NH<sub>3</sub> B, G, H - Ammonia - 181

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
			PPB				Method	Method			
Parameter					Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
											Line 2
Ammonia											
0.1PPM	0.1PPM				W	1/4/17					
Ammonia as N	PASS		0.106		0.106						mg/L
0.2PPM	0.2PPM				W	1/4/17					
Ammonia as N	PASS		0.203		0.203						mg/L
0.4PPM	0.4PPM				W	1/4/17					
Ammonia as N	PASS		0.396		0.396						mg/L
1.0PPM	1.0PPM				W	1/4/17					
Ammonia as N	PASS		0.993		0.993						mg/L
1.3PPM	1.3PPM				W	1/4/17					
Ammonia as N	PASS		1.329		1.33						mg/L
2.0PPM	2.0PPM				W	1/4/17					
Ammonia as N	PASS		2.007		2.01						mg/L
ICV1	ICV1				W	1/4/17					
Ammonia as N	PASS		1.045		1.04	104.0	90	110			mg/L
ICB1	ICB1				W	1/4/17					
Ammonia as N	PASS		0.002		0.002			+/-0.0340			mg/L
CCV1	CCV1				W	1/4/17					
Ammonia as N	PASS		0.954		0.95	95.0	90	110			mg/L
CCB1	CCB1				W	1/4/17					
Ammonia as N	PASS		0.002		0.002			+/-0.0340			mg/L
PB95738BLS	PB95738BLS				S	1/4/17					
Ammonia as N	PASS		0.004		0.180			+/-0.0100			mg/Kg
PB95738BSS	PB95738BSS				S	1/4/17					
Ammonia as N	PASS		1.039		51.90	104.0	80.00	120.00			mg/Kg
H6325-01	LAR1020		1		S	1/4/17					
Ammonia as N	PASS		0.448		49.2						mg/Kg
H6325-01D	LAR1020D		1		S	1/4/17					
Ammonia as N	PASS		0.461		51.000				3.6	20	mg/Kg
H6325-01S	LAR1020S		1		S	1/4/17					
Ammonia as N	PASS		1.465		164.000	103.0	75	125			mg/Kg
H6325-01SD	LAR1020SD		1		S	1/4/17					
Ammonia as N	PASS		1.509		165.000	105.0	75	125	0.6	20	mg/Kg
CCV2	CCV2				W	1/4/17					
Ammonia as N	PASS		1.023		1.02	102.0	90	110			mg/L
CCB2	CCB2				W	1/4/17					
Ammonia as N	PASS		0.019		0.019			+/-0.0340			mg/L