

2385168

=====

Test results Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AIC

1/4/2017 11:53

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.965	0.0	0.109	
ICB1	0.012	0.0	0.005	
CCV1	0.977	0.0	0.110	
CCB1	0.014	0.0	0.006	
PB95737BLS	0.010	0.0	0.005	
PB95737BSS	0.981	0.0	0.110	
H6322-03	0.004	0.0	0.005	
H6322-03D	0.029	0.0	0.007	
H6322-03S	0.891	0.0	0.101	
H6322-03SD	0.876	0.0	0.099	
CCV2	0.975	0.0	0.110	
CCB2	0.009	0.0	0.005	

N	12
Mean	0.479
SD	0.4874
CV%	101.86

LB 85168

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 04 11:14:29 2017

Wed Jan 04 11:52:12 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0575	mg/l	1/4/2017 11:14:29	
0.1PPM	A	Phenolics-	P	0.1075	mg/l	1/4/2017 11:14:30	
0.25PPM	A	Phenolics-	P	0.2573	mg/l	1/4/2017 11:14:31	
0.50PPM	A	Phenolics-	P	0.4924	mg/l	1/4/2017 11:14:32	
1.0PPM	A	Phenolics-	P	0.9698	mg/l	1/4/2017 11:14:33	
2.0PPM	A	Phenolics-	P	2.0155	mg/l	1/4/2017 11:14:34	
ICV1	S	Phenolics-	P	0.9652	mg/l	1/4/2017 11:48:13	
ICB1	S	Phenolics-	P	0.012	mg/l	1/4/2017 11:48:14	
CCV1	S	Phenolics-	P	0.9768	mg/l	1/4/2017 11:48:15	
CCB1	S	Phenolics-	P	0.0136	mg/l	1/4/2017 11:48:16	
PB95737BLS	S	Phenolics-	P	0.0099	mg/l	1/4/2017 11:48:17	
PB95737BSS	S	Phenolics-	P	0.9806	mg/l	1/4/2017 11:48:18	
H6322-03	S	Phenolics-	P	0.004	mg/l	1/4/2017 11:48:19	
H6322-03D	S	Phenolics-	P	0.0293	mg/l	1/4/2017 11:48:20	
H6322-03S	S	Phenolics-	P	0.8913	mg/l	1/4/2017 11:48:21	
H6322-03SD	S	Phenolics-	P	0.876	mg/l	1/4/2017 11:48:22	
CCV2	S	Phenolics-	P	0.9746	mg/l	1/4/2017 11:48:23	
CCB2	S	Phenolics-	P	0.0086	mg/l	1/4/2017 11:48:24	

6885168

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : g1k

1/4/2017 11:23


Test Phenolics-

Accepted 1/4/2017 11:22

Factor 9.224
Bias 0.004

Coeff. of det. 0.999508

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0575	0.0500	
2	Phenol-2	0.016	0.1075	0.1000	
3	Phenol-2	0.032	0.2573	0.2500	
4	Phenol-2	0.057	0.4924	0.5000	
5	Phenol-2	0.109	0.9698	1.0000	
6	Phenol-2	0.223	2.0155	2.0000	