

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/4/17

Analyst : AK

Data File : lb85168.csv

Approved By : 

Approved Date : 1/9/17

Worksheet # :

mgas-Phenolics-11

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Phenol												
0.05PPM	0.05PPM				W	1/4/17						
Phenolics		PASS	0.058		0.058						mg/L	
0.1PPM	0.1PPM				W	1/4/17						
Phenolics		PASS	0.108		0.108						mg/L	
0.25PPM	0.25PPM				W	1/4/17						
Phenolics		PASS	0.257		0.257						mg/L	
0.50PPM	0.50PPM				W	1/4/17						
Phenolics		PASS	0.492		0.492						mg/L	
1.0PPM	1.0PPM				W	1/4/17						
Phenolics		PASS	0.970		0.970						mg/L	
2.0PPM	2.0PPM				W	1/4/17						
Phenolics		PASS	2.016		2.02						mg/L	
ICV1	ICV1				W	1/4/17						
Phenolics		PASS	0.965		0.97	97.0	90	110			mg/L	
ICB1	ICB1				W	1/4/17						
Phenolics		PASS	0.012		0.012			+/-0.0100			mg/L	
CCV1	CCV1				W	1/4/17						
Phenolics		PASS	0.977		0.98	98.0	90	110			mg/L	
CCB1	CCB1				W	1/4/17						
Phenolics		PASS	0.014		0.014			+/-0.0100			mg/L	
PB95737BLS	PB95737BLS				S	1/4/17						
Phenolics		PASS	0.010		0.495			+/-0.0780			mg/Kg	
PB95737BSS	PB95737BSS				S	1/4/17						
Phenolics		PASS	0.981		49.00	98.0	80.00	120.00			mg/Kg	
H6322-03	TP9-COMP9		1		S	1/4/17						
Phenolics		PASS	0.004		0.312						mg/Kg	
H6322-03D	TP9-COMP9D		1		S	1/4/17						
Phenolics		FAIL	0.029		2.270				152	20	mg/Kg	
H6322-03S	TP9-COMP9S		1		S	1/4/17						
Phenolics		PASS	0.891		68.300	89.0	75	125			mg/Kg	
H6322-03SD	TP9-COMP9SD		1		S	1/4/17						
Phenolics		PASS	0.876		67.800	87.0	75	125	1	20	mg/Kg	
CCV2	CCV2				W	1/4/17						
Phenolics		PASS	0.975		0.98	98.0	90	110			mg/L	
CCB2	CCB2				W	1/4/17						
Phenolics		PASS	0.009		0.009			+/-0.0100			mg/L	