

LB85202

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/5/2017 14:34

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.859	0.0	0.098	
ICB1	-0.540	0.0	0.005	
CCV1	238.703	0.0	0.235	
CCB1	-0.206	0.0	0.005	
PB95771BLS	-0.413	0.0	0.005	
II1025-04	-0.922	0.0	0.004	
II1025-04D	-0.928	0.0	0.004	
CCV2	242.968	0.0	0.239	
CCB2	-0.263	0.0	0.005	

N	9
Mean	63.918
SD	105.2542
CV%	164.67

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 05 12:16:47 2017

Thu Jan 05 14:32:58 2017

AK

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-1.0606	µg/l	1/5/2017 14:02:57	
5.0PPBCN	A	Reactive CN	P	4.178	µg/l	1/5/2017 14:02:58	
10PPBCN	A	Reactive CN	P	9.0146	µg/l	1/5/2017 14:02:59	
50PPBCN	A	Reactive CN	P	50.7357	µg/l	1/5/2017 14:03:00	
100PPBCN	A	Reactive CN	P	102.3281	µg/l	1/5/2017 14:03:01	
250PPBCN	A	Reactive CN	P	250.6309	µg/l	1/5/2017 14:03:02	
500PPBCN	A	Reactive CN	P	499.1733	µg/l	1/5/2017 14:03:03	
ICV1	S	Reactive CN	P	96.8591	µg/l	1/5/2017 14:32:49	
ICB1	S	Reactive CN	P	-0.5405	µg/l	1/5/2017 14:32:50	
CCV1	S	Reactive CN	P	238.7033	µg/l	1/5/2017 14:32:51	
CCB1	S	Reactive CN	P	-0.2061	µg/l	1/5/2017 14:32:52	
PB95771BLS	S	Reactive CN	P	-0.4126	µg/l	1/5/2017 14:32:53	
I1025-04	S	Reactive CN	P	-0.9222	µg/l	1/5/2017 14:32:54	
I1025-04D	S	Reactive CN	P	-0.9284	µg/l	1/5/2017 14:32:55	
CCV2	S	Reactive CN	P	242.9678	µg/l	1/5/2017 14:32:56	
CCB2	S	Reactive CN	P	-0.2628	µg/l	1/5/2017 14:32:57	

LB 85202

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ALS

1/5/2017 14:09


Test Total CN

Accepted 1/5/2017 14:09

Factor 1041
Bias 0.005

Coeff. of det. 0.999952

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-1.0606	0.0000	
2	5.0PPBCN	0.009	4.1780	5.0000	
3	10PPBCN	0.014	9.0146	10.0000	
4	50PPBCN	0.054	50.7357	50.0000	
5	100PPBCN	0.104	102.3281	100.0000	
6	250PPBCN	0.246	250.6309	250.0000	
7	500PPBCN	0.485	499.1733	500.0000	