

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/5/17

Analyst : AK

Data File : lb85202.csv

Approved By : 

Approved Date : 1/5/17

Worksheet # :

m9012 A-B Total Amenable @ Reactive cyanide -16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				
			PPB				Method	Method				Line 1
Parameter					Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	1/5/17						
Reactive Cyanide		PASS	-1.061		-0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/5/17						
Reactive Cyanide		PASS	4.178		0.004						mg/L	
10PPBCN	10PPBCN				W	1/5/17						
Reactive Cyanide		PASS	9.015		0.009						mg/L	
50PPBCN	50PPBCN				W	1/5/17						
Reactive Cyanide		PASS	50.736		0.051						mg/L	
100PPBCN	100PPBCN				W	1/5/17						
Reactive Cyanide		PASS	102.328		0.102						mg/L	
250PPBCN	250PPBCN				W	1/5/17						
Reactive Cyanide		PASS	250.631		0.251						mg/L	
500PPBCN	500PPBCN				W	1/5/17						
Reactive Cyanide		PASS	499.173		0.499						mg/L	
ICV1	ICV1				W	1/5/17						
Reactive Cyanide		PASS	96.859		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/5/17						
Reactive Cyanide		PASS	-0.540		-0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	1/5/17						
Reactive Cyanide		PASS	238.703		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/5/17						
Reactive Cyanide		PASS	-0.206		0.000			+/-0.0050			mg/L	
PB95771BLS	PB95771BLS				S	1/5/17						
Reactive Cyanide		PASS	-0.413		-0.004			+/-0.0500			mg/Kg	
I1025-04	TP7-8-COMP10			1	S	1/5/17						
Reactive Cyanide		PASS	-0.922		-0.009						mg/Kg	
I1025-04D	TP7-8-COMP10D			1	S	1/5/17						
Reactive Cyanide		PASS	-0.928		-0.009				0	20	mg/Kg	
CCV2	CCV2				W	1/5/17						
Reactive Cyanide		PASS	242.968		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	1/5/17						
Reactive Cyanide		PASS	-0.263		0.000			+/-0.0050			mg/L	