


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/6/17
Analyst : AK
Data File : lb85204.csv

Approved By : 
Approved Date :
Worksheet # :

079012A-B Total Amenable & Reactive Cyanide-16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				
Parameter			PPB		Final Conc	%Rec	Method	LCL	UCL	RPD	Max RPD	Units
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/5/17						
Cyanide		PASS	-1.061		-0.001							mg/L
5.0PPBCN	5.0PPBCN				W	1/5/17						
Cyanide		PASS	4.178		0.004							mg/L
10PPBCN	10PPBCN				W	1/5/17						
Cyanide		PASS	9.015		0.009							mg/L
50PPBCN	50PPBCN				W	1/5/17						
Cyanide		PASS	50.736		0.051							mg/L
100PPBCN	100PPBCN				W	1/5/17						
Cyanide		PASS	102.328		0.102							mg/L
250PPBCN	250PPBCN				W	1/5/17						
Cyanide		PASS	250.631		0.251							mg/L
500PPBCN	500PPBCN				W	1/5/17						
Cyanide		PASS	499.173		0.499							mg/L
ICV1	ICV1				W	1/5/17						
Cyanide		PASS	257.500		0.26	104.0	90	110				mg/L
ICB1	ICB1				W	1/5/17						
Cyanide		PASS	-4.828		-0.005				+/-0.0025			mg/L
CCV1	CCV1				W	1/5/17						
Cyanide		PASS	240.518		0.24	96.0	90	110				mg/L
CCB1	CCB1				W	1/5/17						
Cyanide		PASS	-2.233		-0.002				+/-0.0025			mg/L
HIGHPB95773	HIGHPB95773				W	1/5/17						
Cyanide		PASS	486.971		0.487							mg/L
PB95773BLS	PB95773BLS				S	1/5/17						
Cyanide		PASS	0.383		0.019				+/-0.0330			mg/Kg
PB95773BSS	PB95773BSS				S	1/5/17						
Cyanide		PASS	100.917		5.05	101.0	85.00	115.00				mg/Kg
LOWPB95773	LOWPB95773				W	1/5/17						
Cyanide		PASS	8.553		0.009							mg/L
I1025-03	TP7-8-COMP10			1	S	1/5/17						
Cyanide		PASS	-0.119		-0.009							mg/Kg
I1025-03D	TP7-8-COMP10D			1	S	1/5/17						
Cyanide		PASS	-0.035		-0.003					0	20	mg/Kg
I1025-03S	TP7-8-COMP10S			1	S	1/5/17						
Cyanide		FAIL	4.287		0.327	11.0	75	125				mg/Kg
I1025-03SD	TP7-8-COMP10SD			1	S	1/5/17						
Cyanide		FAIL	3.005		0.227	8.0	75	125	36.1	20		mg/Kg
CCV2	CCV2				W	1/5/17						
Cyanide		PASS	250.531		0.25	100.0	90	110				mg/L

Chemtech Consulting Group


Analytical Review Report


Date Printed : 1/6/17

Analyst : AK

Data File : lb85204.csv

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method				Line 2
Cyanide											
CCB2	CCB2			W	1/5/17						
Cyanide		PASS		-0.617	-0.001			+/-0.0025		mg/L	