

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
Analyst : AK
Data File : lb85243.csv

Approved By : 
Approved Date : 1/10/17
Worksheet # :

m9012 A-B Total Ammonia & Reactive cyanide

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				
Parameter			PPB		Final Conc	%Rec	Method	Method				Line 1
							LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	1/9/17						
Reactive Cyanide		PASS	-1.569		-0.002						mg/L	
5.0PPBCN	5.0PPBCN				W	1/9/17						
Reactive Cyanide		PASS	5.485		0.005						mg/L	
10PPBCN	10PPBCN				W	1/9/17						
Reactive Cyanide		PASS	8.695		0.009						mg/L	
50PPBCN	50PPBCN				W	1/9/17						
Reactive Cyanide		PASS	50.358		0.050						mg/L	
100PPBCN	100PPBCN				W	1/9/17						
Reactive Cyanide		PASS	101.441		0.101						mg/L	
250PPBCN	250PPBCN				W	1/9/17						
Reactive Cyanide		PASS	251.787		0.252						mg/L	
500PPBCN	500PPBCN				W	1/9/17						
Reactive Cyanide		PASS	498.804		0.499						mg/L	
ICV1	ICV1				W	1/9/17						
Reactive Cyanide		PASS	96.244		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/9/17						
Reactive Cyanide		PASS	-0.628		-0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	1/9/17						
Reactive Cyanide		PASS	242.988		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/9/17						
Reactive Cyanide		PASS	-0.561		-0.001			+/-0.0050			mg/L	
PB95813BLS	PB95813BLS				S	1/9/17						
Reactive Cyanide		PASS	-0.773		-0.008			+/-0.0500			mg/Kg	
I1045-07	P3-SP1			1	S	1/9/17						
Reactive Cyanide		PASS	-1.151		-0.011						mg/Kg	
I1045-07D	P3-SP1D			1	S	1/9/17						
Reactive Cyanide		PASS	-0.930		-0.009				0	20	mg/Kg	
I1045-08	P3-SP2			1	S	1/9/17						
Reactive Cyanide		PASS	-1.071		-0.011						mg/Kg	
I1045-09	RAP-SP1			1	S	1/9/17						
Reactive Cyanide		PASS	-1.036		-0.010						mg/Kg	
I1045-10	RAP-SP2			1	S	1/9/17						
Reactive Cyanide		PASS	-1.509		-0.015						mg/Kg	
I1045-11	RAP-SP3			1	S	1/9/17						
Reactive Cyanide		PASS	-1.219		-0.012						mg/Kg	
I1045-12	RAP-SP4			1	S	1/9/17						
Reactive Cyanide		PASS	-1.240		-0.012						mg/Kg	
CCV2	CCV2				W	1/9/17						
Reactive Cyanide		PASS	259.243		0.26	104.0	90	110			mg/L	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
Analyst : AK
Data File : lb85243.csv

Approved By :
Approved Date : 1/10/17
Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis					Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units		Line 2
Reactive Cyanide												
CCB2	CCB2			W	1/9/17							
Reactive Cyanide		PASS		-1.028	-0.001		+/-0.0050			mg/L		