

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
Analyst : AK
Data File : lb85244.csv

Approved By : 
Approved Date : 1/10/17
Worksheet # :


mq012 A-B Total Amenable & Reactive Cyanide -10

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method				Line 1
Parameter				Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/9/17						
Cyanide		PASS	-1.569	-0.002							mg/L	
5.0PPBCN	5.0PPBCN				W	1/9/17						
Cyanide		PASS	5.485	0.005							mg/L	
10PPBCN	10PPBCN				W	1/9/17						
Cyanide		PASS	8.695	0.009							mg/L	
50PPBCN	50PPBCN				W	1/9/17						
Cyanide		PASS	50.358	0.050							mg/L	
100PPBCN	100PPBCN				W	1/9/17						
Cyanide		PASS	101.441	0.101							mg/L	
250PPBCN	250PPBCN				W	1/9/17						
Cyanide		PASS	251.787	0.252							mg/L	
500PPBCN	500PPBCN				W	1/9/17						
Cyanide		PASS	498.804	0.499							mg/L	
ICV1	ICV1				W	1/9/17						
Cyanide		PASS	95.720	0.10		100.0	90	110			mg/L	
ICB1	ICB1				W	1/9/17						
Cyanide		PASS	-0.171	0.000				+/-0.0025			mg/L	
CCV1	CCV1				W	1/9/17						
Cyanide		PASS	243.903	0.24		96.0	90	110			mg/L	
CCB1	CCB1				W	1/9/17						
Cyanide		PASS	-0.459	0.000				+/-0.0025			mg/L	
PB95805BLS	PB95805BLS				S	1/9/17						
Cyanide		PASS	0.067	0.003				+/-0.0330			mg/Kg	
PB95805BSS	PB95805BSS				S	1/9/17						
Cyanide		PASS	98.087	4.90		98.0	85.00	115.00			mg/Kg	
LOWPB95805	LOWPB95805				W	1/9/17						
Cyanide		PASS	9.914	0.010							mg/L	
HIGHPB95805	HIGHPB95805				W	1/9/17						
Cyanide		PASS	498.293	0.498							mg/L	
I1045-02	P3-SP2				S	1/9/17						
Cyanide		PASS	1.788	0.196							mg/Kg	
I1045-03	RAP-SP1				S	1/9/17						
Cyanide		PASS	1.166	0.061							mg/Kg	
I1045-04	RAP-SP2				S	1/9/17						
Cyanide		PASS	-0.370	-0.039							mg/Kg	
I1045-05	RAP-SP3				S	1/9/17						
Cyanide		PASS	7.285	0.385							mg/Kg	
I1045-06	RAP-SP4				S	1/9/17						
Cyanide		PASS	-0.709	-0.073							mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/9/17
 Analyst : AK
 Data File : lb85244.csv

Approved By : 
 Approved Date : 1/10/17
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
CCV2	CCV2				W	1/9/17						
Cyanide		PASS	249.320		0.25	100.0	90	110			mg/L	
CCB2	CCB2				W	1/9/17						
Cyanide		PASS	-1.046		-0.001			+/-0.0025			mg/L	
I1050-02	POSTEX-29-20170105			1	S	1/9/17						
Cyanide		PASS	0.827		0.048						mg/Kg	
I1050-02D	POSTEX-29-20170105D			1	S	1/9/17						
Cyanide		FAIL	0.662		0.039				22.7	20	mg/Kg	
I1050-02S	POSTEX-29-20170105S			1	S	1/9/17						
Cyanide		PASS	40.008		2.390	98.0	75	125			mg/Kg	
I1045-01	P3-SP1			1	S	1/9/17						
Cyanide		PASS	1.845		0.098						mg/Kg	
I1050-01	POSTEX-28-20170105			1	S	1/9/17						
Cyanide		PASS	1.582		0.090						mg/Kg	
I1050-02SD	POSTEX-29-20170105SD			1	S	1/9/17						
Cyanide		PASS	41.480		2.460	102.0	75	125	2.9	20	mg/Kg	
CCV3	CCV3				W	1/9/17						
Cyanide		PASS	257.358		0.26	104.0	90	110			mg/L	
CCB3	CCB3				W	1/9/17						
Cyanide		PASS	-0.159		0.000			+/-0.0025			mg/L	