

LB85 258

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : Ale

1/10/2017 10:28

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.947	0.0	0.107	
ICB1	-0.001	0.0	0.004	
CCV1	1.052	0.0	0.118	
CCB1	-0.005	0.0	0.004	
PB95863BLS	-0.003	0.0	0.004	
PB95863BSS	0.949	0.0	0.107	
I1064-03	-0.007	0.0	0.004	
I1064-03D	-0.003	0.0	0.004	
I1064-03S	0.777	0.0	0.088	
I1064-03SD	0.774	0.0	0.088	
I1064-09	-0.002	0.0	0.004	
CCV2	1.021	0.0	0.115	
CCB2	-0.005	0.0	0.004	

N	13
Mean	0.423
SD	0.4854
CV%	114.84

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 10 09:36:08 2017

Tue Jan 10 10:24:29 2017

LB85 258

AK

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0491	mg/l	1/10/2017 9:36:08	
0.1PPM	A	Phenolics-	P	0.1023	mg/l	1/10/2017 9:36:09	
0.25PPM	A	Phenolics-	P	0.2533	mg/l	1/10/2017 9:36:10	
0.50PPM	A	Phenolics-	P	0.4963	mg/l	1/10/2017 9:36:11	
1.0PPM	A	Phenolics-	P	0.997	mg/l	1/10/2017 9:36:12	
2.0PPM	A	Phenolics-	P	2.0019	mg/l	1/10/2017 9:36:13	
ICV1	S	Phenolics-	P	0.9473	mg/l	1/10/2017 10:01:28	
ICB1	S	Phenolics-	P	-0.0005	mg/l	1/10/2017 10:01:29	
CCV1	S	Phenolics-	P	1.0516	mg/l	1/10/2017 10:01:30	
CCB1	S	Phenolics-	P	-0.0052	mg/l	1/10/2017 10:01:31	
PB95863BLS	S	Phenolics-	P	-0.0027	mg/l	1/10/2017 10:01:32	
PB95863BSS	S	Phenolics-	P	0.9495	mg/l	1/10/2017 10:01:33	
I1064-03	S	Phenolics-	P	-0.0074	mg/l	1/10/2017 10:01:34	
I1064-03D	S	Phenolics-	P	-0.0027	mg/l	1/10/2017 10:01:35	
I1064-03S	S	Phenolics-	P	0.777	mg/l	1/10/2017 10:01:36	
I1064-03SD	S	Phenolics-	P	0.7744	mg/l	1/10/2017 10:01:37	
I1064-09	S	Phenolics-	P	-0.0018	mg/l	1/10/2017 10:24:27	
CCV2	S	Phenolics-	P	1.0209	mg/l	1/10/2017 10:24:28	
CCB2	S	Phenolics-	P	-0.0054	mg/l	1/10/2017 10:24:29	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/10/2017 9:37


Test Phenolics-

Accepted 1/10/2017 9:37

Factor 9.278
Bias 0.004

Coeff. of det. 0.999985

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0491	0.0500	
2	Phenol-2	0.015	0.1023	0.1000	
3	Phenol-2	0.032	0.2533	0.2500	
4	Phenol-2	0.058	0.4963	0.5000	
5	Phenol-2	0.112	0.9970	1.0000	
6	Phenol-2	0.220	2.0019	2.0000	