

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/10/17

Analyst : AK

Data File : lb85258.csv

Approved By : 

Approved Date : 1/11/17

Worksheet # :

m9065-Phenolics-11

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc	%Rec		LCL	UCL	RPD	Max RPD	Units	Line 2
Phenol											
0.05PPM	0.05PPM			W	1/10/17						
Phenolics	PASS	0.049	0.049							mg/L	
0.1PPM	0.1PPM			W	1/10/17						
Phenolics	PASS	0.102	0.102							mg/L	
0.25PPM	0.25PPM			W	1/10/17						
Phenolics	PASS	0.253	0.253							mg/L	
0.50PPM	0.50PPM			W	1/10/17						
Phenolics	PASS	0.496	0.496							mg/L	
1.0PPM	1.0PPM			W	1/10/17						
Phenolics	PASS	0.997	0.997							mg/L	
2.0PPM	2.0PPM			W	1/10/17						
Phenolics	PASS	2.002	2.00							mg/L	
ICV1	ICV1			W	1/10/17						
Phenolics	PASS	0.947	0.95		95.0	90	110			mg/L	
ICB1	ICB1			W	1/10/17						
Phenolics	PASS	-0.001	-0.001				+/-0.0100			mg/L	
CCV1	CCV1			W	1/10/17						
Phenolics	PASS	1.052	1.05		105.0	90	110			mg/L	
CCB1	CCB1			W	1/10/17						
Phenolics	PASS	-0.005	-0.005				+/-0.0100			mg/L	
PB95863BLS	PB95863BLS			S	1/10/17						
Phenolics	PASS	-0.003	-0.135				+/-0.0780			mg/Kg	
PB95863BSS	PB95863BSS			S	1/10/17						
Phenolics	PASS	0.950	47.50		95.0	80.00	120.00			mg/Kg	
I1064-03	TP5-6-COMP11	1	S	1/10/17							
Phenolics	PASS	-0.007	-0.431							mg/Kg	
I1064-03D	TP5-6-COMP11D	1	S	1/10/17							
Phenolics	PASS	-0.003	-0.159					0	20	mg/Kg	
I1064-03S	TP5-6-COMP11S	1	S	1/10/17							
Phenolics	PASS	0.777	44.900		78.0	75	125			mg/Kg	
I1064-03SD	TP5-6-COMP11SD	1	S	1/10/17							
Phenolics	PASS	0.774	45.200		78.0	75	125	1	20	mg/Kg	
I1064-09	TP4-COMP12	1	S	1/10/17							
Phenolics	PASS	-0.002	-0.221							mg/Kg	
CCV2	CCV2			W	1/10/17						
Phenolics	PASS	1.021	1.02		102.0	90	110			mg/L	
CCB2	CCB2			W	1/10/17						
Phenolics	PASS	-0.005	-0.005				+/-0.0100			mg/L	