

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/11/17

Analyst : AK

Data File : lb85286.csv

Approved By : AK

Approved Date : 1/11/17

Worksheet # : 1/11/17

m9012A-Total Amenable & Reactive Cyanide-16

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1	
Parameter			PPB		Final Conc	%Rec	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide													
0.0PPBCN	0.0PPBCN				W	1/11/17							
Reactive Cyanide		PASS	0.939		0.001							mg/L	
5.0PPBCN	5.0PPBCN				W	1/11/17							
Reactive Cyanide		PASS	6.095		0.006							mg/L	
10PPBCN	10PPBCN				W	1/11/17							
Reactive Cyanide		PASS	10.332		0.010							mg/L	
50PPBCN	50PPBCN				W	1/11/17							
Reactive Cyanide		PASS	53.363		0.053							mg/L	
100PPBCN	100PPBCN				W	1/11/17							
Reactive Cyanide		PASS	99.548		0.100							mg/L	
250PPBCN	250PPBCN				W	1/11/17							
Reactive Cyanide		PASS	239.974		0.240							mg/L	
500PPBCN	500PPBCN				W	1/11/17							
Reactive Cyanide		PASS	504.750		0.505							mg/L	
ICV1	ICV1				W	1/11/17							
Reactive Cyanide		PASS	95.872		0.10	100.0	85	115				mg/L	
ICB1	ICB1				W	1/11/17							
Reactive Cyanide		PASS	1.195		0.001			+/-0.0050				mg/L	
CCV1	CCV1				W	1/11/17							
Reactive Cyanide		PASS	234.327		0.23	92.0	90	110				mg/L	
CCB1	CCB1				W	1/11/17							
Reactive Cyanide		PASS	3.216		0.003			+/-0.0050				mg/L	
PB95864BLS	PB95864BLS				S	1/11/17							
Reactive Cyanide		PASS	1.382		0.014			+/-0.0500				mg/Kg	
I1078-02	ROLL-OFF-0-2-COMP	1			S	1/11/17							
Reactive Cyanide		PASS	1.164		0.012							mg/Kg	
I1078-02D	ROLL-OFF-0-2-COMPD	1			S	1/11/17							
Reactive Cyanide		PASS	2.965		0.030					0	20	mg/Kg	
I1079-02	USDOT-617122-COMP-0-2	1			S	1/11/17							
Reactive Cyanide		PASS	1.139		0.011							mg/Kg	
PB95920BLS	PB95920BLS				S	1/11/17							
Reactive Cyanide		PASS	1.684		0.017			+/-0.0500				mg/Kg	
I1098-05	TP3-COMP13	1			S	1/11/17							
Reactive Cyanide		PASS	1.001		0.010							mg/Kg	
I1098-05D	TP3-COMP13D	1			S	1/11/17							
Reactive Cyanide		PASS	0.848		0.008					0	20	mg/Kg	
CCV2	CCV2				W	1/11/17							
Reactive Cyanide		PASS	242.312		0.24	96.0	90	110				mg/L	
CCB2	CCB2				W	1/11/17							
Reactive Cyanide		PASS	1.666		0.002			+/-0.0050				mg/L	