

LB85314

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/12/2017 14:24

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.028	0.0	0.115	
ICB1	-0.005	0.0	0.004	
CCV1	1.029	0.0	0.116	
CCB1	-0.006	0.0	0.004	
PB95890BLW	-0.007	0.0	0.004	
PB95890BSW	1.024	0.0	0.115	
I1092 -03	0.004	0.0	0.005	
I1092-04	0.104	0.0	0.016	
CCV2	1.031	0.0	0.116	
CCB2	-0.006	0.0	0.004	

N	10
Mean	0.420
SD	0.5246
CV%	124.97

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 12 14:21:16 2017

Thu Jan 12 14:21:25 2017

AL2

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0506	mg/l	1/12/2017 12:53:32	
0.1PPM	A	Phenolics-	P	0.0991	mg/l	1/12/2017 12:53:33	
0.25PPM	A	Phenolics-	P	0.2493	mg/l	1/12/2017 12:53:34	
0.50PPM	A	Phenolics-	P	0.4996	mg/l	1/12/2017 12:53:35	
1.0PPM	A	Phenolics-	P	1.0024	mg/l	1/12/2017 12:53:36	
2.0PPM	A	Phenolics-	P	1.999	mg/l	1/12/2017 12:53:37	
ICV1	S	Phenolics-	P	1.0283	mg/l	1/12/2017 14:21:16	
ICB1	S	Phenolics-	P	-0.0049	mg/l	1/12/2017 14:21:17	
CCV1	S	Phenolics-	P	1.0292	mg/l	1/12/2017 14:21:18	
CCB1	S	Phenolics-	P	-0.0055	mg/l	1/12/2017 14:21:19	
PB95890BLW	S	Phenolics-	P	-0.0065	mg/l	1/12/2017 14:21:20	
PB95890BSW	S	Phenolics-	P	1.0236	mg/l	1/12/2017 14:21:21	
I1092 -03	S	Phenolics-	P	0.0041	mg/l	1/12/2017 14:21:22	
I1092-04	S	Phenolics-	P	0.1044	mg/l	1/12/2017 14:21:23	
CCV2	S	Phenolics-	P	1.031	mg/l	1/12/2017 14:21:24	
CCB2	S	Phenolics-	P	-0.0061	mg/l	1/12/2017 14:21:25	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

1/12/2017 12:54


Test Phenolics-

Accepted 1/12/2017 12:54

Factor 9.28
Bias 0.005

Coeff. of det. 0.999997

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0506	0.0500	
2	Phenol-2	0.015	0.0991	0.1000	
3	Phenol-2	0.031	0.2493	0.2500	
4	Phenol-2	0.058	0.4996	0.5000	
5	Phenol-2	0.113	1.0024	1.0000	
6	Phenol-2	0.220	1.9990	2.0000	