

# Chemtech Consulting Group

## Analytical Review Report

Date Printed : 1/12/17

Analyst : AK

Data File : lb85314.csv

Approved By : 

Approved Date : 1/14/17

Worksheet # :

m420.1 - Phenolics - 9

| Lab Sample ID     | Client ID                      | Raw Amt | Dil      | Matrix     | A. Date        | Prep | Analysis  |     |         |       | Line 1 |
|-------------------|--------------------------------|---------|----------|------------|----------------|------|-----------|-----|---------|-------|--------|
| Parameter         |                                | PPB     |          | Final Conc | %Rec           | LCL  | UCL       | RPD | Max RPD | Units | Line 2 |
| <b>Phenol</b>     |                                |         |          |            |                |      |           |     |         |       |        |
| <b>0.05PPM</b>    | <b>0.05PPM</b>                 |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.051    | 0.051      |                |      |           |     |         | mg/L  |        |
| <b>0.1PPM</b>     | <b>0.1PPM</b>                  |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.099    | 0.099      |                |      |           |     |         | mg/L  |        |
| <b>0.25PPM</b>    | <b>0.25PPM</b>                 |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.249    | 0.249      |                |      |           |     |         | mg/L  |        |
| <b>0.50PPM</b>    | <b>0.50PPM</b>                 |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.500    | 0.500      |                |      |           |     |         | mg/L  |        |
| <b>1.0PPM</b>     | <b>1.0PPM</b>                  |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.002    | 1.00       |                |      |           |     |         | mg/L  |        |
| <b>2.0PPM</b>     | <b>2.0PPM</b>                  |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.999    | 2.00       |                |      |           |     |         | mg/L  |        |
| <b>ICV1</b>       | <b>ICV1</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.028    | 1.03       | 103.0          | 90   | 110       |     |         | mg/L  |        |
| <b>ICB1</b>       | <b>ICB1</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | -0.005   | -0.005     |                |      | +/-0.0020 |     |         | mg/L  |        |
| <b>CCV1</b>       | <b>CCV1</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.029    | 1.03       | 103.0          | 90   | 110       |     |         | mg/L  |        |
| <b>CCB1</b>       | <b>CCB1</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | -0.005   | -0.006     |                |      | +/-0.0020 |     |         | mg/L  |        |
| <b>PB95890BLW</b> | <b>PB95890BLW</b>              |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | -0.007   | -0.007     |                |      | +/-0.0020 |     |         | mg/L  |        |
| <b>PB95890BSW</b> | <b>PB95890BSW</b>              |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.024    | 1.02       | 102.0          | 80   | 120       |     |         | mg/L  |        |
| <b>I1092-03</b>   | <b>Annual -LOD- Water-2017</b> |         | <b>1</b> | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.004    | 0.004      |                |      |           |     |         | mg/L  |        |
| <b>I1092-04</b>   | <b>Annual -LOQ- Water-2017</b> |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 0.104    | 0.10       | 100.0          | 50   | 150       |     |         | mg/L  |        |
| <b>CCV2</b>       | <b>CCV2</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | 1.031    | 1.03       | 103.0          | 90   | 110       |     |         | mg/L  |        |
| <b>CCB2</b>       | <b>CCB2</b>                    |         |          | <b>W</b>   | <b>1/12/17</b> |      |           |     |         |       |        |
| Phenolics         |                                | PASS    | -0.006   | -0.006     |                |      | +/-0.0020 |     |         | mg/L  |        |