

LB85315

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,

Mountainside, NJ 07092

Reviewed by : du

1/12/2017 13:55

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.027	0.0	0.115	
ICB1	-0.001	0.0	0.005	
CCV1	1.025	0.0	0.115	
CCB1	-0.001	0.0	0.004	
PB95933BLS	0.002	0.0	0.005	
PB95933BSS	1.034	0.0	0.116	
I1098-03	0.003	0.0	0.005	
I1098-03D	0.001	0.0	0.005	
I1098-03S	0.952	0.0	0.107	
I1098-03SD	0.948	0.0	0.107	
CCV2	1.043	0.0	0.117	
CCB2	0.002	0.0	0.005	

N	12
Mean	0.503
SD	0.5251
CV%	104.43

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 12 12:53:32 2017

Thu Jan 12 13:54:45 2017

A K

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0506	mg/l	1/12/2017 12:53:32	
0.1PPM	A	Phenolics-	P	0.0991	mg/l	1/12/2017 12:53:33	
0.25PPM	A	Phenolics-	P	0.2493	mg/l	1/12/2017 12:53:34	
0.50PPM	A	Phenolics-	P	0.4996	mg/l	1/12/2017 12:53:35	
1.0PPM	A	Phenolics-	P	1.0024	mg/l	1/12/2017 12:53:36	
2.0PPM	A	Phenolics-	P	1.999	mg/l	1/12/2017 12:53:37	
ICV1	S	Phenolics-	P	1.0271	mg/l	1/12/2017 13:26:05	
ICB1	S	Phenolics-	P	-0.0008	mg/l	1/12/2017 13:26:06	
CCV1	S	Phenolics-	P	1.025	mg/l	1/12/2017 13:26:07	
CCB1	S	Phenolics-	P	-0.0014	mg/l	1/12/2017 13:26:08	
PB95933BLS	S	Phenolics-	P	0.002	mg/l	1/12/2017 13:26:09	
PB95933BSS	S	Phenolics-	P	1.0339	mg/l	1/12/2017 13:26:10	
I1098-03	S	Phenolics-	P	0.0026	mg/l	1/12/2017 13:54:40	
I1098-03D	S	Phenolics-	P	0.0008	mg/l	1/12/2017 13:54:41	
I1098-03S	S	Phenolics-	P	0.9519	mg/l	1/12/2017 13:54:42	
I1098-03SD	S	Phenolics-	P	0.9477	mg/l	1/12/2017 13:54:43	
CCV2	S	Phenolics-	P	1.043	mg/l	1/12/2017 13:54:44	
CCB2	S	Phenolics-	P	0.0017	mg/l	1/12/2017 13:54:45	

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/12/2017 12:54


Test Phenolics-

Accepted 1/12/2017 12:54

Factor 9.28
Bias 0.005

Coeff. of det. 0.999997

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0506	0.0500	
2	Phenol-2	0.015	0.0991	0.1000	
3	Phenol-2	0.031	0.2493	0.2500	
4	Phenol-2	0.058	0.4996	0.5000	
5	Phenol-2	0.113	1.0024	1.0000	
6	Phenol-2	0.220	1.9990	2.0000	