

# Chemtech Consulting Group

## Analytical Review Report

Date Printed : 1/12/17

Analyst : AK

Data File : lb85315.csv

Approved By : 

Approved Date : 1/14/17

Worksheet # :

m9065-Phenolics-11

| Lab Sample ID | Client ID    |      | Raw Amt | Dil | Matrix     | A. Date | Prep  | Analysis  |     |         |       | Line 1 |
|---------------|--------------|------|---------|-----|------------|---------|-------|-----------|-----|---------|-------|--------|
| Parameter     |              |      | PPB     |     | Final Conc | %Rec    | LCL   | UCL       | RPD | Max RPD | Units | Line 2 |
|               |              |      |         |     |            |         |       |           |     |         |       |        |
| Phenol        |              |      |         |     |            |         |       |           |     |         |       |        |
| 0.05PPM       | 0.05PPM      |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.051   |     | 0.051      |         |       |           |     |         | mg/L  |        |
| 0.1PPM        | 0.1PPM       |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.099   |     | 0.099      |         |       |           |     |         | mg/L  |        |
| 0.25PPM       | 0.25PPM      |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.249   |     | 0.249      |         |       |           |     |         | mg/L  |        |
| 0.50PPM       | 0.50PPM      |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.500   |     | 0.500      |         |       |           |     |         | mg/L  |        |
| 1.0PPM        | 1.0PPM       |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.002   |     | 1.00       |         |       |           |     |         | mg/L  |        |
| 2.0PPM        | 2.0PPM       |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.999   |     | 2.00       |         |       |           |     |         | mg/L  |        |
| ICV1          | ICV1         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.027   |     | 1.03       | 103.0   | 90    | 110       |     |         | mg/L  |        |
| ICB1          | ICB1         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | -0.001  |     | -0.001     |         |       | +/-0.0100 |     |         | mg/L  |        |
| CCV1          | CCV1         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.025   |     | 1.02       | 102.0   | 90    | 110       |     |         | mg/L  |        |
| CCB1          | CCB1         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | -0.001  |     | -0.001     |         |       | +/-0.0100 |     |         | mg/L  |        |
| PB95933BLS    | PB95933BLS   |      |         |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.002   |     | 0.100      |         |       | +/-0.0780 |     |         | mg/Kg |        |
| PB95933BSS    | PB95933BSS   |      |         |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.034   |     | 51.70      | 103.0   | 80.00 | 120.00    |     |         | mg/Kg |        |
| I1098-03      | TP3-COMP13   |      | 1       |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.003   |     | 0.151      |         |       |           |     |         | mg/Kg |        |
| I1098-03D     | TP3-COMP13D  |      | 1       |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | FAIL | 0.001   |     | 0.046      |         |       |           | 200 | 20      | mg/Kg |        |
| I1098-03S     | TP3-COMP13S  |      | 1       |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.952   |     | 52.500     | 95.0    | 75    | 125       |     |         | mg/Kg |        |
| I1098-03SD    | TP3-COMP13SD |      | 1       |     | S          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.948   |     | 52.800     | 95.0    | 75    | 125       | 1   | 20      | mg/Kg |        |
| CCV2          | CCV2         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 1.043   |     | 1.04       | 104.0   | 90    | 110       |     |         | mg/L  |        |
| CCB2          | CCB2         |      |         |     | W          | 1/12/17 |       |           |     |         |       |        |
| Phenolics     |              | PASS | 0.002   |     | 0.002      |         |       | +/-0.0100 |     |         | mg/L  |        |