

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : *[Signature]*
 Data File : lb85357.csv


Approved By : *[Signature]*
 Approved Date : 1/17/17
 Worksheet # :



msm45cm - 1712 - B.G. 14 Ammonia - 131

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB		Final Conc	%Rec	Method	Method	LCL	UCL	Line 2
Ammonia										
0.1PPM	0.1PPM			W	1/16/17					
Ammonia as N	PASS	0.110	0.110							mg/L
0.2PPM	0.2PPM			W	1/16/17					
Ammonia as N	PASS	0.203	0.203							mg/L
0.4PPM	0.4PPM			W	1/16/17					
Ammonia as N	PASS	0.384	0.384							mg/L
1.0PPM	1.0PPM			W	1/16/17					
Ammonia as N	PASS	0.998	0.998							mg/L
1.3PPM	1.3PPM			W	1/16/17					
Ammonia as N	PASS	1.337	1.34							mg/L
2.0PPM	2.0PPM			W	1/16/17					
Ammonia as N	PASS	2.000	2.00							mg/L
ICV1	ICV1			W	1/16/17					
Ammonia as N	PASS	1.005	1.00	100.0		90	110			mg/L
ICB1	ICB1			W	1/16/17					
Ammonia as N	PASS	0.013	0.013				+/-0.0340			mg/L
CCV1	CCV1			W	1/16/17					
Ammonia as N	PASS	0.986	0.99	99.0		90	110			mg/L
CCB1	CCB1			W	1/16/17					
Ammonia as N	PASS	0.009	0.010				+/-0.0340			mg/L
PB95977BLS	PB95977BLS			S	1/16/17					
Ammonia as N	PASS	0.012	0.600				+/-0.0100			mg/Kg
PB95977BSS	PB95977BSS			S	1/16/17					
Ammonia as N	PASS	1.044	52.20	104.0		80.00	120.00			mg/Kg
I1092-01	Annual LOD-Soil 2017	1	S	1/16/17						
Ammonia as N	PASS	0.075	3.70							mg/Kg
I1092-02	Annual LOQ-Soil 2017		S	1/16/17						
Ammonia as N	PASS	0.196	9.80	98.0		50.00	150.00			mg/Kg
PB95978BLW	PB95978BLW			W	1/16/17					
Ammonia as N	PASS	0.009	0.009				+/-0.0340			mg/L
PB95978BSW	PB95978BSW			W	1/16/17					
Ammonia as N	PASS	1.037	1.04	104.0		80	120			mg/L
I1092-03	Annual -LOD- Water-2017	1	W	1/16/17						
Ammonia as N	PASS	0.079	0.080							mg/L
I1092-04	Annual -LOQ- Water-2017		W	1/16/17						
Ammonia as N	PASS	0.198	0.20	100.0		50	150			mg/L
I1124-01	001M-WILLETS-PTBLVD(JA	1	W	1/16/17						
Ammonia as N	OR	3.099	3.10							mg/L
I1124-01D	001M-WILLETS-PTBLVD(JA	1	W	1/16/17						
Ammonia as N	PASS	2.988	2.990				3.6	20		mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : 
 Data File : lb85357.csv

Approved By : 
 Approved Date : 
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Ammonia											
CCV2	CCV2			W	1/16/17						
Ammonia as N	PASS	1.047		1.05	105.0	90	110			mg/L	
CCB2	CCB2			W	1/16/17						
Ammonia as N	PASS	0.025		0.025			+/-0.0340			mg/L	
I1124-01S	001M-WILLETS-PTBLVD(JA	1	W	1/16/17							
Ammonia as N	PASS	4.335		4.330	123.0	75	125			mg/L	
I1124-01SD	001M-WILLETS-PTBLVD(JA	1	W	1/16/17							
Ammonia as N	PASS	4.325		4.320	122.0	75	125	0.23	20	mg/L	
I1124-02	002M-35THAVE(JAN)	1	W	1/16/17							
Ammonia as N	OR	3.049		3.05						mg/L	
I1144-01	FINAL-EFFLUENT-19B	1	W	1/16/17							
Ammonia as N	OR	4.071		4.07						mg/L	
CCV3	CCV3			W	1/16/17						
Ammonia as N	PASS	0.982		0.98	98.0	90	110			mg/L	
CCB3	CCB3			W	1/16/17						
Ammonia as N	PASS	0.014		0.014			+/-0.0340			mg/L	
I1124-01DL	001M-WILLETS-PTBLVD(JA	5	W	1/16/17							
Ammonia as N	PASS	0.632		3.16						mg/L	
I1124-01DDL	001M-WILLETS-PTBLVD(JA	5	W	1/16/17							
Ammonia as N	PASS	0.616		3.080				2.6	20	mg/L	
I1124-01SDL	001M-WILLETS-PTBLVD(JA	5	W	1/16/17							
Ammonia as N	PASS	0.821		4.100	94.0	75	125			mg/L	
I1124-01SDDL	001M-WILLETS-PTBLVD(JA	5	W	1/16/17							
Ammonia as N	PASS	0.819		4.100	94.0	75	125	0	20	mg/L	
I1124-02DL	002M-35THAVE(JAN)DL	5	W	1/16/17							
Ammonia as N	PASS	0.585		2.92						mg/L	
I1144-01DL	FINAL-EFFLUENT-19BDL	5	W	1/16/17							
Ammonia as N	PASS	1.039		5.20						mg/L	
CCV4	CCV4			W	1/16/17						
Ammonia as N	PASS	0.948		0.95	95.0	90	110			mg/L	
CCB4	CCB4			W	1/16/17						
Ammonia as N	PASS	0.004		0.004			+/-0.0340			mg/L	