

1885363

=====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Alc

1/17/2017 10:46

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.010	0.0	0.093	
ICB1	1.101	0.0	0.005	
CCV1	243.638	0.0	0.232	
CCB1	1.262	0.0	0.005	
PB96027BLW	1.081	0.0	0.005	
I1219-03	0.848	0.0	0.004	
I1219-03D	1.567	0.0	0.005	
CCV2	246.830	0.0	0.235	
CCB2	1.647	0.0	0.005	

N 9
Mean 65.887
SD 106.2142
CV% 161.21

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 17 10:45:43 2017

Tue Jan 17 10:45:51 2017

AK

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Reactive CN	P	1.3155	µg/l	1/17/2017 9:16:06
5.OPPBCN	A	Reactive CN	P	5.7637	µg/l	1/17/2017 9:16:07
10PPBCN	A	Reactive CN	P	10.8561	µg/l	1/17/2017 9:16:08
50PPBCN	A	Reactive CN	P	52.3963	µg/l	1/17/2017 9:16:09
100PPBCN	A	Reactive CN	P	97.9973	µg/l	1/17/2017 9:16:10
250PPBCN	A	Reactive CN	P	243.0699	µg/l	1/17/2017 9:16:11
500PPBCN	A	Reactive CN	P	503.6011	µg/l	1/17/2017 9:16:12
ICV1	S	Reactive CN	P	95.01	µg/l	1/17/2017 10:45:43
ICB1	S	Reactive CN	P	1.1013	µg/l	1/17/2017 10:45:44
CCV1	S	Reactive CN	P	243.6383	µg/l	1/17/2017 10:45:45
CCB1	S	Reactive CN	P	1.2616	µg/l	1/17/2017 10:45:46
PB96027BLW	S	Reactive CN	P	1.0813	µg/l	1/17/2017 10:45:47
I1219-03	S	Reactive CN	P	0.8476	µg/l	1/17/2017 10:45:48
I1219-03D	S	Reactive CN	P	1.5667	µg/l	1/17/2017 10:45:49
CCV2	S	Reactive CN	P	246.8304	µg/l	1/17/2017 10:45:50
CCB2	S	Reactive CN	P	1.6467	µg/l	1/17/2017 10:45:51

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : dk

1/17/2017 9:18


Test Total CN

Accepted 1/17/2017 9:18

Factor 1066
Bias 0.004

Coeff. of det. 0.999641

Errors



Calibrator	Response	Calc. con.	Conc.	Errors
1	0.005	1.3155	0.0000	
2	0.009	5.7637	5.0000	
3	0.014	10.8561	10.0000	
4	0.053	52.3963	50.0000	
5	0.096	97.9973	100.0000	
6	0.232	243.0699	250.0000	
7	0.476	503.6011	500.0000	