

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AK
 Data File : lb85363.csv

Approved By : *[Signature]*
 Approved Date : 1/17/17
 Worksheet # :

m9012 A-B Total Available @ Reactive Cyanide - 1.6

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	1/17/17						
Reactive Cyanide		PASS	1.316		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/17/17						
Reactive Cyanide		PASS	5.764		0.006						mg/L	
10PPBCN	10PPBCN				W	1/17/17						
Reactive Cyanide		PASS	10.856		0.011						mg/L	
50PPBCN	50PPBCN				W	1/17/17						
Reactive Cyanide		PASS	52.396		0.052						mg/L	
100PPBCN	100PPBCN				W	1/17/17						
Reactive Cyanide		PASS	97.997		0.098						mg/L	
250PPBCN	250PPBCN				W	1/17/17						
Reactive Cyanide		PASS	243.070		0.243						mg/L	
500PPBCN	500PPBCN				W	1/17/17						
Reactive Cyanide		PASS	503.601		0.504						mg/L	
ICV1	ICV1				W	1/17/17						
Reactive Cyanide		PASS	95.010		0.10	100.0	85	115			mg/L	
ICB1	ICB1				W	1/17/17						
Reactive Cyanide		PASS	1.101		0.001			+/-0.0050			mg/L	
CCV1	CCV1				W	1/17/17						
Reactive Cyanide		PASS	243.638		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/17/17						
Reactive Cyanide		PASS	1.262		0.001			+/-0.0050			mg/L	
PB96027BLW	PB96027BLW				W	1/17/17						
Reactive Cyanide		PASS	1.081		0.001			+/-0.0050			mg/L	
I1219-03	WC011217-LIQUID-TODT-HI	1			W	1/17/17						
Reactive Cyanide		PASS	0.848		0.001						mg/L	
I1219-03D	WC011217-LIQUID-TODT-HI	1			W	1/17/17						
Reactive Cyanide		PASS	1.567		0.002				0	20	mg/L	
CCV2	CCV2				W	1/17/17						
Reactive Cyanide		PASS	246.830		0.25	100.0	90	110			mg/L	
CCB2	CCB2				W	1/17/17						
Reactive Cyanide		PASS	1.647		0.002			+/-0.0050			mg/L	