


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17

Approved By : 

Analyst : AK

Approved Date : 1/17/17

Data File : lb85370.csv

Worksheet # :

m9012 A-B Total Amenable & Reactive Cyanide-16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc		%Rec	Method	Method	LCL	UCL	RPD	Line 2
Cyanide											
0.0PPBCN	0.0PPBCN			W	1/17/17						
Cyanide	PASS	1.316	0.001							mg/L	
5.0PPBCN	5.0PPBCN			W	1/17/17						
Cyanide	PASS	5.764	0.006							mg/L	
10PPBCN	10PPBCN			W	1/17/17						
Cyanide	PASS	10.856	0.011							mg/L	
50PPBCN	50PPBCN			W	1/17/17						
Cyanide	PASS	52.396	0.052							mg/L	
100PPBCN	100PPBCN			W	1/17/17						
Cyanide	PASS	97.997	0.098							mg/L	
250PPBCN	250PPBCN			W	1/17/17						
Cyanide	PASS	243.070	0.243							mg/L	
500PPBCN	500PPBCN			W	1/17/17						
Cyanide	PASS	503.601	0.504							mg/L	
ICV1	ICV1			W	1/17/17						
Cyanide	PASS	94.447	0.09	90.0		90	110			mg/L	
ICB1	ICB1			W	1/17/17						
Cyanide	PASS	1.652	0.002				+/-0.0030			mg/L	
CCV1	CCV1			W	1/17/17						
Cyanide	PASS	240.973	0.24	96.0		90	110			mg/L	
CCB1	CCB1			W	1/17/17						
Cyanide	PASS	1.322	0.001				+/-0.0030			mg/L	
PB95953BLS	PB95953BLS			S	1/17/17						
Cyanide	PASS	1.065	0.053				+/-0.0330			mg/Kg	
PB95953BSS	PB95953BSS			S	1/17/17						
Cyanide	PASS	95.081	4.75	95.0		85.00	115.00			mg/Kg	
LOWPB95953	LOWPB95953			W	1/17/17						
Cyanide	PASS	11.353	0.011							mg/L	
HIGHPB95953	HIGHPB95953			W	1/17/17						
Cyanide	PASS	500.856	0.501							mg/L	
I1147-01	TOD-SB-01-0-2		1	S	1/17/17						
Cyanide	PASS	1.879	0.103							mg/Kg	
I1147-01D	TOD-SB-01-0-2D		1	S	1/17/17						
Cyanide	PASS	1.700	0.094					9.1	20	mg/Kg	
I1147-02S	TOD-SB-01-0-2S		1	S	1/17/17						
Cyanide	FAIL	29.153	1.580	68.0		75	125			mg/Kg	
I1147-03SD	TOD-SB-01-0-2SD		1	S	1/17/17						
Cyanide	FAIL	23.563	1.290	54.0		75	125	20.2	20	mg/Kg	
I1147-04	TOD-SB-101-0-2		1	S	1/17/17						
Cyanide	PASS	1.324	0.071							mg/Kg	

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AK
 Data File : lb85370.csv

Approved By : [Signature]
 Approved Date : 1/18/17
 Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter		PPB	Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units
										Line 2
Cyanide										
I1182-01	TOD-BC-0.2		1	S	1/17/17					
Cyanide		PASS	3.114	0.177						mg/Kg
CCV2	CCV2			W	1/17/17					
Cyanide		PASS	242.659	0.24	96.0	90	110			mg/L
CCB2	CCB2			W	1/17/17					
Cyanide		PASS	1.393	0.001			+/-0.0030			mg/L
I1182-04	TOD-SB-78-80		1	S	1/17/17					
Cyanide		PASS	1.408	0.073						mg/Kg
PB96017BLW	PB96017BLW			W	1/17/17					
Cyanide		PASS	1.816	0.002			+/-0.0030			mg/L
PB96017BSW	PB96017BSW			W	1/17/17					
Cyanide		PASS	95.670	0.10	100.0	85	115			mg/L
I1182-06	FB011217-TOD		1	W	1/17/17					
Cyanide		PASS	1.551	0.002						mg/L
I1182-06D	FB011217-TODD		1	W	1/17/17					
Cyanide		FAIL	1.275	0.001				66.7	20	mg/L
I1182-06S	FB011217-TODS		1	W	1/17/17					
Cyanide		PASS	36.222	0.036	90.0	75	125			mg/L
I1182-06SD	FB011217-TODSD		1	W	1/17/17					
Cyanide		PASS	35.874	0.036	90.0	75	125	0	20	mg/L
I1219-02	WC011217-LIQUID-TODT-HILL		1	W	1/17/17					
Cyanide		PASS	1.543	0.002						mg/L
CCV3	CCV3			W	1/17/17					
Cyanide		PASS	249.968	0.25	100.0	90	110			mg/L
CCB3	CCB3			W	1/17/17					
Cyanide		PASS	1.538	0.002			+/-0.0030			mg/L
I1182-03	TOD-SB-23-25		1	S	1/17/17					
Cyanide		PASS	1.923	0.104						mg/Kg
CCV4	CCV4			W	1/17/17					
Cyanide		PASS	256.808	0.26	104.0	90	110			mg/L
CCB4	CCB4			W	1/17/17					
Cyanide		PASS	1.801	0.002			+/-0.0030			mg/L