

CB85371

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/17/2017 13:38

Test: Reactive AK
Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.566	0.0	0.092	
ICB1	1.638	0.0	0.005	
CCV1	237.369	0.0	0.226	
CCB1	1.488	0.0	0.005	
PB96069BLS	1.482	0.0	0.005	
I1246-03	1.020	0.0	0.005	
I1246-03D	1.108	0.0	0.005	
CCV2	245.262	0.0	0.234	
CCB2	1.866	0.0	0.005	

N	9
Mean	65.089
SD	104.4773
CV%	160.52

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 17 13:35:49 2017

Tue Jan 17 13:35:57 2017

AK

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	1.3155	µg/l	1/17/2017 9:16:06	
5.OPPBCN	A	Reactive CN	P	5.7637	µg/l	1/17/2017 9:16:07	
10PPBCN	A	Reactive CN	P	10.8561	µg/l	1/17/2017 9:16:08	
50PPBCN	A	Reactive CN	P	52.3963	µg/l	1/17/2017 9:16:09	
100PPBCN	A	Reactive CN	P	97.9973	µg/l	1/17/2017 9:16:10	
250PPBCN	A	Reactive CN	P	243.0699	µg/l	1/17/2017 9:16:11	
500PPBCN	A	Reactive CN	P	503.6011	µg/l	1/17/2017 9:16:12	
ICV1	S	Reactive CN	P	94.5663	µg/l	1/17/2017 13:35:49	
ICB1	S	Reactive CN	P	1.6375	µg/l	1/17/2017 13:35:50	
CCV1	S	Reactive CN	P	237.369	µg/l	1/17/2017 13:35:51	
CCB1	S	Reactive CN	P	1.4877	µg/l	1/17/2017 13:35:52	
PB96069BLS	S	Reactive CN	P	1.4821	µg/l	1/17/2017 13:35:53	
I1246-03	S	Reactive CN	P	1.02	µg/l	1/17/2017 13:35:54	
I1246-03D	S	Reactive CN	P	1.1085	µg/l	1/17/2017 13:35:55	
CCV2	S	Reactive CN	P	245.2616	µg/l	1/17/2017 13:35:56	
CCB2	S	Reactive CN	P	1.8661	µg/l	1/17/2017 13:35:57	

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/17/2017 9:18


Test Total CN

Accepted 1/17/2017 9:18

Factor 1066
Bias 0.004

Coeff. of det. 0.999641

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	1.3155	0.0000	
2	5.0PPBCN	0.009	5.7637	5.0000	
3	10PPBCN	0.014	10.8561	10.0000	
4	50PPBCN	0.053	52.3963	50.0000	
5	100PPBCN	0.096	97.9973	100.0000	
6	250PPBCN	0.232	243.0699	250.0000	
7	500PPBCN	0.476	503.6011	500.0000	