

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17

Approved By : 

Analyst : AK

Approved Date : 1/17/17

Data File : lb85371.csv

Worksheet # : 1

m9012-4-B Total Amenable of Reactive Cyanide - 16

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc		%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide											
0.0PPBCN	0.0PPBCN			W	1/17/17						
Reactive Cyanide		PASS	1.316	0.001						mg/L	
5.0PPBCN	5.0PPBCN			W	1/17/17						
Reactive Cyanide		PASS	5.764	0.006						mg/L	
10PPBCN	10PPBCN			W	1/17/17						
Reactive Cyanide		PASS	10.856	0.011						mg/L	
50PPBCN	50PPBCN			W	1/17/17						
Reactive Cyanide		PASS	52.396	0.052						mg/L	
100PPBCN	100PPBCN			W	1/17/17						
Reactive Cyanide		PASS	97.997	0.098						mg/L	
250PPBCN	250PPBCN			W	1/17/17						
Reactive Cyanide		PASS	243.070	0.243						mg/L	
500PPBCN	500PPBCN			W	1/17/17						
Reactive Cyanide		PASS	503.601	0.504						mg/L	
ICV1	ICV1			W	1/17/17						
Reactive Cyanide		PASS	94.566	0.09	90.0	85	115			mg/L	
ICB1	ICB1			W	1/17/17						
Reactive Cyanide		PASS	1.638	0.002			+/-0.0050			mg/L	
CCV1	CCV1			W	1/17/17						
Reactive Cyanide		PASS	237.369	0.24	96.0	90	110			mg/L	
CCB1	CCB1			W	1/17/17						
Reactive Cyanide		PASS	1.488	0.001			+/-0.0050			mg/L	
PB96069BLS	PB96069BLS			S	1/17/17						
Reactive Cyanide		PASS	1.482	0.015			+/-0.0500			mg/Kg	
I1246-03	ROW-SP-0-2-COMP		1	S	1/17/17						
Reactive Cyanide		PASS	1.020	0.010						mg/Kg	
I1246-03D	ROW-SP-0-2-COMPD		1	S	1/17/17						
Reactive Cyanide		PASS	1.109	0.011				0	20	mg/Kg	
CCV2	CCV2			W	1/17/17						
Reactive Cyanide		PASS	245.262	0.25	100.0	90	110			mg/L	
CCB2	CCB2			W	1/17/17						
Reactive Cyanide		PASS	1.866	0.002			+/-0.0050			mg/L	