

LB85379

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85379
Instrument: Appolo-9000

m9060-A-TOC-18

Analytical Summary Report

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB85379
Instrument: Appolo-9000


Seq	Lab ID	Client ID	Sample Type	Inst Conc. (ppm/ppb)*	Dil	Analytical Date
1	0.0PPM	0.0PPM	CAL	5652.000	1	12/29/16
2	0.5PPM	0.5PPM	CAL	27564.000	1	12/29/16
3	1.0PPM	1.0PPM	CAL	45665.000	1	12/29/16
4	2.0PPM	2.0PPM	CAL	87909.000	1	12/29/16
5	5.0PPM	5.0PPM	CAL	180418.000	1	12/29/16
6	10.0PPM	10.0PPM	CAL	346324.000	1	12/29/16
7	20.0PPM	20.0PPM	CAL	702526.000	1	12/29/16
8	ICV	ICV	ICV	10.291	1	12/29/16
9	ICB	ICB	ICB	0.000	1	12/29/16
10	CCV	CCV	CCV	10.659	1	1/17/17
11	CCB	CCB	CCB	0.000	1	1/17/17
12	LB85379BLW	LB85379BLW	MB	0.000	1	1/17/17
13	LB85379BSW	LB85379BSW	LCS	10.444	1	1/17/17
14	I1195-01	S45-0036	SAM	1.966	1	1/17/17
15	I1195-04	S45-0037	SAM	3.398	1	1/17/17
16	I1195-05	S45-0040	SAM	2.016	1	1/17/17
17	I1195-06	S45-0038	SAM	8.532	1	1/17/17
18	I1195-07	S45-0039	SAM	8.893	1	1/17/17
19	I1195-02S	S45-0036S	SAM	13.123	1	1/17/17
20	I1195-03SD	S45-0036SD	SAM	13.021	1	1/17/17
21	I1195-08	S45-EB-2017-1	SAM	0.444	1	1/17/17
22	CCV	CCV	CCV	10.119	1	1/17/17
23	CCB	CCB	CCB	0.000	1	1/17/17

Cal results are in raw amount.

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/17/17
 Analyst : AK
 Data File : LB85379.MDB

Approved By : 
 Approved Date :
 Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis		Line 1
Parameter			PPB		Final Conc	%Rec	Method	Method		Line 2
TOC										
0.0PPM	0.0PPM				W	12/29/16				
TOC		PASS	5652.000		5650				mg/L	
0.5PPM	0.5PPM				W	12/29/16				
TOC		PASS	27564.000		27600				mg/L	
1.0PPM	1.0PPM				W	12/29/16				
TOC		PASS	45665.000		45700				mg/L	
2.0PPM	2.0PPM				W	12/29/16				
TOC		PASS	87909.000		87900				mg/L	
5.0PPM	5.0PPM				W	12/29/16				
TOC		PASS	180418.000		180000				mg/L	
10.0PPM	10.0PPM				W	12/29/16				
TOC		PASS	346324.000		346000				mg/L	
20.0PPM	20.0PPM				W	12/29/16				
TOC		PASS	702526.000		703000				mg/L	
ICV1	ICV1				W	12/29/16				
TOC		PASS	10.291		10.30	103.0	90	110	mg/L	
ICB1	ICB1				W	12/29/16				
TOC		PASS	0.000		0.000			+/-0.0970	mg/L	
CCV1	CCV1				W	1/17/17				
TOC		PASS	10.659		10.70	107.0	90	110	mg/L	
CCB1	CCB1				W	1/17/17				
TOC		PASS	0.000		0.000			+/-0.0970	mg/L	
LB85379BLW	LB85379BLW				W	1/17/17				
TOC		PASS	0.000		0.000			+/-0.0970	mg/L	
LB85379BSW	LB85379BSW				W	1/17/17				
TOC		PASS	10.444		10.40	104.0	90	110	mg/L	
I1195-01	S45-0036			1	W	1/17/17				
TOC		PASS	1.966		1.97				mg/L	
I1195-04	S45-0037			1	W	1/17/17				
TOC		PASS	3.398		3.40				mg/L	
I1195-05	S45-0040			1	W	1/17/17				
TOC		PASS	2.016		2.02				mg/L	
I1195-06	S45-0038			1	W	1/17/17				
TOC		PASS	8.532		8.53				mg/L	
I1195-07	S45-0039			1	W	1/17/17				
TOC		PASS	8.893		8.89				mg/L	
I1195-02S	S45-0036S			1	W	1/17/17				
TOC		PASS	13.123		13.10	111.0	75	125	mg/L	
I1195-03SD	S45-0036SD			1	W	1/17/17				
TOC		PASS	13.021		13.00	110.0	75	125	0.8	20
									mg/L	

Cal results are
in raw
amount!

Chemtech Consulting Group

Analytical Review Report


Date Printed : 1/17/17

Analyst : AK

Data File : LB85379.MDB

Approved By : 

Approved Date : 1/17/17

Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Line 1											
Line 2											
TOC											
I1195-08	S45-EB-2017-1			1	W	1/17/17					
TOC		PASS	0.444		0.444						mg/L
CCV2	CCV2				W	1/17/17					
TOC		PASS	10.119		10.10	101.0	90	110			mg/L
CCB2	CCB2				W	1/17/17					
TOC		PASS	0.000		0.000			+/-0.0970			mg/L