


Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/18/17

Analyst : AK

Data File : lb85399.csv

Approved By : 

Approved Date : 1/18/17

Worksheet # :

m9012 A-B Total Amenable & Reactive Cyanide-10

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide												
0.0PPBCN	0.0PPBCN				W	1/18/17						
Cyanide		PASS	0.774		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/18/17						
Cyanide		PASS	5.658		0.006						mg/L	
10PPBCN	10PPBCN				W	1/18/17						
Cyanide		PASS	10.207		0.010						mg/L	
50PPBCN	50PPBCN				W	1/18/17						
Cyanide		PASS	51.638		0.052						mg/L	
100PPBCN	100PPBCN				W	1/18/17						
Cyanide		PASS	97.687		0.098						mg/L	
250PPBCN	250PPBCN				W	1/18/17						
Cyanide		PASS	247.498		0.247						mg/L	
500PPBCN	500PPBCN				W	1/18/17						
Cyanide		PASS	501.539		0.502						mg/L	
ICV1	ICV1				W	1/18/17						
Cyanide		PASS	90.820		0.09	90.0	90	110			mg/L	
ICB1	ICB1				W	1/18/17						
Cyanide		PASS	1.635		0.002			+/-0.0030			mg/L	
CCV1	CCV1				W	1/18/17						
Cyanide		PASS	236.832		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/18/17						
Cyanide		PASS	1.451		0.001			+/-0.0030			mg/L	
PB96086BLS	PB96086BLS				S	1/18/17						
Cyanide		PASS	0.975		0.048			+/-0.0330			mg/Kg	
PB96086BSS	PB96086BSS				S	1/18/17						
Cyanide		PASS	94.326		4.62	92.0	85.00	115.00			mg/Kg	
LOWPB96086	LOWPB96086				W	1/18/17						
Cyanide		PASS	10.534		0.011						mg/L	
HIGHPB96086	HIGHPB96086				W	1/18/17						
Cyanide		PASS	495.916		0.496						mg/L	
I1243-01	PHOENIX-BPM			1	S	1/18/17						
Cyanide		PASS	1.188		0.063						mg/Kg	
I1243-01D	PHOENIX-BPMD			1	S	1/18/17						
Cyanide		FAIL	0.234		0.013				200.00	20	mg/Kg	
I1243-01S	PHOENIX-BPMS			1	S	1/18/17						
Cyanide		FAIL	28.302		1.510	68.0	75	125			mg/Kg	
I1243-01SD	PHOENIX-BPMSD			1	S	1/18/17						
Cyanide		FAIL	28.846		1.560	69.0	75	125	3.3	20	mg/Kg	
I1246-01	ROW-SP-0-2-COMP			1	S	1/18/17						
Cyanide		PASS	0.580		0.031						mg/Kg	

Chemtech Consulting Group

Analytical Review Report


Date Printed : 1/18/17

Analyst : AK

Data File : lb85399.csv

Approved By : 

Approved Date : 1/18/17

Worksheet # : 

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Cyanide											
I1260-01	MH-0-12FT-COMP		1	S	1/18/17						
Cyanide	PASS	1.235		0.064						mg/Kg	
CCV2	CCV2			W	1/18/17						
Cyanide	PASS	247.937		0.25	100.0	90	110			mg/L	
CCB2	CCB2			W	1/18/17						
Cyanide	PASS	1.059		0.001			+/-0.0030			mg/L	