

CB 85411

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

1/19/2017 11:34

Test: *Reactive*
~~Total~~ CN

Sample Id	<i>ug/L</i> Result	<i>W</i> Dil. 1 +	Response	Errors
ICV1	93.185	0.0	0.094	
ICB1	2.182	0.0	0.006	
CCV1	239.730	0.0	0.237	
CCB1	1.515	0.0	0.005	
PB96146BLS	1.579	0.0	0.005	
I1260-04	1.105	0.0	0.005	
I1260-04D	1.552	0.0	0.005	
CCV2	241.259	0.0	0.238	
CCB2	1.852	0.0	0.005	

N	9
Mean	64.884
SD	103.9749
CV%	160.25

CB 85411

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 19 10:47:48 2017

Thu Jan 19 11:32:10 2017

AK

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	1.0742	µg/l	1/19/2017 11:05:53	
5.OPPBCN	A	Reactive CN	P	5.9833	µg/l	1/19/2017 11:05:54	
10PPBCN	A	Reactive CN	P	10.2421	µg/l	1/19/2017 11:05:55	
50PPBCN	A	Reactive CN	P	52.2484	µg/l	1/19/2017 11:05:56	
100PPBCN	A	Reactive CN	P	97.8338	µg/l	1/19/2017 11:05:57	
250PPBCN	A	Reactive CN	P	244.849	µg/l	1/19/2017 11:05:58	
500PPBCN	A	Reactive CN	P	502.7692	µg/l	1/19/2017 11:05:59	
ICV1	S	Reactive CN	P	93.1845	µg/l	1/19/2017 11:32:02	
ICB1	S	Reactive CN	P	2.1821	µg/l	1/19/2017 11:32:03	
CCV1	S	Reactive CN	P	239.7296	µg/l	1/19/2017 11:32:04	
CCB1	S	Reactive CN	P	1.515	µg/l	1/19/2017 11:32:05	
PB96146BLS	S	Reactive CN	P	1.5794	µg/l	1/19/2017 11:32:06	
I1260-04	S	Reactive CN	P	1.1052	µg/l	1/19/2017 11:32:07	
I1260-04D	S	Reactive CN	P	1.5522	µg/l	1/19/2017 11:32:08	
CCV2	S	Reactive CN	P	241.2587	µg/l	1/19/2017 11:32:09	
CCB2	S	Reactive CN	P	1.8523	µg/l	1/19/2017 11:32:10	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *AK*

1/19/2017 11:06


Test Total CN

Accepted 1/19/2017 11:06

Factor 1028
Bias 0.004

Coeff. of det. 0.999776

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.005	1.0742	0.0000	
2	5.0PPBCN	0.009	5.9833	5.0000	
3	10PPBCN	0.014	10.2421	10.0000	
4	50PPBCN	0.054	52.2484	50.0000	
5	100PPBCN	0.099	97.8338	100.0000	
6	250PPBCN	0.242	244.8490	250.0000	
7	500PPBCN	0.493	502.7692	500.0000	