

Chemtech Consulting Group

Analytical Review Report

Date Printed : 1/19/17

Analyst : AK

Data File : lb85411.csv

Approved By : 

Approved Date : 1/20/17

Worksheet # :

m9012A-B Total Amenable & Reactive Cyanide - 15

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Reactive Cyanide												
0.0PPBCN	0.0PPBCN				W	1/19/17						
Reactive Cyanide		PASS	1.074		0.001						mg/L	
5.0PPBCN	5.0PPBCN				W	1/19/17						
Reactive Cyanide		PASS	5.983		0.006						mg/L	
10PPBCN	10PPBCN				W	1/19/17						
Reactive Cyanide		PASS	10.242		0.010						mg/L	
50PPBCN	50PPBCN				W	1/19/17						
Reactive Cyanide		PASS	52.248		0.052						mg/L	
100PPBCN	100PPBCN				W	1/19/17						
Reactive Cyanide		PASS	97.834		0.098						mg/L	
250PPBCN	250PPBCN				W	1/19/17						
Reactive Cyanide		PASS	244.849		0.245						mg/L	
500PPBCN	500PPBCN				W	1/19/17						
Reactive Cyanide		PASS	502.769		0.503						mg/L	
ICV1	ICV1				W	1/19/17						
Reactive Cyanide		PASS	93.185		0.09	90.0	85	115			mg/L	
ICB1	ICB1				W	1/19/17						
Reactive Cyanide		PASS	2.182		0.002			+/-0.0050			mg/L	
CCV1	CCV1				W	1/19/17						
Reactive Cyanide		PASS	239.730		0.24	96.0	90	110			mg/L	
CCB1	CCB1				W	1/19/17						
Reactive Cyanide		PASS	1.515		0.002			+/-0.0050			mg/L	
PB96146BLS	PB96146BLS				S	1/19/17						
Reactive Cyanide		PASS	1.579		0.016			+/-0.0500			mg/Kg	
I1260-04	MH-0-12FT-COMP			I	S	1/19/17						
Reactive Cyanide		PASS	1.105		0.011						mg/Kg	
I1260-04D	MH-0-12FT-COMPD			I	S	1/19/17						
Reactive Cyanide		PASS	1.552		0.014				0	20	mg/Kg	
CCV2	CCV2				W	1/19/17						
Reactive Cyanide		PASS	241.259		0.24	96.0	90	110			mg/L	
CCB2	CCB2				W	1/19/17						
Reactive Cyanide		PASS	1.852		0.002			+/-0.0050			mg/L	