


7471A

LB85441
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44875



A= 0.0000e+000

B= 2.2116e-004

C= 2.2194e-003

Rho= 0.9997940

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	19	0.000	19				
0.2	0.200	0.209	0.009	937	0.0 %	937				
2.5	2.500	2.485	-0.015	11225	0.0 %	11225				
5.0	5.000	4.924	-0.076	22254	0.0 %	22254				
7.5	7.500	7.649	0.149	34575	0.0 %	34575				
10.0	10.000	9.927	-0.073	44875	0.0 %	44875				

LB85441 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 20 Jan 2017 13:15:34

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7471A	PPB	20 Jan 2017 13:59:46	S	Std
0.2 - 1	50.2	937	-	0.2000	7471A	PPB	20 Jan 2017 14:02:03	S	Std
2.5 - 1	82.5	11225	-	2.5000	7471A	PPB	20 Jan 2017 14:04:21	S	Std
5.0 - 1	85	22254	-	5.0000	7471A	PPB	20 Jan 2017 14:06:38	S	Std
7.5 - 1	87.5	34575	-	7.5000	7471A	PPB	20 Jan 2017 14:11:31	S	Std
10.0 - 1	810	44875	-	10.0000	7471A	PPB	20 Jan 2017 14:16:26	S	Std
ICV05 - 1	ICV05	17668	3.9097	-	7471A	PPB	20 Jan 2017 14:19:08	U	SMPL
ICB05 - 1	ICB05	6	0.0035	-	7471A	PPB	20 Jan 2017 14:21:23	U	SMPL
CCV01 - 1	CCV01	22347	4.9445	-	7471A	PPB	20 Jan 2017 14:23:39	U	SMPL
CCB01 - 1	CCB01	31	0.0091	-	7471A	PPB	20 Jan 2017 14:25:55	U	SMPL
CRA - 1	CRA	943	0.2108	-	7471A	PPB	20 Jan 2017 14:28:11	U	SMPL
HighStd - 1	HighStd	45710	10.1114	-	7471A	PPB	20 Jan 2017 14:30:27	U	SMPL
ChkStd - 1	ChkStd	31636	6.9988	-	7471A	PPB	20 Jan 2017 14:32:43	U	SMPL
PB96167BL - 1	PBS	-56	-0.0102	-	7471A	PPB	20 Jan 2017 14:34:59	U	SMPL
PB96167BS - 1	LCSS	15208	3.3656	-	7471A	PPB	20 Jan 2017 14:37:16	U	SMPL
I1304-01 - 1	ROLL-OFF-15-41-COMP	1918	0.4264	-	7471A	PPB	20 Jan 2017 14:39:33	U	SMPL
I1304-04 - 1	1634-NOTTINGHAM-COMP	2100	0.4667	-	7471A	PPB	20 Jan 2017 14:41:50	U	SMPL
I1304-07 - 1	ROLL-OFF-15939-COMP	4126	0.9147	-	7471A	PPB	20 Jan 2017 14:44:08	U	SMPL
I1308-01 - 1	MH-2-0-12FT-COMP	240	0.0553	-	7471A	PPB	20 Jan 2017 14:46:26	U	SMPL
I1309-01 - 1	ELT-SB-01-0-2	935	0.2090	-	7471A	PPB	20 Jan 2017 14:48:42	U	SMPL
CCV02 - 1	CCV02	22287	4.9312	-	7471A	PPB	20 Jan 2017 14:50:58	U	SMPL
CCB02 - 1	CCB02	12	0.0049	-	7471A	PPB	20 Jan 2017 14:53:14	U	SMPL
I1309-01DUP - 1	ELT-SB-01-0-2DUP	919	0.2055	-	7471A	PPB	20 Jan 2017 14:58:10	U	SMPL
I1309-02 - 1	I1309-01MS	16896	3.7389	-	7471A	PPB	20 Jan 2017 15:00:25	U	SMPL
I1309-03 - 1	I1309-01MSD	16974	3.7562	-	7471A	PPB	20 Jan 2017 15:02:41	U	SMPL
I1309-04 - 1	ELT-SB-101-0-2	2141	0.4757	-	7471A	PPB	20 Jan 2017 15:04:58	U	SMPL
I1309-05 - 1	ELT-SB-02-0-2	1101	0.2457	-	7471A	PPB	20 Jan 2017 15:07:14	U	SMPL
I1309-06 - 1	ELT-SB-03-0-2	412	0.0933	-	7471A	PPB	20 Jan 2017 15:09:31	U	SMPL
I1312-01 - 1	CAS-SB-01-0-2	2658	0.5901	-	7471A	PPB	20 Jan 2017 15:11:48	U	SMPL
I1312-02 - 1	I1312-01MS	21137	4.6769	-	7471A	PPB	20 Jan 2017 15:14:06	U	SMPL
I1312-03 - 1	I1312-01MSD	19462	4.3064	-	7471A	PPB	20 Jan 2017 15:16:23	U	SMPL
I1312-04 - 1	CAS-SB-101-0-2	4050	0.8979	-	7471A	PPB	20 Jan 2017 15:18:39	U	SMPL
CCV03 - 1	CCV03	22254	4.9239	-	7471A	PPB	20 Jan 2017 15:20:55	U	SMPL
CCB03 - 1	CCB03	4	0.0031	-	7471A	PPB	20 Jan 2017 15:23:11	U	SMPL
I1312-05 - 1	CAS-SB-02-0-2	6348	1.4061	-	7471A	PPB	20 Jan 2017 15:25:28	U	SMPL
I1312-06 - 1	CAS-SB-03-0-2	3311	0.7345	-	7471A	PPB	20 Jan 2017 15:27:44	U	SMPL
I1312-07 - 1	CAS-BALLARD01-0-2	4582	1.0156	-	7471A	PPB	20 Jan 2017 15:30:01	U	SMPL
I1312-08 - 1	CAS-SB-02-5-7	7816	1.7308	-	7471A	PPB	20 Jan 2017 15:32:17	U	SMPL
I1312-11 - 1	CAS-SB-03-5-7	64786	14.3302	-	7471A	PPB	20 Jan 2017 15:34:34	U	SMPL
I1326-01 - 1	FK-GF-02-001-A	365	0.0829	-	7471A	PPB	20 Jan 2017 15:36:51	U	SMPL
I1326-03 - 1	FK-GF-02-001-B	506	0.1141	-	7471A	PPB	20 Jan 2017 15:39:08	U	SMPL
I1326-05 - 1	FK-GF-02-001-C	577	0.1298	-	7471A	PPB	20 Jan 2017 15:41:26	U	SMPL
I1309-01LX5 - 1		196	0.0456	-	7471A	PPB	20 Jan 2017 15:43:44	U	SMPL
I1309-01A - 1		17892	3.9592	-	7471A	PPB	20 Jan 2017 15:46:00	U	SMPL
CCV04 - 1	CCV04	22162	4.9035	-	7471A	PPB	20 Jan 2017 15:48:16	U	SMPL
CCB04 - 1	CCB04	-6	0.0009	-	7471A	PPB	20 Jan 2017 15:50:33	U	SMPL
I1312-11DLX5 - 1		12997	2.8766	-	7471A	PPB	20 Jan 2017 15:52:49	U	SMPL
CCV05 - 1	CCV05	22204	4.9128	-	7471A	PPB	20 Jan 2017 15:55:05	U	SMPL
CCB05 - 1	CCB05	0	0.0022	-	7471A	PPB	20 Jan 2017 15:57:22	U	SMPL